

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102622\
 Data File : PP052702.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Oct 2022 15:56
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 26 21:53:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 26 21:51:00 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.411	3.655	178.1E6	158.2E6	73.788	74.182
2) SA Decachlor...	10.225	8.745	101.0E6	133.6E6	75.118	71.601
Target Compounds						
16) L4 AR-1242-1	5.586	4.757	39755529	39266671	721.461	720.030
17) L4 AR-1242-2	5.608	4.776	58664150	54826944	727.594	724.711
18) L4 AR-1242-3	5.671	4.955	35694773	29439192	720.615	719.551
19) L4 AR-1242-4	5.770	5.040	28568402	29835404	726.388	713.108
20) L4 AR-1242-5	6.511	5.571	27922136	35784404	712.706	715.566



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102622\
Data File : PP052702.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Oct 2022 15:56
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 21:53:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 26 21:51:00 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.411 min
Delta R.T.: 0.000 min
Response: 178077236
Conc: 73.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.655 min
Delta R.T.: 0.000 min
Response: 158185374
Conc: 74.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.225 min
Delta R.T.: 0.000 min
Response: 101033455
Conc: 75.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.745 min
Delta R.T.: 0.000 min
Response: 133563781
Conc: 71.60 ng/ml



#16 AR-1242-1


R.T.: 5.586 min
Delta R.T.: 0.000 min
Response: 39755529
Conc: 721.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#16 AR-1242-1

R.T.: 4.757 min
Delta R.T.: 0.000 min
Response: 39266671
Conc: 720.03 ng/ml




#17 AR-1242-2

R.T.: 5.608 min
Delta R.T.: 0.000 min
Response: 58664150
Conc: 727.59 ng/ml



#17 AR-1242-2


R.T.: 4.776 min
Delta R.T.: 0.000 min
Response: 54826944
Conc: 724.71 ng/ml



#18 AR-1242-3


R.T.: 5.671 min
Delta R.T.: 0.000 min
Response: 35694773
Conc: 720.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#18 AR-1242-3

R.T.: 4.955 min
Delta R.T.: 0.000 min
Response: 29439192
Conc: 719.55 ng/ml




#19 AR-1242-4

R.T.: 5.770 min
Delta R.T.: 0.000 min
Response: 28568402
Conc: 726.39 ng/ml



#19 AR-1242-4


R.T.: 5.040 min
Delta R.T.: 0.000 min
Response: 29835404
Conc: 713.11 ng/ml



#20 AR-1242-5

R.T.: 6.511 min
Delta R.T.: 0.000 min
Response: 27922136
Conc: 712.71 ng/ml

Instrument :
ECD_P
ClientSampleId :



#20 AR-1242-5

R.T.: 5.571 min
Delta R.T.: 0.000 min
Response: 35784404
Conc: 715.57 ng/ml