

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102721\
 Data File : PP040486.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Oct 2021 8:11
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 08:18:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 27 13:26:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.751	3.773	2370914	1432271	58.210	52.575
2)	SA Decachlor...	10.682	9.058	1585780	1352364	52.343	48.203
Target Compounds							
3)	L1 AR-1016-1	6.073	5.025	21549	14760	16.820	14.956
4)	L1 AR-1016-2	6.100	5.043	25011	12999	12.503	8.927 #
5)	L1 AR-1016-3	6.160	5.234	11491	8546	9.150	10.669
6)	L1 AR-1016-4	6.267	5.288	12729	321851	12.876	484.839 #
7)	L1 AR-1016-5	6.583	5.515	305357	182378	314.943	247.516
8)	L2 AR-1221-1	4.990	0.000	13372	0	27.592	N.D. #
9)	L2 AR-1221-2	5.101	4.123	31499	17281	92.800	66.260 #
10)	L2 AR-1221-3	0.000	4.216	0	27264	N.D.	32.060 #
11)	L3 AR-1232-1	0.000	4.216	0	27264	N.D.	38.607 #
12)	L3 AR-1232-2	5.776	5.043	12096	12999	27.300	21.734
13)	L3 AR-1232-3	6.100	5.234	25011	8546	30.357	27.509
14)	L3 AR-1232-4	6.267	5.331	12729	91103	32.331	355.534 #
15)	L3 AR-1232-5	6.367	5.515	488656	182378	1837.431	678.093 #
16)	L4 AR-1242-1	6.073	5.025	21549	14760	22.268	20.333
17)	L4 AR-1242-2	6.100	5.043	25011	12999	16.646	12.251 #
18)	L4 AR-1242-3	6.160	5.234	11491	8546	12.183	14.696
19)	L4 AR-1242-4	6.267	5.331	12729	91103	16.877	177.594 #
20)	L4 AR-1242-5	7.055	5.896	297972	750855	376.332	1106.377 #
21)	L5 AR-1248-1	6.073	5.025	21549	14760	28.846	28.210
22)	L5 AR-1248-2	6.367	5.288	488656	321851	472.032	403.840
23)	L5 AR-1248-3	6.583	5.331	305357	91103	237.342	122.659 #
24)	L5 AR-1248-4	7.014	5.515	458573	182378	312.018	218.251 #
25)	L5 AR-1248-5	7.055	5.938	297972	117454	208.359	119.476 #
26)	L6 AR-1254-1	6.983	5.896	821607	750855	568.491	531.874
27)	L6 AR-1254-2	7.213	6.057	1278733	661658	568.059	532.425
28)	L6 AR-1254-3	7.594	6.478	1403409	998219	582.340	524.539
29)	L6 AR-1254-4	7.892	6.718	978522	632220	577.005	546.716
30)	L6 AR-1254-5	8.321	7.149	1004332	932811	552.094	539.333
31)	L7 AR-1260-1	7.762	6.616	579144	520458	371.113	364.065
32)	L7 AR-1260-2	8.029	6.814	580942	435513	315.227	260.176
33)	L7 AR-1260-3	8.381f	6.966	134278	428533	96.822	270.990 #
34)	L7 AR-1260-4	8.617	7.453	414027	47602	262.223	35.217 #
35)	L7 AR-1260-5	8.943	7.703	137417	110652	46.623	36.854
36)	L8 AR-1262-1	8.381f	7.218f	134278	77566	63.532	44.751 #
37)	L8 AR-1262-2	8.943	7.703	137417	110652	38.364	36.765
38)	L8 AR-1262-3	9.261f	7.991	104898	7727	59.790	6.032 #
39)	L8 AR-1262-4	9.309f	8.051	26830	116438	23.308	49.801 #
40)	L8 AR-1262-5	0.000	8.554	0	4477	N.D.	3.928 #
41)	L9 AR-1268-1	9.261f	7.991	104898	7727	25.402	2.186 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102721\
 Data File : PP040486.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Oct 2021 8:11
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 08:18:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 27 13:26:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.309f	8.051	26830	116438	6.841	34.551 #
43) L9	AR-1268-3	0.000	8.264	0	13105	N.D.	4.460 #
44) L9	AR-1268-4	0.000	8.554	0	4477	N.D.	3.687 #
45) L9	AR-1268-5	0.000	8.827	0	8100	N.D.	0.902 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102721\
Data File : PP040486.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Oct 2021 8:11
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 08:18:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 27 13:26:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.751 min
Delta R.T.: 0.000 min
Response: 2370914
Conc: 58.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.773 min
Delta R.T.: 0.000 min
Response: 1432271
Conc: 52.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.682 min
Delta R.T.: -0.009 min
Response: 1585780
Conc: 52.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.058 min
Delta R.T.: -0.008 min
Response: 1352364
Conc: 48.20 ng/ml



#3 AR-1016-1


R.T.: 6.073 min
Delta R.T.: 0.000 min
Response: 21549
Conc: 16.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.025 min
Delta R.T.: -0.001 min
Response: 14760
Conc: 14.96 ng/ml




#4 AR-1016-2

R.T.: 6.100 min
Delta R.T.: 0.002 min
Response: 25011
Conc: 12.50 ng/ml



#4 AR-1016-2


R.T.: 5.043 min
Delta R.T.: -0.003 min
Response: 12999
Conc: 8.93 ng/ml



#5 AR-1016-3


R.T.: 6.160 min
Delta R.T.: -0.003 min
Response: 11491
Conc: 9.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.234 min
Delta R.T.: -0.004 min
Response: 8546
Conc: 10.67 ng/ml




#6 AR-1016-4

R.T.: 6.267 min
Delta R.T.: -0.002 min
Response: 12729
Conc: 12.88 ng/ml



#6 AR-1016-4


R.T.: 5.288 min
Delta R.T.: -0.003 min
Response: 321851
Conc: 484.84 ng/ml



#7 AR-1016-5


R.T.: 6.583 min
Delta R.T.: -0.003 min
Response: 305357
Conc: 314.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.515 min
Delta R.T.: -0.003 min
Response: 182378
Conc: 247.52 ng/ml




#8 AR-1221-1

R.T.: 4.990 min
Delta R.T.: -0.013 min
Response: 13372
Conc: 27.59 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 4.032 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 5.101 min
Delta R.T.: 0.000 min
Response: 31499
Conc: 92.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.123 min
Delta R.T.: -0.008 min
Response: 17281
Conc: 66.26 ng/ml




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 5.189 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 4.216 min
Delta R.T.: -0.002 min
Response: 27264
Conc: 32.06 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T.: 5.189 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.216 min
Delta R.T.: -0.002 min
Response: 27264
Conc: 38.61 ng/ml




#12 AR-1232-2

R.T.: 5.776 min
Delta R.T.: -0.001 min
Response: 12096
Conc: 27.30 ng/ml



#12 AR-1232-2


R.T.: 5.043 min
Delta R.T.: -0.003 min
Response: 12999
Conc: 21.73 ng/ml



#13 AR-1232-3


R.T.: 6.100 min
Delta R.T.: 0.002 min
Response: 25011
Conc: 30.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.234 min
Delta R.T.: -0.003 min
Response: 8546
Conc: 27.51 ng/ml




#14 AR-1232-4

R.T.: 6.267 min
Delta R.T.: -0.003 min
Response: 12729
Conc: 32.33 ng/ml



#14 AR-1232-4


R.T.: 5.331 min
Delta R.T.: -0.004 min
Response: 91103
Conc: 355.53 ng/ml



#15 AR-1232-5


R.T.: 6.367 min
Delta R.T.: -0.002 min
Response: 488656
Conc: 1837.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.515 min
Delta R.T.: -0.003 min
Response: 182378
Conc: 678.09 ng/ml




#16 AR-1242-1

R.T.: 6.073 min
Delta R.T.: -0.002 min
Response: 21549
Conc: 22.27 ng/ml



#16 AR-1242-1


R.T.: 5.025 min
Delta R.T.: -0.002 min
Response: 14760
Conc: 20.33 ng/ml



#17 AR-1242-2


R.T.: 6.100 min
Delta R.T.: 0.001 min
Response: 25011
Conc: 16.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.043 min
Delta R.T.: -0.003 min
Response: 12999
Conc: 12.25 ng/ml




#18 AR-1242-3

R.T.: 6.160 min
Delta R.T.: -0.004 min
Response: 11491
Conc: 12.18 ng/ml



#18 AR-1242-3


R.T.: 5.234 min
Delta R.T.: -0.004 min
Response: 8546
Conc: 14.70 ng/ml



#19 AR-1242-4


R.T.: 6.267 min
Delta R.T.: -0.004 min
Response: 12729
Conc: 16.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.331 min
Delta R.T.: -0.005 min
Response: 91103
Conc: 177.59 ng/ml




#20 AR-1242-5

R.T.: 7.055 min
Delta R.T.: -0.002 min
Response: 297972
Conc: 376.33 ng/ml



#20 AR-1242-5


R.T.: 5.896 min
Delta R.T.: -0.004 min
Response: 750855
Conc: 1106.38 ng/ml



#21 AR-1248-1


R.T.: 6.073 min
Delta R.T.: -0.002 min
Response: 21549
Conc: 28.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.025 min
Delta R.T.: -0.001 min
Response: 14760
Conc: 28.21 ng/ml




#22 AR-1248-2

R.T.: 6.367 min
Delta R.T.: -0.002 min
Response: 488656
Conc: 472.03 ng/ml



#22 AR-1248-2


R.T.: 5.288 min
Delta R.T.: -0.003 min
Response: 321851
Conc: 403.84 ng/ml



#23 AR-1248-3


R.T.: 6.583 min
Delta R.T.: -0.001 min
Response: 305357
Conc: 237.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.331 min
Delta R.T.: -0.006 min
Response: 91103
Conc: 122.66 ng/ml




#24 AR-1248-4

R.T.: 7.014 min
Delta R.T.: -0.002 min
Response: 458573
Conc: 312.02 ng/ml



#24 AR-1248-4


R.T.: 5.515 min
Delta R.T.: -0.002 min
Response: 182378
Conc: 218.25 ng/ml



#25 AR-1248-5


R.T.: 7.055 min
Delta R.T.: 0.000 min
Response: 297972
Conc: 208.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.938 min
Delta R.T.: -0.004 min
Response: 117454
Conc: 119.48 ng/ml




#26 AR-1254-1

R.T.: 6.983 min
Delta R.T.: -0.002 min
Response: 821607
Conc: 568.49 ng/ml



#26 AR-1254-1


R.T.: 5.896 min
Delta R.T.: -0.003 min
Response: 750855
Conc: 531.87 ng/ml



#27 AR-1254-2


R.T.: 7.213 min
Delta R.T.: 0.000 min
Response: 1278733
Conc: 568.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.057 min
Delta R.T.: -0.003 min
Response: 661658
Conc: 532.43 ng/ml




#28 AR-1254-3

R.T.: 7.594 min
Delta R.T.: -0.002 min
Response: 1403409
Conc: 582.34 ng/ml



#28 AR-1254-3


R.T.: 6.478 min
Delta R.T.: -0.004 min
Response: 998219
Conc: 524.54 ng/ml



#29 AR-1254-4


R.T.: 7.892 min
Delta R.T.: -0.003 min
Response: 978522
Conc: 577.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.718 min
Delta R.T.: -0.005 min
Response: 632220
Conc: 546.72 ng/ml




#30 AR-1254-5

R.T.: 8.321 min
Delta R.T.: 0.000 min
Response: 1004332
Conc: 552.09 ng/ml



#30 AR-1254-5


R.T.: 7.149 min
Delta R.T.: -0.005 min
Response: 932811
Conc: 539.33 ng/ml



#31 AR-1260-1


R.T.: 7.762 min
Delta R.T.: -0.002 min
Response: 579144
Conc: 371.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.616 min
Delta R.T.: -0.004 min
Response: 520458
Conc: 364.06 ng/ml




#32 AR-1260-2

R.T.: 8.029 min
Delta R.T.: -0.003 min
Response: 580942
Conc: 315.23 ng/ml



#32 AR-1260-2


R.T.: 6.814 min
Delta R.T.: -0.005 min
Response: 435513
Conc: 260.18 ng/ml



#33 AR-1260-3


R.T.: 8.381 min
Delta R.T.: -0.017 min
Response: 134278
Conc: 96.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.966 min
Delta R.T.: -0.006 min
Response: 428533
Conc: 270.99 ng/ml




#34 AR-1260-4

R.T.: 8.617 min
Delta R.T.: -0.012 min
Response: 414027
Conc: 262.22 ng/ml



#34 AR-1260-4


R.T.: 7.453 min
Delta R.T.: -0.006 min
Response: 47602
Conc: 35.22 ng/ml



#35 AR-1260-5


R.T.: 8.943 min
Delta R.T.: -0.003 min
Response: 137417
Conc: 46.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.703 min
Delta R.T.: -0.007 min
Response: 110652
Conc: 36.85 ng/ml




#36 AR-1262-1

R.T.: 8.381 min
Delta R.T.: -0.016 min
Response: 134278
Conc: 63.53 ng/ml



#36 AR-1262-1


R.T.: 7.218 min
Delta R.T.: 0.023 min
Response: 77566
Conc: 44.75 ng/ml



#37 AR-1262-2


R.T.: 8.943 min
Delta R.T.: -0.003 min
Response: 137417
Conc: 38.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.703 min
Delta R.T.: -0.006 min
Response: 110652
Conc: 36.76 ng/ml




#38 AR-1262-3

R.T.: 9.261 min
Delta R.T.: 0.017 min
Response: 104898
Conc: 59.79 ng/ml



#38 AR-1262-3


R.T.: 7.991 min
Delta R.T.: -0.005 min
Response: 7727
Conc: 6.03 ng/ml



#39 AR-1262-4


R.T.: 9.309 min
Delta R.T.: -0.024 min
Response: 26830
Conc: 23.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.051 min
Delta R.T.: -0.010 min
Response: 116438
Conc: 49.80 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 9.975 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.554 min
Delta R.T.: -0.008 min
Response: 4477
Conc: 3.93 ng/ml



#41 AR-1268-1


R.T.: 9.261 min
Delta R.T.: 0.018 min
Response: 104898
Conc: 25.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.991 min
Delta R.T.: -0.005 min
Response: 7727
Conc: 2.19 ng/ml




#42 AR-1268-2

R.T.: 9.309 min
Delta R.T.: -0.027 min
Response: 26830
Conc: 6.84 ng/ml



#42 AR-1268-2


R.T.: 8.051 min
Delta R.T.: -0.011 min
Response: 116438
Conc: 34.55 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.551 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.264 min
Delta R.T.: -0.004 min
Response: 13105
Conc: 4.46 ng/ml




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 9.974 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 8.554 min
Delta R.T.: -0.008 min
Response: 4477
Conc: 3.69 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.369 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.827 min
Delta R.T.: -0.008 min
Response: 8100
Conc: 0.90 ng/ml