

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102721\
 Data File : PP040479.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Oct 2021 2:48
 Operator : AJ\MA
 Sample : M4363-01
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 33509-512

Manual Integrations
 APPROVED

mohammad
 10/28/2021 1:35:26 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 07:18:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 27 13:26:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.752	3.773	625136	472477	15.348m	17.343m
2) SA Decachlor...	10.686	9.060	422012	320508	13.930	11.424m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102721\
Data File : PP040479.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Oct 2021 2:48
Operator : AJ\MA
Sample : M4363-01
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
33509-512


Manual Integrations
APPROVED

mohammad
10/28/2021 1:35:26 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 07:18:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 27 13:26:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.752 min
Delta R.T.: 0.000 min
Response: 625136
Conc: 15.35 ng/ml m

Instrument :
ECD_P
ClientSampleId :
33509-512


Manual Integrations
APPROVED

mohammad
10/28/2021 1:35:26 PM




#1 Tetrachloro-m-xylene

R.T.: 3.773 min
Delta R.T.: 0.000 min
Response: 472477
Conc: 17.34 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.686 min
Delta R.T.: -0.006 min
Response: 422012
Conc: 13.93 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.060 min
Delta R.T.: -0.007 min
Response: 320508
Conc: 11.42 ng/ml m