

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102722\
 Data File : PP052746.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2022 14:13
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 14:48:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 27 04:46:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.409	3.653	107.1E6	100.5E6	48.449	53.483
2) SA Decachlor...	10.220	8.741	66286746	83556744	52.223	49.822
Target Compounds						
26) L6 AR-1254-1	6.444	5.569	33244263	49406549	437.964	460.094
27) L6 AR-1254-2	6.664	5.718	49403941	43639853	442.029	456.560
28) L6 AR-1254-3	7.032	6.128	49749456	69282728	446.946	453.465
29) L6 AR-1254-4	7.319	6.359	33195737	41617885	471.394	458.376
30) L6 AR-1254-5	7.740	6.782	37478648	62841924	476.021	451.680



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102722\
Data File : PP052746.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2022 14:13
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 14:48:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 27 04:46:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.409 min
Delta R.T.: -0.001 min
Response: 107056785
Conc: 48.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.653 min
Delta R.T.: 0.000 min
Response: 100506999
Conc: 53.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: -0.003 min
Response: 66286746
Conc: 52.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.741 min
Delta R.T.: -0.002 min
Response: 83556744
Conc: 49.82 ng/ml



#26 AR-1254-1


R.T.: 6.444 min
Delta R.T.: -0.004 min
Response: 33244263
Conc: 437.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.569 min
Delta R.T.: -0.003 min
Response: 49406549
Conc: 460.09 ng/ml




#27 AR-1254-2

R.T.: 6.664 min
Delta R.T.: -0.003 min
Response: 49403941
Conc: 442.03 ng/ml



#27 AR-1254-2


R.T.: 5.718 min
Delta R.T.: -0.003 min
Response: 43639853
Conc: 456.56 ng/ml



#28 AR-1254-3


R.T.: 7.032 min
Delta R.T.: -0.004 min
Response: 49749456
Conc: 446.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.128 min
Delta R.T.: -0.004 min
Response: 69282728
Conc: 453.46 ng/ml




#29 AR-1254-4

R.T.: 7.319 min
Delta R.T.: -0.003 min
Response: 33195737
Conc: 471.39 ng/ml



#29 AR-1254-4


R.T.: 6.359 min
Delta R.T.: -0.004 min
Response: 41617885
Conc: 458.38 ng/ml



#30 AR-1254-5

R.T.: 7.740 min
Delta R.T.: -0.004 min
Response: 37478648
Conc: 476.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.782 min
Delta R.T.: -0.004 min
Response: 62841924
Conc: 451.68 ng/ml