

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102722\
 Data File : PP052791.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2022 04:13
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 05:43:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 27 04:46:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.405	3.650	102.8E6	96612832	46.535	51.411
2) SA Decachlor...	10.210	8.735	64412454	81971090	50.746	48.876
Target Compounds						
21) L5 AR-1248-1	5.579	4.750	19007460	19833817	467.971	495.545
22) L5 AR-1248-2	5.855	4.991	27401355	26466669	460.378	467.435
23) L5 AR-1248-3	6.060	5.033	29726277	28059476	457.900	477.916
24) L5 AR-1248-4	6.465	5.208	31052441	33552437	482.149	481.847
25) L5 AR-1248-5	6.504	5.606	30463467	28141442	480.890	443.587



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102722\
Data File : PP052791.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2022 04:13
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 05:43:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 27 04:46:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.405 min
Delta R.T.: -0.005 min
Response: 102829169
Conc: 46.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.650 min
Delta R.T.: -0.004 min
Response: 96612832
Conc: 51.41 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: -0.013 min
Response: 64412454
Conc: 50.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.735 min
Delta R.T.: -0.008 min
Response: 81971090
Conc: 48.88 ng/ml



#21 AR-1248-1


R.T.: 5.579 min
Delta R.T.: -0.007 min
Response: 19007460
Conc: 467.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.750 min
Delta R.T.: -0.006 min
Response: 19833817
Conc: 495.55 ng/ml




#22 AR-1248-2

R.T.: 5.855 min
Delta R.T.: -0.007 min
Response: 27401355
Conc: 460.38 ng/ml



#22 AR-1248-2


R.T.: 4.991 min
Delta R.T.: -0.006 min
Response: 26466669
Conc: 467.44 ng/ml



#23 AR-1248-3


R.T.: 6.060 min
Delta R.T.: -0.008 min
Response: 29726277
Conc: 457.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.033 min
Delta R.T.: -0.006 min
Response: 28059476
Conc: 477.92 ng/ml




#24 AR-1248-4

R.T.: 6.465 min
Delta R.T.: -0.008 min
Response: 31052441
Conc: 482.15 ng/ml



#24 AR-1248-4


R.T.: 5.208 min
Delta R.T.: -0.007 min
Response: 33552437
Conc: 481.85 ng/ml



#25 AR-1248-5

R.T.: 6.504 min
Delta R.T.: -0.009 min
Response: 30463467
Conc: 480.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.606 min
Delta R.T.: -0.008 min
Response: 28141442
Conc: 443.59 ng/ml