

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102821\
 Data File : PP040515.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Oct 2021 16:19
 Operator : AJ\MA
 Sample : PB140316BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 17:52:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 28 08:58:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.750	3.772	888487	478595	21.814	17.568
2)	SA Decachlor...	10.681	9.057	634600	566785	20.947	20.202
Target Compounds							
3)	L1 AR-1016-1	6.071	5.023	603853	377917	471.352	382.916
4)	L1 AR-1016-2	6.094	5.042	923712	535390	461.765	367.667
5)	L1 AR-1016-3	6.160	5.234	580903	299154	462.562	373.451
6)	L1 AR-1016-4	6.267	5.287	483477	238515	489.054	359.301 #
7)	L1 AR-1016-5	6.582	5.515	466435	303753	481.077	412.241
8)	L2 AR-1221-1	4.995	4.029	9094	36354	18.764	110.107 #
9)	L2 AR-1221-2	5.100	4.129	75291	48247	221.817	184.995
10)	L2 AR-1221-3	5.187	4.216	350303	233419	326.070	274.475
11)	L3 AR-1232-1	5.187	4.216	350303	233419	395.013	330.526
12)	L3 AR-1232-2	5.774	5.042	451043	535390	1018.005	895.173
13)	L3 AR-1232-3	6.094	5.234	923712	299154	1121.180	962.896
14)	L3 AR-1232-4	6.267	5.331	483477	295416	1228.002	1152.881
15)	L3 AR-1232-5	6.366	5.515	368336	303753	1385.004	1129.370
16)	L4 AR-1242-1	6.071	5.023	603853	377917	624.004	520.586
17)	L4 AR-1242-2	6.094	5.042	923712	535390	614.795	504.592
18)	L4 AR-1242-3	6.160	5.234	580903	299154	615.905	514.398
19)	L4 AR-1242-4	6.267	5.331	483477	295416	641.043	575.879
20)	L4 AR-1242-5	7.052	5.895	67020	238763	84.645	351.815 #
21)	L5 AR-1248-1	6.071	5.023	603853	377917	808.351	722.275
22)	L5 AR-1248-2	6.366	5.287	368336	238515	355.804	299.275
23)	L5 AR-1248-3	6.582	5.331	466435	295416	362.542	397.742
24)	L5 AR-1248-4	7.012	5.515	76690	303753	52.180	363.499 #
25)	L5 AR-1248-5	7.052	5.938	67020	37013	46.864	37.650
26)	L6 AR-1254-1	6.981	5.895	343116	238763	232.573	169.130 #
27)	L6 AR-1254-2	7.211	6.056	388274	233120	172.485	187.588
28)	L6 AR-1254-3	7.593	6.495	225206	406177	93.449	213.436 #
29)	L6 AR-1254-4	7.889	6.717	144389	46211	84.066	39.961 #
30)	L6 AR-1254-5	8.319	7.148	1182540	802862	650.057	464.199 #
31)	L7 AR-1260-1	7.761	6.614	832415	570988	533.407	399.411 #
32)	L7 AR-1260-2	8.028	6.813	969212	674151	525.907	402.739
33)	L7 AR-1260-3	8.394	6.967	620780	639751	447.619	404.557
34)	L7 AR-1260-4	8.624	7.452	741125	486178	469.389	359.690
35)	L7 AR-1260-5	8.941	7.703	1322558	1106520	448.715	368.543
36)	L8 AR-1262-1	8.394	7.189	620780	525567	293.716	303.219
37)	L8 AR-1262-2	8.941	7.703	1322558	1106520	369.227	367.651
38)	L8 AR-1262-3	9.259f	7.989	871017	234692	496.464	183.207 #
39)	L8 AR-1262-4	9.311f	8.051	506937	841039	440.387	359.719
40)	L8 AR-1262-5	9.968	8.554	333116	274606	251.694	240.933
41)	L9 AR-1268-1	9.259f	7.989	871017	234692	210.928	66.387 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102821\
 Data File : PP040515.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Oct 2021 16:19
 Operator : AJ\MA
 Sample : PB140316BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 17:52:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 28 08:58:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.311f	8.051	506937	841039	129.253	249.563 #
43) L9	AR-1268-3	9.545	8.261	26791	24282	8.191	8.264
44) L9	AR-1268-4	9.968	8.554	333116	274606	238.354	226.171
45) L9	AR-1268-5	10.362	8.827	108411	98202	10.839	10.934



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102821\
Data File : PP040515.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Oct 2021 16:19
Operator : AJ\MA
Sample : PB140316BS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 17:52:37 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 28 08:58:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.750 min
Delta R.T.: -0.001 min
Response: 888487
Conc: 21.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.772 min
Delta R.T.: 0.000 min
Response: 478595
Conc: 17.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.681 min
Delta R.T.: -0.010 min
Response: 634600
Conc: 20.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.057 min
Delta R.T.: -0.010 min
Response: 566785
Conc: 20.20 ng/ml



#3 AR-1016-1


R.T.: 6.071 min
Delta R.T.: -0.002 min
Response: 603853
Conc: 471.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.023 min
Delta R.T.: -0.003 min
Response: 377917
Conc: 382.92 ng/ml




#4 AR-1016-2

R.T.: 6.094 min
Delta R.T.: -0.003 min
Response: 923712
Conc: 461.77 ng/ml



#4 AR-1016-2


R.T.: 5.042 min
Delta R.T.: -0.004 min
Response: 535390
Conc: 367.67 ng/ml



#5 AR-1016-3


R.T.: 6.160 min
Delta R.T.: -0.003 min
Response: 580903
Conc: 462.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.234 min
Delta R.T.: -0.004 min
Response: 299154
Conc: 373.45 ng/ml




#6 AR-1016-4

R.T.: 6.267 min
Delta R.T.: -0.002 min
Response: 483477
Conc: 489.05 ng/ml



#6 AR-1016-4


R.T.: 5.287 min
Delta R.T.: -0.004 min
Response: 238515
Conc: 359.30 ng/ml



#7 AR-1016-5


R.T.: 6.582 min
Delta R.T.: -0.004 min
Response: 466435
Conc: 481.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.515 min
Delta R.T.: -0.003 min
Response: 303753
Conc: 412.24 ng/ml




#8 AR-1221-1

R.T.: 4.995 min
Delta R.T.: -0.008 min
Response: 9094
Conc: 18.76 ng/ml



#8 AR-1221-1


R.T.: 4.029 min
Delta R.T.: -0.003 min
Response: 36354
Conc: 110.11 ng/ml



#9 AR-1221-2


R.T.: 5.100 min
Delta R.T.: -0.002 min
Response: 75291
Conc: 221.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.129 min
Delta R.T.: -0.001 min
Response: 48247
Conc: 185.00 ng/ml




#10 AR-1221-3

R.T.: 5.187 min
Delta R.T.: -0.002 min
Response: 350303
Conc: 326.07 ng/ml



#10 AR-1221-3


R.T.: 4.216 min
Delta R.T.: -0.002 min
Response: 233419
Conc: 274.47 ng/ml



#11 AR-1232-1


R.T.: 5.187 min
Delta R.T.: -0.002 min
Response: 350303
Conc: 395.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.216 min
Delta R.T.: -0.002 min
Response: 233419
Conc: 330.53 ng/ml




#12 AR-1232-2

R.T.: 5.774 min
Delta R.T.: -0.003 min
Response: 451043
Conc: 1018.01 ng/ml



#12 AR-1232-2


R.T.: 5.042 min
Delta R.T.: -0.003 min
Response: 535390
Conc: 895.17 ng/ml



#13 AR-1232-3


R.T.: 6.094 min
Delta R.T.: -0.003 min
Response: 923712
Conc: 1121.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.234 min
Delta R.T.: -0.003 min
Response: 299154
Conc: 962.90 ng/ml




#14 AR-1232-4

R.T.: 6.267 min
Delta R.T.: -0.003 min
Response: 483477
Conc: 1228.00 ng/ml



#14 AR-1232-4


R.T.: 5.331 min
Delta R.T.: -0.004 min
Response: 295416
Conc: 1152.88 ng/ml



#15 AR-1232-5


R.T.: 6.366 min
Delta R.T.: -0.002 min
Response: 368336
Conc: 1385.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.515 min
Delta R.T.: -0.003 min
Response: 303753
Conc: 1129.37 ng/ml




#16 AR-1242-1

R.T.: 6.071 min
Delta R.T.: -0.003 min
Response: 603853
Conc: 624.00 ng/ml



#16 AR-1242-1


R.T.: 5.023 min
Delta R.T.: -0.004 min
Response: 377917
Conc: 520.59 ng/ml



#17 AR-1242-2


R.T.: 6.094 min
Delta R.T.: -0.005 min
Response: 923712
Conc: 614.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.042 min
Delta R.T.: -0.004 min
Response: 535390
Conc: 504.59 ng/ml




#18 AR-1242-3

R.T.: 6.160 min
Delta R.T.: -0.004 min
Response: 580903
Conc: 615.90 ng/ml



#18 AR-1242-3


R.T.: 5.234 min
Delta R.T.: -0.004 min
Response: 299154
Conc: 514.40 ng/ml



#19 AR-1242-4


R.T.: 6.267 min
Delta R.T.: -0.004 min
Response: 483477
Conc: 641.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.331 min
Delta R.T.: -0.005 min
Response: 295416
Conc: 575.88 ng/ml




#20 AR-1242-5

R.T.: 7.052 min
Delta R.T.: -0.005 min
Response: 67020
Conc: 84.65 ng/ml



#20 AR-1242-5


R.T.: 5.895 min
Delta R.T.: -0.005 min
Response: 238763
Conc: 351.82 ng/ml



#21 AR-1248-1


R.T.: 6.071 min
Delta R.T.: -0.003 min
Response: 603853
Conc: 808.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.023 min
Delta R.T.: -0.003 min
Response: 377917
Conc: 722.28 ng/ml




#22 AR-1248-2

R.T.: 6.366 min
Delta R.T.: -0.002 min
Response: 368336
Conc: 355.80 ng/ml



#22 AR-1248-2


R.T.: 5.287 min
Delta R.T.: -0.004 min
Response: 238515
Conc: 299.28 ng/ml



#23 AR-1248-3


R.T.: 6.582 min
Delta R.T.: -0.002 min
Response: 466435
Conc: 362.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.331 min
Delta R.T.: -0.006 min
Response: 295416
Conc: 397.74 ng/ml




#24 AR-1248-4

R.T.: 7.012 min
Delta R.T.: -0.003 min
Response: 76690
Conc: 52.18 ng/ml



#24 AR-1248-4


R.T.: 5.515 min
Delta R.T.: -0.002 min
Response: 303753
Conc: 363.50 ng/ml



#25 AR-1248-5


R.T.: 7.052 min
Delta R.T.: -0.003 min
Response: 67020
Conc: 46.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.938 min
Delta R.T.: -0.004 min
Response: 37013
Conc: 37.65 ng/ml




#26 AR-1254-1

R.T.: 6.981 min
Delta R.T.: -0.003 min
Response: 343116
Conc: 232.57 ng/ml



#26 AR-1254-1


R.T.: 5.895 min
Delta R.T.: -0.004 min
Response: 238763
Conc: 169.13 ng/ml



#27 AR-1254-2


R.T.: 7.211 min
Delta R.T.: -0.003 min
Response: 388274
Conc: 172.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.056 min
Delta R.T.: -0.004 min
Response: 233120
Conc: 187.59 ng/ml




#28 AR-1254-3

R.T.: 7.593 min
Delta R.T.: -0.003 min
Response: 225206
Conc: 93.45 ng/ml



#28 AR-1254-3


R.T.: 6.495 min
Delta R.T.: 0.013 min
Response: 406177
Conc: 213.44 ng/ml



#29 AR-1254-4


R.T.: 7.889 min
Delta R.T.: -0.006 min
Response: 144389
Conc: 84.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.717 min
Delta R.T.: -0.006 min
Response: 46211
Conc: 39.96 ng/ml




#30 AR-1254-5

R.T.: 8.319 min
Delta R.T.: -0.003 min
Response: 1182540
Conc: 650.06 ng/ml



#30 AR-1254-5


R.T.: 7.148 min
Delta R.T.: -0.006 min
Response: 802862
Conc: 464.20 ng/ml



#31 AR-1260-1


R.T.: 7.761 min
Delta R.T.: -0.003 min
Response: 832415
Conc: 533.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.614 min
Delta R.T.: -0.005 min
Response: 570988
Conc: 399.41 ng/ml




#32 AR-1260-2

R.T.: 8.028 min
Delta R.T.: -0.004 min
Response: 969212
Conc: 525.91 ng/ml



#32 AR-1260-2


R.T.: 6.813 min
Delta R.T.: -0.006 min
Response: 674151
Conc: 402.74 ng/ml



#33 AR-1260-3


R.T.: 8.394 min
Delta R.T.: -0.005 min
Response: 620780
Conc: 447.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.967 min
Delta R.T.: -0.006 min
Response: 639751
Conc: 404.56 ng/ml




#34 AR-1260-4

R.T.: 8.624 min
Delta R.T.: -0.005 min
Response: 741125
Conc: 469.39 ng/ml



#34 AR-1260-4


R.T.: 7.452 min
Delta R.T.: -0.007 min
Response: 486178
Conc: 359.69 ng/ml



#35 AR-1260-5


R.T.: 8.941 min
Delta R.T.: -0.005 min
Response: 1322558
Conc: 448.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.703 min
Delta R.T.: -0.007 min
Response: 1106520
Conc: 368.54 ng/ml




#36 AR-1262-1

R.T.: 8.394 min
Delta R.T.: -0.003 min
Response: 620780
Conc: 293.72 ng/ml



#36 AR-1262-1


R.T.: 7.189 min
Delta R.T.: -0.006 min
Response: 525567
Conc: 303.22 ng/ml



#37 AR-1262-2


R.T.: 8.941 min
Delta R.T.: -0.005 min
Response: 1322558
Conc: 369.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.703 min
Delta R.T.: -0.006 min
Response: 1106520
Conc: 367.65 ng/ml




#38 AR-1262-3

R.T.: 9.259 min
Delta R.T.: 0.015 min
Response: 871017
Conc: 496.46 ng/ml



#38 AR-1262-3


R.T.: 7.989 min
Delta R.T.: -0.007 min
Response: 234692
Conc: 183.21 ng/ml



#39 AR-1262-4


R.T.: 9.311 min
Delta R.T.: -0.023 min
Response: 506937
Conc: 440.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.051 min
Delta R.T.: -0.009 min
Response: 841039
Conc: 359.72 ng/ml




#40 AR-1262-5

R.T.: 9.968 min
Delta R.T.: -0.007 min
Response: 333116
Conc: 251.69 ng/ml



#40 AR-1262-5


R.T.: 8.554 min
Delta R.T.: -0.008 min
Response: 274606
Conc: 240.93 ng/ml



#41 AR-1268-1


R.T.: 9.259 min
Delta R.T.: 0.017 min
Response: 871017
Conc: 210.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.989 min
Delta R.T.: -0.007 min
Response: 234692
Conc: 66.39 ng/ml




#42 AR-1268-2

R.T.: 9.311 min
Delta R.T.: -0.025 min
Response: 506937
Conc: 129.25 ng/ml



#42 AR-1268-2


R.T.: 8.051 min
Delta R.T.: -0.010 min
Response: 841039
Conc: 249.56 ng/ml



#43 AR-1268-3


R.T.: 9.545 min
Delta R.T.: -0.006 min
Response: 26791
Conc: 8.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.261 min
Delta R.T.: -0.007 min
Response: 24282
Conc: 8.26 ng/ml




#44 AR-1268-4

R.T.: 9.968 min
Delta R.T.: -0.007 min
Response: 333116
Conc: 238.35 ng/ml



#44 AR-1268-4


R.T.: 8.554 min
Delta R.T.: -0.008 min
Response: 274606
Conc: 226.17 ng/ml



#45 AR-1268-5

R.T.: 10.362 min
Delta R.T.: -0.007 min
Response: 108411
Conc: 10.84 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.827 min
Delta R.T.: -0.008 min
Response: 98202
Conc: 10.93 ng/ml