

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102821\
 Data File : PP040517.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Oct 2021 17:20
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 17:53:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 28 08:58:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.749	3.772	2014126	1170040	49.450	42.949
2)	SA Decachlor...	10.677	9.055	1392828	1229579	45.974	43.826
Target Compounds							
3)	L1 AR-1016-1	6.070	5.023	694687	431112	542.255	436.816
4)	L1 AR-1016-2	6.093	5.042	1058540	618583	529.166	424.797
5)	L1 AR-1016-3	6.159	5.233	681517	348023	542.678	434.457
6)	L1 AR-1016-4	6.266	5.286	560257	285464	566.720	430.025
7)	L1 AR-1016-5	6.581	5.514	546102	350754	563.245	476.028
8)	L2 AR-1221-1	4.992	4.028	21453	47119	44.265	142.709 #
9)	L2 AR-1221-2	5.098	4.128	98238	67456	289.422	258.648
10)	L2 AR-1221-3	5.185	4.215	407037	280206	378.879	329.490
11)	L3 AR-1232-1	5.185	4.215	407037	280206	458.988	396.776
12)	L3 AR-1232-2	5.773	5.042	516247	618583	1165.171	1034.271
13)	L3 AR-1232-3	6.093	5.233	1058540	348023	1284.830	1120.193
14)	L3 AR-1232-4	6.266	5.330	560257	345953	1423.017	1350.104
15)	L3 AR-1232-5	6.365	5.514	437186	350754	1643.893	1304.123
16)	L4 AR-1242-1	6.070	5.023	694687	431112	717.869	593.864
17)	L4 AR-1242-2	6.093	5.042	1058540	618583	704.533	582.999
18)	L4 AR-1242-3	6.159	5.233	681517	348023	722.581	598.430
19)	L4 AR-1242-4	6.266	5.330	560257	345953	742.845	674.395
20)	L4 AR-1242-5	7.050	5.894	86743	258101	109.554	380.308 #
21)	L5 AR-1248-1	6.070	5.023	694687	431112	929.946	823.943
22)	L5 AR-1248-2	6.365	5.286	437186	285464	422.312	358.184
23)	L5 AR-1248-3	6.581	5.330	546102	345953	424.464	465.784
24)	L5 AR-1248-4	7.011	5.514	104176	350754	70.882	419.744 #
25)	L5 AR-1248-5	7.050	5.937	86743	45769	60.655	46.557
26)	L6 AR-1254-1	6.979	5.894	354586	258101	240.348	182.828
27)	L6 AR-1254-2	7.210	6.055	403620	237551	179.302	191.153
28)	L6 AR-1254-3	7.591	6.494	248485	451110	103.108	237.047 #
29)	L6 AR-1254-4	7.887	6.717	161637	66402	94.109	57.422 #
30)	L6 AR-1254-5	8.316	7.146	1199030	797284	659.123	460.974 #
31)	L7 AR-1260-1	7.758	6.613	872914	605593	559.359	423.617
32)	L7 AR-1260-2	8.025	6.812	1025458	715807	556.427	427.624
33)	L7 AR-1260-3	8.391	6.966	781283	667905	563.350	422.361 #
34)	L7 AR-1260-4	8.621	7.451	883974	587934	559.862	434.973
35)	L7 AR-1260-5	8.938	7.701	1616104	1326706	548.309	441.879
36)	L8 AR-1262-1	8.391	7.188	781283	637723	369.656	367.925
37)	L8 AR-1262-2	8.938	7.701	1616104	1326706	451.178	440.809
38)	L8 AR-1262-3	9.256	7.988	1090072	332344	621.321	259.437 #
39)	L8 AR-1262-4	9.324	8.051	317932	1029611	276.195	440.372 #
40)	L8 AR-1262-5	9.963	8.552	490975	400274	370.968	351.191
41)	L9 AR-1268-1	9.256	7.988	1090072	332344	263.975	94.010 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102821\
 Data File : PP040517.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Oct 2021 17:20
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 17:53:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 28 08:58:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.324	8.051	317932	1029611	81.063	305.519 #
43) L9	AR-1268-3	9.540	8.258	27369	18451	8.368	6.279
44) L9	AR-1268-4	9.963	8.552	490975	400274	351.307	329.673
45) L9	AR-1268-5	10.358	8.825	132726	116195	13.270	12.937



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102821\
Data File : PP040517.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Oct 2021 17:20
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 17:53:20 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 28 08:58:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.749 min
Delta R.T.: -0.002 min
Response: 2014126
Conc: 49.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.772 min
Delta R.T.: -0.001 min
Response: 1170040
Conc: 42.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.677 min
Delta R.T.: -0.014 min
Response: 1392828
Conc: 45.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.055 min
Delta R.T.: -0.011 min
Response: 1229579
Conc: 43.83 ng/ml



#3 AR-1016-1


R.T.: 6.070 min
Delta R.T.: -0.004 min
Response: 694687
Conc: 542.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.023 min
Delta R.T.: -0.004 min
Response: 431112
Conc: 436.82 ng/ml




#4 AR-1016-2

R.T.: 6.093 min
Delta R.T.: -0.005 min
Response: 1058540
Conc: 529.17 ng/ml



#4 AR-1016-2


R.T.: 5.042 min
Delta R.T.: -0.004 min
Response: 618583
Conc: 424.80 ng/ml



#5 AR-1016-3


R.T.: 6.159 min
Delta R.T.: -0.004 min
Response: 681517
Conc: 542.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.233 min
Delta R.T.: -0.005 min
Response: 348023
Conc: 434.46 ng/ml




#6 AR-1016-4

R.T.: 6.266 min
Delta R.T.: -0.004 min
Response: 560257
Conc: 566.72 ng/ml



#6 AR-1016-4


R.T.: 5.286 min
Delta R.T.: -0.004 min
Response: 285464
Conc: 430.03 ng/ml



#7 AR-1016-5


R.T.: 6.581 min
Delta R.T.: -0.005 min
Response: 546102
Conc: 563.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.514 min
Delta R.T.: -0.003 min
Response: 350754
Conc: 476.03 ng/ml




#8 AR-1221-1

R.T.: 4.992 min
Delta R.T.: -0.012 min
Response: 21453
Conc: 44.26 ng/ml



#8 AR-1221-1


R.T.: 4.028 min
Delta R.T.: -0.003 min
Response: 47119
Conc: 142.71 ng/ml



#9 AR-1221-2


R.T.: 5.098 min
Delta R.T.: -0.003 min
Response: 98238
Conc: 289.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.128 min
Delta R.T.: -0.003 min
Response: 67456
Conc: 258.65 ng/ml




#10 AR-1221-3

R.T.: 5.185 min
Delta R.T.: -0.003 min
Response: 407037
Conc: 378.88 ng/ml



#10 AR-1221-3


R.T.: 4.215 min
Delta R.T.: -0.003 min
Response: 280206
Conc: 329.49 ng/ml



#11 AR-1232-1


R.T.: 5.185 min
Delta R.T.: -0.004 min
Response: 407037
Conc: 458.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.215 min
Delta R.T.: -0.003 min
Response: 280206
Conc: 396.78 ng/ml




#12 AR-1232-2

R.T.: 5.773 min
Delta R.T.: -0.004 min
Response: 516247
Conc: 1165.17 ng/ml



#12 AR-1232-2


R.T.: 5.042 min
Delta R.T.: -0.004 min
Response: 618583
Conc: 1034.27 ng/ml



#13 AR-1232-3


R.T.: 6.093 min
Delta R.T.: -0.005 min
Response: 1058540
Conc: 1284.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.233 min
Delta R.T.: -0.004 min
Response: 348023
Conc: 1120.19 ng/ml




#14 AR-1232-4

R.T.: 6.266 min
Delta R.T.: -0.004 min
Response: 560257
Conc: 1423.02 ng/ml



#14 AR-1232-4


R.T.: 5.330 min
Delta R.T.: -0.005 min
Response: 345953
Conc: 1350.10 ng/ml



#15 AR-1232-5


R.T.: 6.365 min
Delta R.T.: -0.004 min
Response: 437186
Conc: 1643.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.514 min
Delta R.T.: -0.004 min
Response: 350754
Conc: 1304.12 ng/ml




#16 AR-1242-1

R.T.: 6.070 min
Delta R.T.: -0.005 min
Response: 694687
Conc: 717.87 ng/ml



#16 AR-1242-1


R.T.: 5.023 min
Delta R.T.: -0.004 min
Response: 431112
Conc: 593.86 ng/ml



#17 AR-1242-2


R.T.: 6.093 min
Delta R.T.: -0.006 min
Response: 1058540
Conc: 704.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.042 min
Delta R.T.: -0.004 min
Response: 618583
Conc: 583.00 ng/ml




#18 AR-1242-3

R.T.: 6.159 min
Delta R.T.: -0.005 min
Response: 681517
Conc: 722.58 ng/ml



#18 AR-1242-3


R.T.: 5.233 min
Delta R.T.: -0.005 min
Response: 348023
Conc: 598.43 ng/ml



#19 AR-1242-4


R.T.: 6.266 min
Delta R.T.: -0.005 min
Response: 560257
Conc: 742.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.330 min
Delta R.T.: -0.006 min
Response: 345953
Conc: 674.40 ng/ml




#20 AR-1242-5

R.T.: 7.050 min
Delta R.T.: -0.006 min
Response: 86743
Conc: 109.55 ng/ml



#20 AR-1242-5


R.T.: 5.894 min
Delta R.T.: -0.006 min
Response: 258101
Conc: 380.31 ng/ml



#21 AR-1248-1


R.T.: 6.070 min
Delta R.T.: -0.005 min
Response: 694687
Conc: 929.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.023 min
Delta R.T.: -0.004 min
Response: 431112
Conc: 823.94 ng/ml




#22 AR-1248-2

R.T.: 6.365 min
Delta R.T.: -0.004 min
Response: 437186
Conc: 422.31 ng/ml



#22 AR-1248-2


R.T.: 5.286 min
Delta R.T.: -0.004 min
Response: 285464
Conc: 358.18 ng/ml



#23 AR-1248-3


R.T.: 6.581 min
Delta R.T.: -0.004 min
Response: 546102
Conc: 424.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.330 min
Delta R.T.: -0.006 min
Response: 345953
Conc: 465.78 ng/ml




#24 AR-1248-4

R.T.: 7.011 min
Delta R.T.: -0.005 min
Response: 104176
Conc: 70.88 ng/ml



#24 AR-1248-4


R.T.: 5.514 min
Delta R.T.: -0.003 min
Response: 350754
Conc: 419.74 ng/ml



#25 AR-1248-5


R.T.: 7.050 min
Delta R.T.: -0.005 min
Response: 86743
Conc: 60.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.937 min
Delta R.T.: -0.005 min
Response: 45769
Conc: 46.56 ng/ml




#26 AR-1254-1

R.T.: 6.979 min
Delta R.T.: -0.005 min
Response: 354586
Conc: 240.35 ng/ml



#26 AR-1254-1


R.T.: 5.894 min
Delta R.T.: -0.005 min
Response: 258101
Conc: 182.83 ng/ml



#27 AR-1254-2


R.T.: 7.210 min
Delta R.T.: -0.004 min
Response: 403620
Conc: 179.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.055 min
Delta R.T.: -0.005 min
Response: 237551
Conc: 191.15 ng/ml




#28 AR-1254-3

R.T.: 7.591 min
Delta R.T.: -0.005 min
Response: 248485
Conc: 103.11 ng/ml



#28 AR-1254-3


R.T.: 6.494 min
Delta R.T.: 0.012 min
Response: 451110
Conc: 237.05 ng/ml



#29 AR-1254-4


R.T.: 7.887 min
Delta R.T.: -0.008 min
Response: 161637
Conc: 94.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.717 min
Delta R.T.: -0.006 min
Response: 66402
Conc: 57.42 ng/ml




#30 AR-1254-5

R.T.: 8.316 min
Delta R.T.: -0.006 min
Response: 1199030
Conc: 659.12 ng/ml



#30 AR-1254-5


R.T.: 7.146 min
Delta R.T.: -0.007 min
Response: 797284
Conc: 460.97 ng/ml



#31 AR-1260-1


R.T.: 7.758 min
Delta R.T.: -0.006 min
Response: 872914
Conc: 559.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.613 min
Delta R.T.: -0.007 min
Response: 605593
Conc: 423.62 ng/ml




#32 AR-1260-2

R.T.: 8.025 min
Delta R.T.: -0.007 min
Response: 1025458
Conc: 556.43 ng/ml



#32 AR-1260-2


R.T.: 6.812 min
Delta R.T.: -0.007 min
Response: 715807
Conc: 427.62 ng/ml



#33 AR-1260-3


R.T.: 8.391 min
Delta R.T.: -0.008 min
Response: 781283
Conc: 563.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.966 min
Delta R.T.: -0.006 min
Response: 667905
Conc: 422.36 ng/ml




#34 AR-1260-4

R.T.: 8.621 min
Delta R.T.: -0.007 min
Response: 883974
Conc: 559.86 ng/ml



#34 AR-1260-4


R.T.: 7.451 min
Delta R.T.: -0.008 min
Response: 587934
Conc: 434.97 ng/ml



#35 AR-1260-5


R.T.: 8.938 min
Delta R.T.: -0.008 min
Response: 1616104
Conc: 548.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.701 min
Delta R.T.: -0.009 min
Response: 1326706
Conc: 441.88 ng/ml




#36 AR-1262-1

R.T.: 8.391 min
Delta R.T.: -0.007 min
Response: 781283
Conc: 369.66 ng/ml



#36 AR-1262-1


R.T.: 7.188 min
Delta R.T.: -0.007 min
Response: 637723
Conc: 367.93 ng/ml



#37 AR-1262-2


R.T.: 8.938 min
Delta R.T.: -0.008 min
Response: 1616104
Conc: 451.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.701 min
Delta R.T.: -0.008 min
Response: 1326706
Conc: 440.81 ng/ml




#38 AR-1262-3

R.T.: 9.256 min
Delta R.T.: 0.012 min
Response: 1090072
Conc: 621.32 ng/ml



#38 AR-1262-3


R.T.: 7.988 min
Delta R.T.: -0.008 min
Response: 332344
Conc: 259.44 ng/ml



#39 AR-1262-4


R.T.: 9.324 min
Delta R.T.: -0.010 min
Response: 317932
Conc: 276.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.051 min
Delta R.T.: -0.010 min
Response: 1029611
Conc: 440.37 ng/ml




#40 AR-1262-5

R.T.: 9.963 min
Delta R.T.: -0.012 min
Response: 490975
Conc: 370.97 ng/ml



#40 AR-1262-5


R.T.: 8.552 min
Delta R.T.: -0.010 min
Response: 400274
Conc: 351.19 ng/ml



#41 AR-1268-1


R.T.: 9.256 min
Delta R.T.: 0.013 min
Response: 1090072
Conc: 263.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.988 min
Delta R.T.: -0.008 min
Response: 332344
Conc: 94.01 ng/ml




#42 AR-1268-2

R.T.: 9.324 min
Delta R.T.: -0.012 min
Response: 317932
Conc: 81.06 ng/ml



#42 AR-1268-2


R.T.: 8.051 min
Delta R.T.: -0.011 min
Response: 1029611
Conc: 305.52 ng/ml



#43 AR-1268-3


R.T.: 9.540 min
Delta R.T.: -0.011 min
Response: 27369
Conc: 8.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.258 min
Delta R.T.: -0.010 min
Response: 18451
Conc: 6.28 ng/ml




#44 AR-1268-4

R.T.: 9.963 min
Delta R.T.: -0.011 min
Response: 490975
Conc: 351.31 ng/ml



#44 AR-1268-4


R.T.: 8.552 min
Delta R.T.: -0.009 min
Response: 400274
Conc: 329.67 ng/ml



#45 AR-1268-5

R.T.: 10.358 min
Delta R.T.: -0.011 min
Response: 132726
Conc: 13.27 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.825 min
Delta R.T.: -0.010 min
Response: 116195
Conc: 12.94 ng/ml