

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102821\
 Data File : PP040530.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Oct 2021 21:25
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 00:38:05 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 28 08:58:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.748	3.770	2116664	1203003	51.968	44.159
2)	SA Decachlor...	10.678	9.054	1413301	1274115	46.650	45.414
Target Compounds							
3)	L1 AR-1016-1	6.070	5.022	703942	438666	549.479	444.470
4)	L1 AR-1016-2	6.094	5.041	1068258	634037	534.024	435.410
5)	L1 AR-1016-3	6.159	5.232	681702	357882	542.826	446.765
6)	L1 AR-1016-4	6.266	5.286	563341	292059	569.840	439.961
7)	L1 AR-1016-5	6.581	5.514	562630	329577	580.291	447.287
31)	L7 AR-1260-1	7.760	6.612	896049	622501	574.184	435.444
32)	L7 AR-1260-2	8.026	6.811	1053014	737693	571.379	440.699
33)	L7 AR-1260-3	8.392	6.965	797742	696123	575.218	440.205
34)	L7 AR-1260-4	8.622	7.450	903406	607416	572.169	449.386
35)	L7 AR-1260-5	8.938	7.700	1654522	1372684	561.343	457.192



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102821\
Data File : PP040530.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Oct 2021 21:25
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 00:38:05 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 28 08:58:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.748 min
Delta R.T.: -0.003 min
Response: 2116664
Conc: 51.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.770 min
Delta R.T.: -0.002 min
Response: 1203003
Conc: 44.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.678 min
Delta R.T.: -0.014 min
Response: 1413301
Conc: 46.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.054 min
Delta R.T.: -0.013 min
Response: 1274115
Conc: 45.41 ng/ml



#3 AR-1016-1


R.T.: 6.070 min
Delta R.T.: -0.004 min
Response: 703942
Conc: 549.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.022 min
Delta R.T.: -0.004 min
Response: 438666
Conc: 444.47 ng/ml




#4 AR-1016-2

R.T.: 6.094 min
Delta R.T.: -0.004 min
Response: 1068258
Conc: 534.02 ng/ml



#4 AR-1016-2


R.T.: 5.041 min
Delta R.T.: -0.004 min
Response: 634037
Conc: 435.41 ng/ml



#5 AR-1016-3


R.T.: 6.159 min
Delta R.T.: -0.004 min
Response: 681702
Conc: 542.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.232 min
Delta R.T.: -0.006 min
Response: 357882
Conc: 446.76 ng/ml




#6 AR-1016-4

R.T.: 6.266 min
Delta R.T.: -0.003 min
Response: 563341
Conc: 569.84 ng/ml



#6 AR-1016-4


R.T.: 5.286 min
Delta R.T.: -0.005 min
Response: 292059
Conc: 439.96 ng/ml



#7 AR-1016-5


R.T.: 6.581 min
Delta R.T.: -0.005 min
Response: 562630
Conc: 580.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.514 min
Delta R.T.: -0.004 min
Response: 329577
Conc: 447.29 ng/ml




#31 AR-1260-1

R.T.: 7.760 min
Delta R.T.: -0.004 min
Response: 896049
Conc: 574.18 ng/ml



#31 AR-1260-1


R.T.: 6.612 min
Delta R.T.: -0.007 min
Response: 622501
Conc: 435.44 ng/ml



#32 AR-1260-2


R.T.: 8.026 min
Delta R.T.: -0.007 min
Response: 1053014
Conc: 571.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.811 min
Delta R.T.: -0.008 min
Response: 737693
Conc: 440.70 ng/ml




#33 AR-1260-3

R.T.: 8.392 min
Delta R.T.: -0.007 min
Response: 797742
Conc: 575.22 ng/ml



#33 AR-1260-3


R.T.: 6.965 min
Delta R.T.: -0.007 min
Response: 696123
Conc: 440.21 ng/ml



#34 AR-1260-4


R.T.: 8.622 min
Delta R.T.: -0.007 min
Response: 903406
Conc: 572.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.450 min
Delta R.T.: -0.009 min
Response: 607416
Conc: 449.39 ng/ml



#35 AR-1260-5

R.T.: 8.938 min
Delta R.T.: -0.008 min
Response: 1654522
Conc: 561.34 ng/ml



#35 AR-1260-5

R.T.: 7.700 min
Delta R.T.: -0.010 min
Response: 1372684
Conc: 457.19 ng/ml