

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102920\
 Data File : PP030930.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Oct 2020 19:37
 Operator : DD\AJ
 Sample : MDL-AR1248-S-2
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MDL-AR1248-S-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 10:05:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\AR1248PP102920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 28 16:58:01 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.803	4.081	1125202	985364	18.873	18.269
2) SA Decachlor...	10.690	9.416	663603	711067	19.510	19.517
Target Compounds						
21) L5 AR-1248-1	6.109	5.345	26675	19119	26.720	22.476
22) L5 AR-1248-2	6.403	5.609	33031	24719	26.368	21.386
23) L5 AR-1248-3	6.617	5.655	39101	24179	25.506	19.759
24) L5 AR-1248-4	7.044	5.839	42739	32051	26.707	21.753
25) L5 AR-1248-5	7.083	6.264	41430	27917	25.923	20.903



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102920\
Data File : PP030930.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Oct 2020 19:37
Operator : DD\AJ
Sample : MDL-AR1248-S-2
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
MDL-AR1248-S-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 10:05:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\AR1248PP102920.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 28 16:58:01 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.803 min
Delta R.T.: 0.000 min
Response: 1125202
Conc: 18.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
MDL-AR1248-S-2




#1 Tetrachloro-m-xylene

R.T.: 4.081 min
Delta R.T.: 0.000 min
Response: 985364
Conc: 18.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.690 min
Delta R.T.: 0.000 min
Response: 663603
Conc: 19.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.416 min
Delta R.T.: 0.001 min
Response: 711067
Conc: 19.52 ng/ml



#21 AR-1248-1


R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 26675
Conc: 26.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
MDL-AR1248-S-2




#21 AR-1248-1

R.T.: 5.345 min
Delta R.T.: 0.000 min
Response: 19119
Conc: 22.48 ng/ml




#22 AR-1248-2

R.T.: 6.403 min
Delta R.T.: 0.000 min
Response: 33031
Conc: 26.37 ng/ml



#22 AR-1248-2


R.T.: 5.609 min
Delta R.T.: 0.000 min
Response: 24719
Conc: 21.39 ng/ml



#23 AR-1248-3


R.T.: 6.617 min
Delta R.T.: -0.001 min
Response: 39101
Conc: 25.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
MDL-AR1248-S-2




#23 AR-1248-3

R.T.: 5.655 min
Delta R.T.: 0.000 min
Response: 24179
Conc: 19.76 ng/ml




#24 AR-1248-4

R.T.: 7.044 min
Delta R.T.: 0.000 min
Response: 42739
Conc: 26.71 ng/ml



#24 AR-1248-4


R.T.: 5.839 min
Delta R.T.: 0.000 min
Response: 32051
Conc: 21.75 ng/ml



#25 AR-1248-5

R.T.: 7.083 min
Delta R.T.: 0.000 min
Response: 41430
Conc: 25.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
MDL-AR1248-S-2



#25 AR-1248-5

R.T.: 6.264 min
Delta R.T.: 0.001 min
Response: 27917
Conc: 20.90 ng/ml