

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102921\
 Data File : PP040575.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Oct 2021 17:01
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 02:21:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 28 08:58:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.750	3.770	2142602	1246721	52.605	45.764
2)	SA Decachlor...	10.679	9.053	1462976	1307118	48.289	46.590
Target Compounds							
3)	L1 AR-1016-1	6.071	5.021	20302	11747	15.847	11.903
4)	L1 AR-1016-2	6.098	5.041	25766	11537	12.880	7.923 #
5)	L1 AR-1016-3	6.161	5.232	12108	8207	9.641	10.246
6)	L1 AR-1016-4	0.000	5.285	0	283495	N.D.	427.060 #
7)	L1 AR-1016-5	6.581	5.512	273380	159813	281.962	216.892
8)	L2 AR-1221-1	4.987f	0.000	26225	0	54.111	N.D. #
9)	L2 AR-1221-2	5.101	4.120	27683	16811	81.558	64.460
10)	L2 AR-1221-3	0.000	4.214	0	35490	N.D.	41.732 #
11)	L3 AR-1232-1	0.000	4.214	0	35490	N.D.	50.255 #
12)	L3 AR-1232-2	5.772	5.041	8519	11537	19.229	19.291
13)	L3 AR-1232-3	6.098	5.232	25766	8207	31.274	26.417
14)	L3 AR-1232-4	0.000	5.328	0	86156	N.D.	336.228 #
15)	L3 AR-1232-5	6.364	5.512	442802	159813	1665.012	594.195 #
16)	L4 AR-1242-1	6.071	5.021	20302	11747	20.979	16.182
17)	L4 AR-1242-2	6.098	5.041	25766	11537	17.149	10.874 #
18)	L4 AR-1242-3	6.161	5.232	12108	8207	12.838	14.113
19)	L4 AR-1242-4	0.000	5.328	0	86156	N.D.	167.951 #
20)	L4 AR-1242-5	7.052	5.893	257548	643048	325.278	947.524 #
21)	L5 AR-1248-1	6.071	5.021	20302	11747	27.177	22.452
22)	L5 AR-1248-2	6.364	5.285	442802	283495	427.738	355.714
23)	L5 AR-1248-3	6.581	5.328	273380	86156	212.488	115.998 #
24)	L5 AR-1248-4	7.012	5.512	404638	159813	275.320	191.247 #
25)	L5 AR-1248-5	7.052	5.936	257548	83787	180.092	85.229 #
26)	L6 AR-1254-1	6.980	5.893	744198	643048	504.437	455.508
27)	L6 AR-1254-2	7.210	6.053	1149845	566283	510.802	455.678
28)	L6 AR-1254-3	7.592	6.474	1241143	837820	515.008	440.253
29)	L6 AR-1254-4	7.890	6.715	910058	557110	529.856	481.764
30)	L6 AR-1254-5	8.318	7.145	964986	815400	530.466	471.448
31)	L7 AR-1260-1	7.760	6.612	497417	456939	318.742	319.633
32)	L7 AR-1260-2	8.027	6.811	565915	380264	307.073	227.170 #
33)	L7 AR-1260-3	8.379f	6.963	137220	379241	98.944	239.819 #
34)	L7 AR-1260-4	8.615	7.449	385034	42280	243.860	31.280 #
35)	L7 AR-1260-5	8.939	7.699	129654	97779	43.989	32.567 #
36)	L8 AR-1262-1	8.379f	7.214f	137220	66633	64.924	38.443 #
37)	L8 AR-1262-2	8.939	7.699	129654	97779	36.196	32.488
38)	L8 AR-1262-3	9.259	7.993	107473	10431	61.258	8.143 #
39)	L8 AR-1262-4	9.304f	8.047	27375	112590	23.781	48.155 #
40)	L8 AR-1262-5	0.000	8.550	0	4512	N.D.	3.958 #
41)	L9 AR-1268-1	9.259f	7.993	107473	10431	26.026	2.951 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102921\
 Data File : PP040575.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Oct 2021 17:01
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 02:21:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 28 08:58:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	0.000	8.047	0	112590	N.D.	33.409 #
43) L9	AR-1268-3	0.000	8.260	0	18671	N.D.	6.354 #
44) L9	AR-1268-4	0.000	8.550	0	4512	N.D.	3.716 #
45) L9	AR-1268-5	0.000	8.825	0	9517	N.D.	1.060 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102921\
Data File : PP040575.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Oct 2021 17:01
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:21:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 28 08:58:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.750 min
Delta R.T.: -0.001 min
Response: 2142602
Conc: 52.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.770 min
Delta R.T.: -0.002 min
Response: 1246721
Conc: 45.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.679 min
Delta R.T.: -0.012 min
Response: 1462976
Conc: 48.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.053 min
Delta R.T.: -0.013 min
Response: 1307118
Conc: 46.59 ng/ml



#3 AR-1016-1


R.T.: 6.071 min
Delta R.T.: -0.003 min
Response: 20302
Conc: 15.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.021 min
Delta R.T.: -0.005 min
Response: 11747
Conc: 11.90 ng/ml




#4 AR-1016-2

R.T.: 6.098 min
Delta R.T.: 0.000 min
Response: 25766
Conc: 12.88 ng/ml



#4 AR-1016-2


R.T.: 5.041 min
Delta R.T.: -0.005 min
Response: 11537
Conc: 7.92 ng/ml



#5 AR-1016-3


R.T.: 6.161 min
Delta R.T.: -0.002 min
Response: 12108
Conc: 9.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.232 min
Delta R.T.: -0.006 min
Response: 8207
Conc: 10.25 ng/ml




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T. : 6.270 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 5.285 min
Delta R.T.: -0.006 min
Response: 283495
Conc: 427.06 ng/ml



#7 AR-1016-5


R.T.: 6.581 min
Delta R.T.: -0.005 min
Response: 273380
Conc: 281.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.512 min
Delta R.T.: -0.005 min
Response: 159813
Conc: 216.89 ng/ml




#8 AR-1221-1

R.T.: 4.987 min
Delta R.T.: -0.016 min
Response: 26225
Conc: 54.11 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 4.032 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 5.101 min
Delta R.T.: -0.001 min
Response: 27683
Conc: 81.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.120 min
Delta R.T.: -0.011 min
Response: 16811
Conc: 64.46 ng/ml




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 5.189 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 4.214 min
Delta R.T.: -0.004 min
Response: 35490
Conc: 41.73 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T.: 5.189 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.214 min
Delta R.T.: -0.004 min
Response: 35490
Conc: 50.25 ng/ml




#12 AR-1232-2

R.T.: 5.772 min
Delta R.T.: -0.005 min
Response: 8519
Conc: 19.23 ng/ml



#12 AR-1232-2


R.T.: 5.041 min
Delta R.T.: -0.004 min
Response: 11537
Conc: 19.29 ng/ml



#13 AR-1232-3


R.T.: 6.098 min
Delta R.T.: 0.000 min
Response: 25766
Conc: 31.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.232 min
Delta R.T.: -0.005 min
Response: 8207
Conc: 26.42 ng/ml




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T. : 6.270 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 5.328 min
Delta R.T.: -0.007 min
Response: 86156
Conc: 336.23 ng/ml



#15 AR-1232-5


R.T.: 6.364 min
Delta R.T.: -0.004 min
Response: 442802
Conc: 1665.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.512 min
Delta R.T.: -0.006 min
Response: 159813
Conc: 594.19 ng/ml




#16 AR-1242-1

R.T.: 6.071 min
Delta R.T.: -0.003 min
Response: 20302
Conc: 20.98 ng/ml



#16 AR-1242-1


R.T.: 5.021 min
Delta R.T.: -0.006 min
Response: 11747
Conc: 16.18 ng/ml



#17 AR-1242-2


R.T.: 6.098 min
Delta R.T.: -0.001 min
Response: 25766
Conc: 17.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.041 min
Delta R.T.: -0.005 min
Response: 11537
Conc: 10.87 ng/ml




#18 AR-1242-3

R.T.: 6.161 min
Delta R.T.: -0.003 min
Response: 12108
Conc: 12.84 ng/ml



#18 AR-1242-3


R.T.: 5.232 min
Delta R.T.: -0.006 min
Response: 8207
Conc: 14.11 ng/ml



#19 AR-1242-4


R.T.: 0.000 min
Exp R.T.: 6.271 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.328 min
Delta R.T.: -0.009 min
Response: 86156
Conc: 167.95 ng/ml




#20 AR-1242-5

R.T.: 7.052 min
Delta R.T.: -0.004 min
Response: 257548
Conc: 325.28 ng/ml



#20 AR-1242-5


R.T.: 5.893 min
Delta R.T.: -0.007 min
Response: 643048
Conc: 947.52 ng/ml



#21 AR-1248-1


R.T.: 6.071 min
Delta R.T.: -0.003 min
Response: 20302
Conc: 27.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.021 min
Delta R.T.: -0.005 min
Response: 11747
Conc: 22.45 ng/ml




#22 AR-1248-2

R.T.: 6.364 min
Delta R.T.: -0.004 min
Response: 442802
Conc: 427.74 ng/ml



#22 AR-1248-2


R.T.: 5.285 min
Delta R.T.: -0.006 min
Response: 283495
Conc: 355.71 ng/ml



#23 AR-1248-3


R.T.: 6.581 min
Delta R.T.: -0.004 min
Response: 273380
Conc: 212.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.328 min
Delta R.T.: -0.009 min
Response: 86156
Conc: 116.00 ng/ml




#24 AR-1248-4

R.T.: 7.012 min
Delta R.T.: -0.003 min
Response: 404638
Conc: 275.32 ng/ml



#24 AR-1248-4


R.T.: 5.512 min
Delta R.T.: -0.005 min
Response: 159813
Conc: 191.25 ng/ml



#25 AR-1248-5


R.T.: 7.052 min
Delta R.T.: -0.003 min
Response: 257548
Conc: 180.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.936 min
Delta R.T.: -0.006 min
Response: 83787
Conc: 85.23 ng/ml




#26 AR-1254-1

R.T.: 6.980 min
Delta R.T.: -0.004 min
Response: 744198
Conc: 504.44 ng/ml



#26 AR-1254-1


R.T.: 5.893 min
Delta R.T.: -0.006 min
Response: 643048
Conc: 455.51 ng/ml



#27 AR-1254-2


R.T.: 7.210 min
Delta R.T.: -0.003 min
Response: 1149845
Conc: 510.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.053 min
Delta R.T.: -0.007 min
Response: 566283
Conc: 455.68 ng/ml




#28 AR-1254-3

R.T.: 7.592 min
Delta R.T.: -0.004 min
Response: 1241143
Conc: 515.01 ng/ml



#28 AR-1254-3


R.T.: 6.474 min
Delta R.T.: -0.007 min
Response: 837820
Conc: 440.25 ng/ml



#29 AR-1254-4


R.T.: 7.890 min
Delta R.T.: -0.005 min
Response: 910058
Conc: 529.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.715 min
Delta R.T.: -0.008 min
Response: 557110
Conc: 481.76 ng/ml




#30 AR-1254-5

R.T.: 8.318 min
Delta R.T.: -0.004 min
Response: 964986
Conc: 530.47 ng/ml



#30 AR-1254-5


R.T.: 7.145 min
Delta R.T.: -0.008 min
Response: 815400
Conc: 471.45 ng/ml



#31 AR-1260-1


R.T.: 7.760 min
Delta R.T.: -0.004 min
Response: 497417
Conc: 318.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.612 min
Delta R.T.: -0.007 min
Response: 456939
Conc: 319.63 ng/ml




#32 AR-1260-2

R.T.: 8.027 min
Delta R.T.: -0.006 min
Response: 565915
Conc: 307.07 ng/ml



#32 AR-1260-2


R.T.: 6.811 min
Delta R.T.: -0.008 min
Response: 380264
Conc: 227.17 ng/ml



#33 AR-1260-3


R.T.: 8.379 min
Delta R.T.: -0.020 min
Response: 137220
Conc: 98.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.963 min
Delta R.T.: -0.010 min
Response: 379241
Conc: 239.82 ng/ml




#34 AR-1260-4

R.T.: 8.615 min
Delta R.T.: -0.014 min
Response: 385034
Conc: 243.86 ng/ml



#34 AR-1260-4


R.T.: 7.449 min
Delta R.T.: -0.009 min
Response: 42280
Conc: 31.28 ng/ml



#35 AR-1260-5


R.T.: 8.939 min
Delta R.T.: -0.007 min
Response: 129654
Conc: 43.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.699 min
Delta R.T.: -0.011 min
Response: 97779
Conc: 32.57 ng/ml




#36 AR-1262-1

R.T.: 8.379 min
Delta R.T.: -0.018 min
Response: 137220
Conc: 64.92 ng/ml



#36 AR-1262-1


R.T.: 7.214 min
Delta R.T.: 0.019 min
Response: 66633
Conc: 38.44 ng/ml



#37 AR-1262-2


R.T.: 8.939 min
Delta R.T.: -0.007 min
Response: 129654
Conc: 36.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.699 min
Delta R.T.: -0.010 min
Response: 97779
Conc: 32.49 ng/ml




#38 AR-1262-3

R.T.: 9.259 min
Delta R.T.: 0.015 min
Response: 107473
Conc: 61.26 ng/ml



#38 AR-1262-3


R.T.: 7.993 min
Delta R.T.: -0.003 min
Response: 10431
Conc: 8.14 ng/ml



#39 AR-1262-4


R.T.: 9.304 min
Delta R.T.: -0.029 min
Response: 27375
Conc: 23.78 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.047 min
Delta R.T.: -0.014 min
Response: 112590
Conc: 48.16 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 9.975 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.550 min
Delta R.T.: -0.011 min
Response: 4512
Conc: 3.96 ng/ml



#41 AR-1268-1


R.T.: 9.259 min
Delta R.T.: 0.016 min
Response: 107473
Conc: 26.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.993 min
Delta R.T.: -0.003 min
Response: 10431
Conc: 2.95 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 9.336 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 8.047 min
Delta R.T.: -0.015 min
Response: 112590
Conc: 33.41 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.551 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.260 min
Delta R.T.: -0.008 min
Response: 18671
Conc: 6.35 ng/ml




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 9.974 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 8.550 min
Delta R.T.: -0.011 min
Response: 4512
Conc: 3.72 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.369 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.825 min
Delta R.T.: -0.010 min
Response: 9517
Conc: 1.06 ng/ml