

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102921\
 Data File : PP040553.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Oct 2021 10:21
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 02:14:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 28 08:58:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.750	3.771	2095910	1233022	51.458	45.261
2)	SA Decachlor...	10.681	9.055	1425473	1272711	47.051	45.364
Target Compounds							
26)	L6 AR-1254-1	6.980	5.894	734399	641268	497.796	454.247
27)	L6 AR-1254-2	7.211	6.055	1128618	564442	501.372	454.197
28)	L6 AR-1254-3	7.594	6.475	1212493	842123	503.120	442.514
29)	L6 AR-1254-4	7.890	6.716	873137	556466	508.360	481.208
30)	L6 AR-1254-5	8.319	7.146	935123	818149	514.049	473.038



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102921\
Data File : PP040553.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Oct 2021 10:21
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:14:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 28 08:58:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.750 min
Delta R.T.: -0.001 min
Response: 2095910
Conc: 51.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.771 min
Delta R.T.: -0.001 min
Response: 1233022
Conc: 45.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.681 min
Delta R.T.: -0.010 min
Response: 1425473
Conc: 47.05 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.055 min
Delta R.T.: -0.012 min
Response: 1272711
Conc: 45.36 ng/ml



#26 AR-1254-1


R.T.: 6.980 min
Delta R.T.: -0.003 min
Response: 734399
Conc: 497.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.894 min
Delta R.T.: -0.005 min
Response: 641268
Conc: 454.25 ng/ml




#27 AR-1254-2

R.T.: 7.211 min
Delta R.T.: -0.003 min
Response: 1128618
Conc: 501.37 ng/ml



#27 AR-1254-2


R.T.: 6.055 min
Delta R.T.: -0.006 min
Response: 564442
Conc: 454.20 ng/ml



#28 AR-1254-3


R.T.: 7.594 min
Delta R.T.: -0.003 min
Response: 1212493
Conc: 503.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.475 min
Delta R.T.: -0.006 min
Response: 842123
Conc: 442.51 ng/ml




#29 AR-1254-4

R.T.: 7.890 min
Delta R.T.: -0.004 min
Response: 873137
Conc: 508.36 ng/ml



#29 AR-1254-4


R.T.: 6.716 min
Delta R.T.: -0.007 min
Response: 556466
Conc: 481.21 ng/ml



#30 AR-1254-5

R.T.: 8.319 min
Delta R.T.: -0.003 min
Response: 935123
Conc: 514.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.146 min
Delta R.T.: -0.008 min
Response: 818149
Conc: 473.04 ng/ml