

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102925\
 Data File : PP076157.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2025 12:33
 Operator : YP\AJ
 Sample : Q3480-01MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 RT3449MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/30/2025
 Supervised By :Sohil Jodhani 10/30/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 01:23:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.642	3.781	35157930	156.9E6	24.602	23.707
2) SA Decachlor...	10.410	8.776	24486121	149.0E6	22.745	18.118
Target Compounds						
3) L1 AR-1016-1	5.793	4.875	28404408	348.3E6	480.404	470.527
4) L1 AR-1016-2	5.815	4.934	43008392	157.7E6	486.278	461.098
5) L1 AR-1016-3	5.877	5.053	26685512	91371929	482.114	444.973
6) L1 AR-1016-4	5.974	5.093	24000943	92207420	528.187	493.390
7) L1 AR-1016-5	6.266	5.306	20390550	103.7E6	469.769	454.907m
31) L7 AR-1260-1	7.382	6.335	34190870	255.1E6	497.274m	465.765
32) L7 AR-1260-2	7.637	6.521	41746235	375.0E6	466.036	460.820
33) L7 AR-1260-3	7.994	6.674	27821083	263.8E6	401.537	468.713
34) L7 AR-1260-4	8.220	7.143	32509626	230.2E6	474.099m	443.244
35) L7 AR-1260-5	8.547	7.383	59247424	535.0E6	400.451	394.672

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102925\
Data File : PP076157.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2025 12:33
Operator : YP\AJ
Sample : Q3480-01MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
RT3449MSD

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 10/30/2025
Supervised By : Sohil Jodhani 10/30/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 01:23:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

