

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103020\
 Data File : PP030964.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Oct 2020 13:22
 Operator : DD\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 17:11:05 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\AR1248PP102920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 28 16:58:01 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.802	4.080	3132634	2845240	52.544	52.752
2) SA Decachlor...	10.693	9.416	1781655	1926554	52.381	52.879
Target Compounds						
21) L5 AR-1248-1	6.109	5.344	514879	440022	515.757	517.266
22) L5 AR-1248-2	6.404	5.609	633993	596334	506.107	515.939
23) L5 AR-1248-3	6.618	5.653	785840	627640	512.606	512.904
24) L5 AR-1248-4	7.044	5.839	807360	712371	504.507	483.497
25) L5 AR-1248-5	7.083	6.263	826333	697425	517.036	522.212



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103020\
Data File : PP030964.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Oct 2020 13:22
Operator : DD\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 17:11:05 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\AR1248PP102920.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 28 16:58:01 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.802 min
Delta R.T.: -0.001 min
Response: 3132634
Conc: 52.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.080 min
Delta R.T.: 0.000 min
Response: 2845240
Conc: 52.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.693 min
Delta R.T.: 0.002 min
Response: 1781655
Conc: 52.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.416 min
Delta R.T.: 0.000 min
Response: 1926554
Conc: 52.88 ng/ml



#21 AR-1248-1


R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 514879
Conc: 515.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.344 min
Delta R.T.: 0.000 min
Response: 440022
Conc: 517.27 ng/ml




#22 AR-1248-2

R.T.: 6.404 min
Delta R.T.: 0.000 min
Response: 633993
Conc: 506.11 ng/ml



#22 AR-1248-2


R.T.: 5.609 min
Delta R.T.: 0.000 min
Response: 596334
Conc: 515.94 ng/ml



#23 AR-1248-3


R.T.: 6.618 min
Delta R.T.: 0.000 min
Response: 785840
Conc: 512.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.653 min
Delta R.T.: -0.002 min
Response: 627640
Conc: 512.90 ng/ml




#24 AR-1248-4

R.T.: 7.044 min
Delta R.T.: 0.000 min
Response: 807360
Conc: 504.51 ng/ml



#24 AR-1248-4


R.T.: 5.839 min
Delta R.T.: -0.001 min
Response: 712371
Conc: 483.50 ng/ml



#25 AR-1248-5

R.T.: 7.083 min
Delta R.T.: 0.000 min
Response: 826333
Conc: 517.04 ng/ml

Instrument :
ECD_P
ClientSampleId :



#25 AR-1248-5

R.T.: 6.263 min
Delta R.T.: 0.000 min
Response: 697425
Conc: 522.21 ng/ml