

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103025\
 Data File : PP076236.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2025 06:54
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 07:12:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.783	77255383	352.2E6	54.059	53.220
2) SA Decachlor...	10.429	8.782	55835422	371.2E6	51.866	45.137
Target Compounds						
3) L1 AR-1016-1	5.800	4.877	31376634	384.9E6	530.673	519.956
4) L1 AR-1016-2	5.821	4.935	47171300	188.1E6	533.346	550.015
5) L1 AR-1016-3	5.884	5.054	29268926	112.9E6	528.787	549.882
6) L1 AR-1016-4	5.981	5.095	24361119	106.4E6	536.114	569.151
7) L1 AR-1016-5	6.274	5.311	22837564	118.6E6	526.144	520.271
31) L7 AR-1260-1	7.393	6.338	40720958	299.3E6	592.248	546.546
32) L7 AR-1260-2	7.645	6.524	47673270	438.6E6	532.203	539.017
33) L7 AR-1260-3	8.004	6.678	38895628	304.8E6	561.373	541.613
34) L7 AR-1260-4	8.233	7.146	40632107	306.7E6	592.552	590.601
35) L7 AR-1260-5	8.559	7.387	79502961	723.5E6	537.358	533.754



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103025\
Data File : PP076236.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 06:54
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 07:12:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.647 min
Delta R.T.: 0.005 min
Response: 77255383
Conc: 54.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.783 min
Delta R.T.: 0.001 min
Response: 35222247
Conc: 53.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.429 min
Delta R.T.: 0.011 min
Response: 55835422
Conc: 51.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.782 min
Delta R.T.: 0.000 min
Response: 371161868
Conc: 45.14 ng/ml



#3 AR-1016-1


R.T.: 5.800 min
Delta R.T.: 0.005 min
Response: 31376634
Conc: 530.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.877 min
Delta R.T.: -0.001 min
Response: 384857212
Conc: 519.96 ng/ml




#4 AR-1016-2

R.T.: 5.821 min
Delta R.T.: 0.005 min
Response: 47171300
Conc: 533.35 ng/ml



#4 AR-1016-2


R.T.: 4.935 min
Delta R.T.: 0.000 min
Response: 188068020
Conc: 550.02 ng/ml



#5 AR-1016-3


R.T.: 5.884 min
Delta R.T.: 0.004 min
Response: 29268926
Conc: 528.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.054 min
Delta R.T.: 0.000 min
Response: 112914234
Conc: 549.88 ng/ml




#6 AR-1016-4

R.T.: 5.981 min
Delta R.T.: 0.004 min
Response: 24361119
Conc: 536.11 ng/ml



#6 AR-1016-4


R.T.: 5.095 min
Delta R.T.: -0.002 min
Response: 106366062
Conc: 569.15 ng/ml



#7 AR-1016-5


R.T.: 6.274 min
Delta R.T.: 0.005 min
Response: 22837564
Conc: 526.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.311 min
Delta R.T.: 0.000 min
Response: 118612978
Conc: 520.27 ng/ml




#31 AR-1260-1

R.T.: 7.393 min
Delta R.T.: 0.006 min
Response: 40720958
Conc: 592.25 ng/ml



#31 AR-1260-1


R.T.: 6.338 min
Delta R.T.: 0.000 min
Response: 299340681
Conc: 546.55 ng/ml



#32 AR-1260-2


R.T.: 7.645 min
Delta R.T.: 0.005 min
Response: 47673270
Conc: 532.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.524 min
Delta R.T.: 0.000 min
Response: 438616345
Conc: 539.02 ng/ml




#33 AR-1260-3

R.T.: 8.004 min
Delta R.T.: 0.005 min
Response: 38895628
Conc: 561.37 ng/ml



#33 AR-1260-3


R.T.: 6.678 min
Delta R.T.: 0.000 min
Response: 304796586
Conc: 541.61 ng/ml



#34 AR-1260-4


R.T.: 8.233 min
Delta R.T.: 0.006 min
Response: 40632107
Conc: 592.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.146 min
Delta R.T.: 0.000 min
Response: 306717792
Conc: 590.60 ng/ml



#35 AR-1260-5

R.T.: 8.559 min
Delta R.T.: 0.006 min
Response: 79502961
Conc: 537.36 ng/ml



#35 AR-1260-5

R.T.: 7.387 min
Delta R.T.: 0.000 min
Response: 723533396
Conc: 533.75 ng/ml