

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103122\
 Data File : PP052888.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2022 15:07
 Operator : YP\AJ
 Sample : N5310-17
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 22:28:17 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 27 04:46:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.409	3.652	49939542	45079927	22.600	23.989
2)	SA Decachlor...	10.224	8.741	33284254	41927812	26.222	25.000
Target Compounds							
3)	L1 AR-1016-1	5.569f	0.000	100088	0	1.493	N.D. #
4)	L1 AR-1016-2	5.602	4.787	164218	105451	1.644	1.149 #
5)	L1 AR-1016-3	5.672	0.000	99556	0	1.636	N.D. #
6)	L1 AR-1016-4	5.763	4.994	40958	2694406	0.846	69.159 #
7)	L1 AR-1016-5	6.067	5.210	1326103	633315	27.533	12.604 #
8)	L2 AR-1221-1	4.618	0.000	4071702	0	158.793	N.D. #
9)	L2 AR-1221-2	4.717	3.940f	702243	3299417	36.916	188.478 #
10)	L2 AR-1221-3	4.791	0.000	20893	0	0.367	N.D. #
11)	L3 AR-1232-1	4.791	0.000	20893	0	0.482	N.D. #
12)	L3 AR-1232-2	5.334f	4.787	997488	105451	40.203	2.544 #
13)	L3 AR-1232-3	5.602	0.000	164218	0	3.751	N.D. #
14)	L3 AR-1232-4	5.763	5.036	40958	1672866	1.919	81.395 #
15)	L3 AR-1232-5	5.860	5.210	941247	633315	54.683	27.279 #
16)	L4 AR-1242-1	5.602f	4.787f	164218	105451	2.918	1.890 #
17)	L4 AR-1242-2	5.602	4.787	164218	105451	1.988	1.367 #
18)	L4 AR-1242-3	5.672	0.000	99556	0	1.958	N.D. #
19)	L4 AR-1242-4	5.763	5.036	40958	1672866	1.011	39.442 #
20)	L4 AR-1242-5	6.516	5.569	861665	3058344	21.613	60.503 #
21)	L5 AR-1248-1	5.602f	0.000	164218	0	4.043	N.D. #
22)	L5 AR-1248-2	5.860	4.994	941247	2694406	15.814	47.587 #
23)	L5 AR-1248-3	6.067	5.036	1326103	1672866	20.427	28.493 #
24)	L5 AR-1248-4	6.470	5.210	716883	633315	11.131	9.095
25)	L5 AR-1248-5	6.516	5.611	861665	321855	13.602	5.073 #
26)	L6 AR-1254-1	6.445	5.569	2514058	3058344	33.121	28.481
27)	L6 AR-1254-2	6.666	5.718	3371244	2753525	30.163	28.807
28)	L6 AR-1254-3	7.033	6.128	4277610	5360373	38.430	35.084
29)	L6 AR-1254-4	7.320	6.358	1972602	2439076	28.012	26.864
30)	L6 AR-1254-5	7.741	6.781	3551715	4785931	45.111	34.399
31)	L7 AR-1260-1	7.199	6.259	1790999	2956729	20.966	29.070 #
32)	L7 AR-1260-2	7.456	6.449	1801983	2678876	19.297	22.028
33)	L7 AR-1260-3	7.818	6.604	126934	2358695	2.009	20.221 #
34)	L7 AR-1260-4	8.029f	7.081	1935516	297447	27.032	3.303 #
35)	L7 AR-1260-5	8.368	7.326	446220	700505	3.511	3.420
36)	L8 AR-1262-1	7.818	6.861f	126934	1048438	1.324	18.306 #
37)	L8 AR-1262-2	8.368	7.081	446220	297447	3.000	2.402
38)	L8 AR-1262-3	8.699	7.599	332928	4251882	3.355	42.132 #
39)	L8 AR-1262-4	8.776	7.673	217844	786474	4.676	4.396
40)	L8 AR-1262-5	9.443	8.186	196305	397062	4.290	5.302
41)	L9 AR-1268-1	8.699	7.599	332928	4251882	1.770	13.871 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103122\
 Data File : PP052888.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2022 15:07
 Operator : YP\AJ
 Sample : N5310-17
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 22:28:17 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 27 04:46:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.776	7.673	217844	786474	1.331	2.900 #
43) L9	AR-1268-3	9.011	7.885	160954	278384	1.062	1.188
44) L9	AR-1268-4	9.443	8.186	196305	397062	3.860	4.674
45) L9	AR-1268-5	9.879	8.477	455996	313889	1.061	0.479 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103122\
Data File : PP052888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2022 15:07
Operator : YP\AJ
Sample : N5310-17
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 22:28:17 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 27 04:46:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.409 min
Delta R.T.: 0.000 min
Response: 49939542
Conc: 22.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.652 min
Delta R.T.: -0.002 min
Response: 45079927
Conc: 23.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.224 min
Delta R.T.: 0.000 min
Response: 33284254
Conc: 26.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.741 min
Delta R.T.: -0.002 min
Response: 41927812
Conc: 25.00 ng/ml



#3 AR-1016-1


R.T.: 5.569 min
Delta R.T.: -0.015 min
Response: 100088
Conc: 1.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.755 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.602 min
Delta R.T.: -0.005 min
Response: 164218
Conc: 1.64 ng/ml



#4 AR-1016-2


R.T.: 4.787 min
Delta R.T.: 0.013 min
Response: 105451
Conc: 1.15 ng/ml



#5 AR-1016-3


R.T.: 5.672 min
Delta R.T.: 0.001 min
Response: 99556
Conc: 1.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.954 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.763 min
Delta R.T.: -0.006 min
Response: 40958
Conc: 0.85 ng/ml



#6 AR-1016-4


R.T.: 4.994 min
Delta R.T.: -0.002 min
Response: 2694406
Conc: 69.16 ng/ml



#7 AR-1016-5


R.T.: 6.067 min
Delta R.T.: 0.001 min
Response: 1326103
Conc: 27.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.210 min
Delta R.T.: -0.003 min
Response: 633315
Conc: 12.60 ng/ml




#8 AR-1221-1

R.T.: 4.618 min
Delta R.T.: -0.003 min
Response: 4071702
Conc: 158.79 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.872 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.717 min
Delta R.T.: 0.010 min
Response: 702243
Conc: 36.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.940 min
Delta R.T.: -0.018 min
Response: 3299417
Conc: 188.48 ng/ml




#10 AR-1221-3

R.T.: 4.791 min
Delta R.T.: 0.007 min
Response: 20893
Conc: 0.37 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 4.036 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.791 min
Delta R.T.: 0.006 min
Response: 20893
Conc: 0.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.036 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.334 min
Delta R.T.: 0.017 min
Response: 997488
Conc: 40.20 ng/ml



#12 AR-1232-2


R.T.: 4.787 min
Delta R.T.: 0.011 min
Response: 105451
Conc: 2.54 ng/ml



#13 AR-1232-3


R.T.: 5.602 min
Delta R.T.: -0.008 min
Response: 164218
Conc: 3.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.955 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.763 min
Delta R.T.: -0.008 min
Response: 40958
Conc: 1.92 ng/ml



#14 AR-1232-4


R.T.: 5.036 min
Delta R.T.: -0.004 min
Response: 1672866
Conc: 81.40 ng/ml



#15 AR-1232-5


R.T.: 5.860 min
Delta R.T.: -0.002 min
Response: 941247
Conc: 54.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.210 min
Delta R.T.: -0.004 min
Response: 633315
Conc: 27.28 ng/ml




#16 AR-1242-1

R.T.: 5.602 min
Delta R.T.: 0.015 min
Response: 164218
Conc: 2.92 ng/ml



#16 AR-1242-1


R.T.: 4.787 min
Delta R.T.: 0.030 min
Response: 105451
Conc: 1.89 ng/ml



#17 AR-1242-2


R.T.: 5.602 min
Delta R.T.: -0.007 min
Response: 164218
Conc: 1.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.787 min
Delta R.T.: 0.011 min
Response: 105451
Conc: 1.37 ng/ml




#18 AR-1242-3

R.T.: 5.672 min
Delta R.T.: 0.000 min
Response: 99556
Conc: 1.96 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T.: 4.956 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.763 min
Delta R.T.: -0.007 min
Response: 40958
Conc: 1.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.036 min
Delta R.T.: -0.004 min
Response: 1672866
Conc: 39.44 ng/ml




#20 AR-1242-5

R.T.: 6.516 min
Delta R.T.: 0.005 min
Response: 861665
Conc: 21.61 ng/ml



#20 AR-1242-5


R.T.: 5.569 min
Delta R.T.: -0.002 min
Response: 3058344
Conc: 60.50 ng/ml



#21 AR-1248-1


R.T.: 5.602 min
Delta R.T.: 0.015 min
Response: 164218
Conc: 4.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.757 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.860 min
Delta R.T.: -0.002 min
Response: 941247
Conc: 15.81 ng/ml



#22 AR-1248-2


R.T.: 4.994 min
Delta R.T.: -0.003 min
Response: 2694406
Conc: 47.59 ng/ml



#23 AR-1248-3


R.T.: 6.067 min
Delta R.T.: 0.000 min
Response: 1326103
Conc: 20.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.036 min
Delta R.T.: -0.004 min
Response: 1672866
Conc: 28.49 ng/ml




#24 AR-1248-4

R.T.: 6.470 min
Delta R.T.: -0.003 min
Response: 716883
Conc: 11.13 ng/ml



#24 AR-1248-4


R.T.: 5.210 min
Delta R.T.: -0.004 min
Response: 633315
Conc: 9.10 ng/ml



#25 AR-1248-5


R.T.: 6.516 min
Delta R.T.: 0.004 min
Response: 861665
Conc: 13.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.611 min
Delta R.T.: -0.003 min
Response: 321855
Conc: 5.07 ng/ml




#26 AR-1254-1

R.T.: 6.445 min
Delta R.T.: -0.002 min
Response: 2514058
Conc: 33.12 ng/ml



#26 AR-1254-1


R.T.: 5.569 min
Delta R.T.: -0.004 min
Response: 3058344
Conc: 28.48 ng/ml



#27 AR-1254-2


R.T.: 6.666 min
Delta R.T.: -0.002 min
Response: 3371244
Conc: 30.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.718 min
Delta R.T.: -0.004 min
Response: 2753525
Conc: 28.81 ng/ml




#28 AR-1254-3

R.T.: 7.033 min
Delta R.T.: -0.003 min
Response: 4277610
Conc: 38.43 ng/ml



#28 AR-1254-3


R.T.: 6.128 min
Delta R.T.: -0.004 min
Response: 5360373
Conc: 35.08 ng/ml



#29 AR-1254-4


R.T.: 7.320 min
Delta R.T.: -0.001 min
Response: 1972602
Conc: 28.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.358 min
Delta R.T.: -0.005 min
Response: 2439076
Conc: 26.86 ng/ml




#30 AR-1254-5

R.T.: 7.741 min
Delta R.T.: -0.002 min
Response: 3551715
Conc: 45.11 ng/ml



#30 AR-1254-5


R.T.: 6.781 min
Delta R.T.: -0.005 min
Response: 4785931
Conc: 34.40 ng/ml



#31 AR-1260-1


R.T.: 7.199 min
Delta R.T.: 0.000 min
Response: 1790999
Conc: 20.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.259 min
Delta R.T.: -0.002 min
Response: 2956729
Conc: 29.07 ng/ml




#32 AR-1260-2

R.T.: 7.456 min
Delta R.T.: 0.000 min
Response: 1801983
Conc: 19.30 ng/ml



#32 AR-1260-2


R.T.: 6.449 min
Delta R.T.: -0.002 min
Response: 2678876
Conc: 22.03 ng/ml



#33 AR-1260-3


R.T.: 7.818 min
Delta R.T.: 0.000 min
Response: 126934
Conc: 2.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.604 min
Delta R.T.: -0.004 min
Response: 2358695
Conc: 20.22 ng/ml




#34 AR-1260-4

R.T.: 8.029 min
Delta R.T.: -0.016 min
Response: 1935516
Conc: 27.03 ng/ml



#34 AR-1260-4


R.T.: 7.081 min
Delta R.T.: -0.002 min
Response: 297447
Conc: 3.30 ng/ml



#35 AR-1260-5


R.T.: 8.368 min
Delta R.T.: 0.000 min
Response: 446220
Conc: 3.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.326 min
Delta R.T.: -0.001 min
Response: 700505
Conc: 3.42 ng/ml




#36 AR-1262-1

R.T.: 7.818 min
Delta R.T.: 0.000 min
Response: 126934
Conc: 1.32 ng/ml



#36 AR-1262-1


R.T.: 6.861 min
Delta R.T.: -0.016 min
Response: 1048438
Conc: 18.31 ng/ml



#37 AR-1262-2


R.T.: 8.368 min
Delta R.T.: 0.000 min
Response: 446220
Conc: 3.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.081 min
Delta R.T.: -0.003 min
Response: 297447
Conc: 2.40 ng/ml




#38 AR-1262-3

R.T.: 8.699 min
Delta R.T.: 0.012 min
Response: 332928
Conc: 3.36 ng/ml



#38 AR-1262-3


R.T.: 7.599 min
Delta R.T.: -0.015 min
Response: 4251882
Conc: 42.13 ng/ml



#39 AR-1262-4


R.T.: 8.776 min
Delta R.T.: -0.001 min
Response: 217844
Conc: 4.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.673 min
Delta R.T.: -0.006 min
Response: 786474
Conc: 4.40 ng/ml




#40 AR-1262-5

R.T.: 9.443 min
Delta R.T.: -0.006 min
Response: 196305
Conc: 4.29 ng/ml



#40 AR-1262-5


R.T.: 8.186 min
Delta R.T.: 0.005 min
Response: 397062
Conc: 5.30 ng/ml



#41 AR-1268-1


R.T.: 8.699 min
Delta R.T.: 0.015 min
Response: 332928
Conc: 1.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.599 min
Delta R.T.: -0.015 min
Response: 4251882
Conc: 13.87 ng/ml




#42 AR-1268-2

R.T.: 8.776 min
Delta R.T.: -0.004 min
Response: 217844
Conc: 1.33 ng/ml



#42 AR-1268-2


R.T.: 7.673 min
Delta R.T.: -0.007 min
Response: 786474
Conc: 2.90 ng/ml



#43 AR-1268-3


R.T.: 9.011 min
Delta R.T.: -0.004 min
Response: 160954
Conc: 1.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.885 min
Delta R.T.: -0.004 min
Response: 278384
Conc: 1.19 ng/ml




#44 AR-1268-4

R.T.: 9.443 min
Delta R.T.: -0.006 min
Response: 196305
Conc: 3.86 ng/ml



#44 AR-1268-4


R.T.: 8.186 min
Delta R.T.: 0.006 min
Response: 397062
Conc: 4.67 ng/ml



#45 AR-1268-5

R.T.: 9.879 min
Delta R.T.: 0.003 min
Response: 455996
Conc: 1.06 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.477 min
Delta R.T.: -0.004 min
Response: 313889
Conc: 0.48 ng/ml