

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103122\
 Data File : PP052894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2022 16:46
 Operator : YP\AJ
 Sample : N5310-21
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BIN 1-2-(2840)-COMP-5

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/01/2022
 Supervised By :Ankita Jodhani 11/01/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 21:59:34 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 27 04:46:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.409	3.653	47402500	43157192	21.452	22.965
2) SA Decachlor...	10.220	8.738	31235325	39060173	24.608	23.290
Target Compounds						
26) L6 AR-1254-1	6.445	5.568	1339678	1523606	17.649m	14.188
27) L6 AR-1254-2	6.664	5.723	1578439	1962206	14.123	20.529 #
28) L6 AR-1254-3	7.031	6.127	2387559	2302185	21.450m	15.068m#
29) L6 AR-1254-4	7.317	6.358	864952	907702	12.283m	9.997m
30) L6 AR-1254-5	7.738	6.779	1564176	2301748	19.867m	16.544

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103122\
Data File : PP052894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2022 16:46
Operator : YP\AJ
Sample : N5310-21
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
BIN 1-2-(2840)-COMP-5


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/01/2022
Supervised By :Ankita Jodhani 11/01/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 21:59:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 27 04:46:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.409 min
Delta R.T.: 0.000 min
Response: 47402500
Conc: 21.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
BIN 1-2-(2840)-COMP-5


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/01/2022
Supervised By :Ankita Jodhani 11/01/2022




#1 Tetrachloro-m-xylene

R.T.: 3.653 min
Delta R.T.: -0.001 min
Response: 43157192
Conc: 22.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: -0.003 min
Response: 31235325
Conc: 24.61 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.738 min
Delta R.T.: -0.005 min
Response: 39060173
Conc: 23.29 ng/ml




#26 AR-1254-1

R.T.: 6.445 min
Delta R.T.: -0.003 min
Response: 1339678
Conc: 17.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
BIN 1-2-(2840)-COMP-5


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/01/2022
Supervised By :Ankita Jodhani 11/01/2022




#26 AR-1254-1

R.T.: 5.568 min
Delta R.T.: -0.004 min
Response: 1523606
Conc: 14.19 ng/ml




#27 AR-1254-2

R.T.: 6.664 min
Delta R.T.: -0.003 min
Response: 1578439
Conc: 14.12 ng/ml



#27 AR-1254-2

R.T.: 5.723 min
Delta R.T.: 0.001 min
Response: 1962206
Conc: 20.53 ng/ml




#28 AR-1254-3

R.T.: 7.031 min
Delta R.T.: -0.005 min
Response: 2387559
Conc: 21.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
BIN 1-2-(2840)-COMP-5


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/01/2022
Supervised By :Ankita Jodhani 11/01/2022




#28 AR-1254-3

R.T.: 6.127 min
Delta R.T.: -0.005 min
Response: 2302185
Conc: 15.07 ng/ml m




#29 AR-1254-4

R.T.: 7.317 min
Delta R.T.: -0.005 min
Response: 864952
Conc: 12.28 ng/ml m



#29 AR-1254-4

R.T.: 6.358 min
Delta R.T.: -0.005 min
Response: 907702
Conc: 10.00 ng/ml m




#30 AR-1254-5

R.T.: 7.738 min
Delta R.T.: -0.005 min
Response: 1564176
Conc: 19.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
BIN 1-2-(2840)-COMP-5

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/01/2022
Supervised By :Ankita Jodhani 11/01/2022



#30 AR-1254-5

R.T.: 6.779 min
Delta R.T.: -0.007 min
Response: 2301748
Conc: 16.54 ng/ml