

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103122\
 Data File : PP052907.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2022 22:01
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 02:04:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 27 04:46:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.408	3.652	96400484	88780392	43.626	47.243
2) SA Decachlor...	10.217	8.737	65542432	75483454	51.636	45.008
Target Compounds						
21) L5 AR-1248-1	5.583	4.752	17832507	18836258	439.043	470.621
22) L5 AR-1248-2	5.858	4.992	25474641	25211046	428.007	445.259
23) L5 AR-1248-3	6.063	5.034	27753421	26321062	427.510	448.307
24) L5 AR-1248-4	6.468	5.209	28409387	31332224	441.110	449.963
25) L5 AR-1248-5	6.507	5.608	27682912	26533368	436.996	418.239



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103122\
Data File : PP052907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2022 22:01
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:04:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 27 04:46:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.408 min
Delta R.T.: -0.002 min
Response: 96400484
Conc: 43.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.652 min
Delta R.T.: -0.002 min
Response: 88780392
Conc: 47.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.217 min
Delta R.T.: -0.007 min
Response: 65542432
Conc: 51.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.737 min
Delta R.T.: -0.006 min
Response: 75483454
Conc: 45.01 ng/ml



#21 AR-1248-1


R.T.: 5.583 min
Delta R.T.: -0.004 min
Response: 17832507
Conc: 439.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.752 min
Delta R.T.: -0.004 min
Response: 18836258
Conc: 470.62 ng/ml




#22 AR-1248-2

R.T.: 5.858 min
Delta R.T.: -0.004 min
Response: 25474641
Conc: 428.01 ng/ml



#22 AR-1248-2


R.T.: 4.992 min
Delta R.T.: -0.005 min
Response: 25211046
Conc: 445.26 ng/ml



#23 AR-1248-3


R.T.: 6.063 min
Delta R.T.: -0.005 min
Response: 27753421
Conc: 427.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.034 min
Delta R.T.: -0.005 min
Response: 26321062
Conc: 448.31 ng/ml




#24 AR-1248-4

R.T.: 6.468 min
Delta R.T.: -0.005 min
Response: 28409387
Conc: 441.11 ng/ml



#24 AR-1248-4


R.T.: 5.209 min
Delta R.T.: -0.005 min
Response: 31332224
Conc: 449.96 ng/ml



#25 AR-1248-5

R.T.: 6.507 min
Delta R.T.: -0.005 min
Response: 27682912
Conc: 437.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.608 min
Delta R.T.: -0.006 min
Response: 26533368
Conc: 418.24 ng/ml