

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103123\
 Data File : PP061396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2023 13:11
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 01:03:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 27 18:44:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.564	3.697	141.2E6	84225188	57.178	56.357
2) SA Decachlor...	10.551	8.771	85616928	66671834	45.403	43.415
Target Compounds						
16) L4 AR-1242-1	5.758	4.798	34124993	22274483	521.659	526.812
17) L4 AR-1242-2	5.781	4.817	47895919	30565205	516.854	537.642
18) L4 AR-1242-3	5.844	4.995	28961534	17396270	507.411	542.029
19) L4 AR-1242-4	5.945	5.080	23531175	17210809	484.047	535.121
20) L4 AR-1242-5	6.694	5.609	23700037	21529022	458.186	542.365



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103123\
Data File : PP061396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2023 13:11
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 01:03:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 27 18:44:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.564 min
Delta R.T.: 0.005 min
Response: 141240235
Conc: 57.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.697 min
Delta R.T.: 0.003 min
Response: 84225188
Conc: 56.36 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.551 min
Delta R.T.: 0.015 min
Response: 85616928
Conc: 45.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.771 min
Delta R.T.: 0.007 min
Response: 66671834
Conc: 43.41 ng/ml



#16 AR-1242-1


R.T.: 5.758 min
Delta R.T.: 0.001 min
Response: 34124993
Conc: 521.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 22274483
Conc: 526.81 ng/ml




#17 AR-1242-2

R.T.: 5.781 min
Delta R.T.: 0.002 min
Response: 47895919
Conc: 516.85 ng/ml



#17 AR-1242-2


R.T.: 4.817 min
Delta R.T.: 0.000 min
Response: 30565205
Conc: 537.64 ng/ml



#18 AR-1242-3


R.T.: 5.844 min
Delta R.T.: 0.001 min
Response: 28961534
Conc: 507.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.995 min
Delta R.T.: 0.000 min
Response: 17396270
Conc: 542.03 ng/ml




#19 AR-1242-4

R.T.: 5.945 min
Delta R.T.: 0.001 min
Response: 23531175
Conc: 484.05 ng/ml



#19 AR-1242-4


R.T.: 5.080 min
Delta R.T.: 0.000 min
Response: 17210809
Conc: 535.12 ng/ml



#20 AR-1242-5

R.T.: 6.694 min
Delta R.T.: 0.002 min
Response: 23700037
Conc: 458.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.609 min
Delta R.T.: 0.000 min
Response: 21529022
Conc: 542.36 ng/ml