

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
 Data File : PP076288.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2025 03:28
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/03/2025
 Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 04:47:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.784	83083421	410.8E6	58.137	62.072
2) SA Decachlor...	10.422	8.781	60077765	413.0E6	55.807	50.221
Target Compounds						
21) L5 AR-1248-1	5.799	4.877	19459760	206.4E6	542.499	549.574
22) L5 AR-1248-2	6.070	5.097	26336725	130.9E6	540.303	532.379
23) L5 AR-1248-3	6.273	5.137	28856940	150.3E6	531.781	523.875m
24) L5 AR-1248-4	6.671	5.309	35165705	160.3E6	529.890	566.834
25) L5 AR-1248-5	6.710	5.696	34137362	274.4E6	526.319	528.111m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076288.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 03:28
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 04:47:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.647 min
Delta R.T.: 0.005 min
Response: 83083421
Conc: 58.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025




#1 Tetrachloro-m-xylene

R.T.: 3.784 min
Delta R.T.: 0.002 min
Response: 410812496
Conc: 62.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.422 min
Delta R.T.: 0.003 min
Response: 60077765
Conc: 55.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.781 min
Delta R.T.: -0.001 min
Response: 412968713
Conc: 50.22 ng/ml




#21 AR-1248-1

R.T.: 5.799 min
Delta R.T.: 0.000 min
Response: 19459760
Conc: 542.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025




#21 AR-1248-1

R.T.: 4.877 min
Delta R.T.: 0.000 min
Response: 206429488
Conc: 549.57 ng/ml




#22 AR-1248-2

R.T.: 6.070 min
Delta R.T.: -0.001 min
Response: 26336725
Conc: 540.30 ng/ml



#22 AR-1248-2

R.T.: 5.097 min
Delta R.T.: 0.001 min
Response: 130916297
Conc: 532.38 ng/ml




#23 AR-1248-3

R.T.: 6.273 min
Delta R.T.: -0.001 min
Response: 28856940
Conc: 531.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025




#23 AR-1248-3

R.T.: 5.137 min
Delta R.T.: 0.000 min
Response: 150295962
Conc: 523.88 ng/ml m




#24 AR-1248-4

R.T.: 6.671 min
Delta R.T.: -0.001 min
Response: 35165705
Conc: 529.89 ng/ml



#24 AR-1248-4

R.T.: 5.309 min
Delta R.T.: 0.000 min
Response: 160250494
Conc: 566.83 ng/ml




#25 AR-1248-5

R.T.: 6.710 min
Delta R.T.: 0.000 min
Response: 34137362
Conc: 526.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025



#25 AR-1248-5

R.T.: 5.696 min
Delta R.T.: -0.001 min
Response: 274437343
Conc: 528.11 ng/ml m