

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110121\
 Data File : PP040601.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Nov 2021 18:32
 Operator : AJ\MA
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 05:35:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 02 05:35:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.742	3.768	1205672	1688135	71.506	71.652
2)	SA Decachlor...	10.654	9.039	1031566	1541714	74.844	70.734
Target Compounds							
3)	L1 AR-1016-1	6.061	5.019	409460	477313	706.070	698.316
4)	L1 AR-1016-2	6.084	5.038	607897	705528	706.362	704.534
5)	L1 AR-1016-3	6.150	5.229	387355	392197	705.389	708.479
6)	L1 AR-1016-4	6.257	5.282	319632	313228	715.180	705.473
7)	L1 AR-1016-5	6.571	5.510	311611	394194	720.222	718.214
31)	L7 AR-1260-1	7.747	6.605	525159	634032	711.708	701.447
32)	L7 AR-1260-2	8.013	6.804	613022	741692	702.331	708.790
33)	L7 AR-1260-3	8.379	6.957	482648	710264	702.887	708.371
34)	L7 AR-1260-4	8.609	7.440	582647	625828	714.102	706.873
35)	L7 AR-1260-5	8.924	7.690	1095436	1444233	724.953	718.440



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110121\
Data File : PP040601.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Nov 2021 18:32
Operator : AJ\MA
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 02 05:35:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 02 05:35:16 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.742 min
Delta R.T.: 0.000 min
Response: 1205672
Conc: 71.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.768 min
Delta R.T.: 0.000 min
Response: 1688135
Conc: 71.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.654 min
Delta R.T.: 0.000 min
Response: 1031566
Conc: 74.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.039 min
Delta R.T.: 0.000 min
Response: 1541714
Conc: 70.73 ng/ml



#3 AR-1016-1


R.T.: 6.061 min
Delta R.T.: 0.000 min
Response: 409460
Conc: 706.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 5.019 min
Delta R.T.: 0.000 min
Response: 477313
Conc: 698.32 ng/ml




#4 AR-1016-2

R.T.: 6.084 min
Delta R.T.: 0.000 min
Response: 607897
Conc: 706.36 ng/ml



#4 AR-1016-2


R.T.: 5.038 min
Delta R.T.: 0.000 min
Response: 705528
Conc: 704.53 ng/ml



#5 AR-1016-3


R.T.: 6.150 min
Delta R.T.: 0.000 min
Response: 387355
Conc: 705.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 392197
Conc: 708.48 ng/ml




#6 AR-1016-4

R.T.: 6.257 min
Delta R.T.: 0.000 min
Response: 319632
Conc: 715.18 ng/ml



#6 AR-1016-4


R.T.: 5.282 min
Delta R.T.: 0.000 min
Response: 313228
Conc: 705.47 ng/ml



#7 AR-1016-5


R.T.: 6.571 min
Delta R.T.: 0.000 min
Response: 311611
Conc: 720.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.510 min
Delta R.T.: 0.000 min
Response: 394194
Conc: 718.21 ng/ml




#31 AR-1260-1

R.T.: 7.747 min
Delta R.T.: 0.000 min
Response: 525159
Conc: 711.71 ng/ml



#31 AR-1260-1


R.T.: 6.605 min
Delta R.T.: 0.000 min
Response: 634032
Conc: 701.45 ng/ml



#32 AR-1260-2


R.T.: 8.013 min
Delta R.T.: 0.000 min
Response: 613022
Conc: 702.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.804 min
Delta R.T.: 0.001 min
Response: 741692
Conc: 708.79 ng/ml




#33 AR-1260-3

R.T.: 8.379 min
Delta R.T.: 0.000 min
Response: 482648
Conc: 702.89 ng/ml



#33 AR-1260-3


R.T.: 6.957 min
Delta R.T.: 0.000 min
Response: 710264
Conc: 708.37 ng/ml



#34 AR-1260-4


R.T.: 8.609 min
Delta R.T.: 0.001 min
Response: 582647
Conc: 714.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 7.440 min
Delta R.T.: 0.000 min
Response: 625828
Conc: 706.87 ng/ml



#35 AR-1260-5

R.T.: 8.924 min
Delta R.T.: 0.000 min
Response: 1095436
Conc: 724.95 ng/ml



#35 AR-1260-5

R.T.: 7.690 min
Delta R.T.: 0.000 min
Response: 1444233
Conc: 718.44 ng/ml