

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110122\
 Data File : PP052966.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Nov 2022 04:18
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 05:20:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 27 04:46:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.406	3.650	98034474	88741328	44.366	47.222
2) SA Decachlor...	10.211	8.734	64674765	79517132	50.953	47.413
Target Compounds						
16) L4 AR-1242-1	5.580	4.751	22315771	23890432	396.508	428.087
17) L4 AR-1242-2	5.602	4.770	32798462	32857279	397.039	426.013
18) L4 AR-1242-3	5.665	4.949	20451990	17821324	402.143	434.369
19) L4 AR-1242-4	5.764	5.033	16466857	18625344	406.493	439.142
20) L4 AR-1242-5	6.505	5.564	19607826	20052079	491.819	396.688



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110122\
Data File : PP052966.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Nov 2022 04:18
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 02 05:20:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 27 04:46:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.406 min
Delta R.T.: -0.004 min
Response: 98034474
Conc: 44.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.650 min
Delta R.T.: -0.003 min
Response: 88741328
Conc: 47.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.211 min
Delta R.T.: -0.012 min
Response: 64674765
Conc: 50.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.734 min
Delta R.T.: -0.009 min
Response: 79517132
Conc: 47.41 ng/ml



#16 AR-1242-1


R.T.: 5.580 min
Delta R.T.: -0.007 min
Response: 22315771
Conc: 396.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.751 min
Delta R.T.: -0.006 min
Response: 23890432
Conc: 428.09 ng/ml




#17 AR-1242-2

R.T.: 5.602 min
Delta R.T.: -0.006 min
Response: 32798462
Conc: 397.04 ng/ml



#17 AR-1242-2


R.T.: 4.770 min
Delta R.T.: -0.006 min
Response: 32857279
Conc: 426.01 ng/ml



#18 AR-1242-3


R.T.: 5.665 min
Delta R.T.: -0.006 min
Response: 20451990
Conc: 402.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.949 min
Delta R.T.: -0.007 min
Response: 17821324
Conc: 434.37 ng/ml




#19 AR-1242-4

R.T.: 5.764 min
Delta R.T.: -0.007 min
Response: 16466857
Conc: 406.49 ng/ml



#19 AR-1242-4


R.T.: 5.033 min
Delta R.T.: -0.007 min
Response: 18625344
Conc: 439.14 ng/ml



#20 AR-1242-5

R.T.: 6.505 min
Delta R.T.: -0.007 min
Response: 19607826
Conc: 491.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.564 min
Delta R.T.: -0.006 min
Response: 20052079
Conc: 396.69 ng/ml