

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110124\
 Data File : PP068345.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2024 15:11
 Operator : YP\AJ
 Sample : PB164590BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB164590BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 00:44:31 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 05:48:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.761	4.055	19412421	19801100	20.972	19.597
2) SA Decachlor...	10.679	9.213	26828761	24824387	23.303	22.136
Target Compounds						
3) L1 AR-1016-1	5.926	5.161	15895830	14241675	480.720	407.053
4) L1 AR-1016-2	5.948	5.181	23549461	20032499	482.859	415.883
5) L1 AR-1016-3	6.011	5.362	15659222	11139411	496.930	413.528
6) L1 AR-1016-4	6.110	5.400	11944497	9979509	463.979	414.344
7) L1 AR-1016-5	6.404	5.619	11528999	12365842	425.945	406.639
31) L7 AR-1260-1	7.528	6.663	24792052	24404000	458.911	428.426
32) L7 AR-1260-2	7.780	6.849	28347851	28491402	448.040	427.129
33) L7 AR-1260-3	8.142	7.007	20858764	27507689	402.300	431.824
34) L7 AR-1260-4	8.380	7.482	25463553	21042129	423.022	380.559
35) L7 AR-1260-5	8.718	7.720	43902210	46430189	407.588	379.646



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110124\
Data File : PP068345.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2024 15:11
Operator : YP\AJ
Sample : PB164590BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB164590BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 00:44:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 05:48:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.761 min
Delta R.T.: 0.006 min
Response: 19412421
Conc: 20.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB164590BS




#1 Tetrachloro-m-xylene

R.T.: 4.055 min
Delta R.T.: 0.003 min
Response: 19801100
Conc: 19.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.679 min
Delta R.T.: 0.010 min
Response: 26828761
Conc: 23.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.213 min
Delta R.T.: -0.006 min
Response: 24824387
Conc: 22.14 ng/ml



#3 AR-1016-1


R.T.: 5.926 min
Delta R.T.: 0.007 min
Response: 15895830
Conc: 480.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB164590BS




#3 AR-1016-1

R.T.: 5.161 min
Delta R.T.: 0.000 min
Response: 14241675
Conc: 407.05 ng/ml




#4 AR-1016-2

R.T.: 5.948 min
Delta R.T.: 0.006 min
Response: 23549461
Conc: 482.86 ng/ml



#4 AR-1016-2


R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 20032499
Conc: 415.88 ng/ml



#5 AR-1016-3


R.T.: 6.011 min
Delta R.T.: 0.006 min
Response: 15659222
Conc: 496.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB164590BS




#5 AR-1016-3

R.T.: 5.362 min
Delta R.T.: 0.000 min
Response: 11139411
Conc: 413.53 ng/ml




#6 AR-1016-4

R.T.: 6.110 min
Delta R.T.: 0.007 min
Response: 11944497
Conc: 463.98 ng/ml



#6 AR-1016-4


R.T.: 5.400 min
Delta R.T.: 0.000 min
Response: 9979509
Conc: 414.34 ng/ml



#7 AR-1016-5


R.T.: 6.404 min
Delta R.T.: 0.006 min
Response: 11528999
Conc: 425.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB164590BS




#7 AR-1016-5

R.T.: 5.619 min
Delta R.T.: 0.000 min
Response: 12365842
Conc: 406.64 ng/ml




#31 AR-1260-1

R.T.: 7.528 min
Delta R.T.: 0.006 min
Response: 24792052
Conc: 458.91 ng/ml



#31 AR-1260-1


R.T.: 6.663 min
Delta R.T.: -0.001 min
Response: 24404000
Conc: 428.43 ng/ml



#32 AR-1260-2


R.T.: 7.780 min
Delta R.T.: 0.005 min
Response: 28347851
Conc: 448.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB164590BS




#32 AR-1260-2

R.T.: 6.849 min
Delta R.T.: -0.001 min
Response: 28491402
Conc: 427.13 ng/ml




#33 AR-1260-3

R.T.: 8.142 min
Delta R.T.: 0.005 min
Response: 20858764
Conc: 402.30 ng/ml



#33 AR-1260-3


R.T.: 7.007 min
Delta R.T.: -0.002 min
Response: 27507689
Conc: 431.82 ng/ml



#34 AR-1260-4


R.T.: 8.380 min
Delta R.T.: 0.005 min
Response: 25463553
Conc: 423.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB164590BS




#34 AR-1260-4

R.T.: 7.482 min
Delta R.T.: -0.002 min
Response: 21042129
Conc: 380.56 ng/ml



#35 AR-1260-5

R.T.: 8.718 min
Delta R.T.: 0.006 min
Response: 43902210
Conc: 407.59 ng/ml



#35 AR-1260-5

R.T.: 7.720 min
Delta R.T.: -0.002 min
Response: 46430189
Conc: 379.65 ng/ml