

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110220\
 Data File : PP031028.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Nov 2020 9:51
 Operator : DD\AJ
 Sample : AR1268CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 01:35:03 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\AR1268PP110120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 31 02:37:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.802	4.077	2619387	2334563	43.955	44.351
2) SA Decachlor...	10.692	9.410	2645337	2888027	46.098	45.743
Target Compounds						
41) L9 AR-1268-1	9.264	8.306	2195651	2427769	450.750	451.300
42) L9 AR-1268-2	9.356	8.370	2165920	2380752	447.176	450.988
43) L9 AR-1268-3	9.573	8.579	1816820	2009552	449.010	455.711
44) L9 AR-1268-4	9.984	8.866	748679	820614	466.145	456.248
45) L9 AR-1268-5	10.375	9.157	5993855	6634501	466.326	464.229



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110220\
Data File : PP031028.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Nov 2020 9:51
Operator : DD\AJ
Sample : AR1268CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 01:35:03 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\AR1268PP110120.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 31 02:37:11 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.802 min
Delta R.T.: 0.000 min
Response: 2619387
Conc: 43.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.077 min
Delta R.T.: -0.002 min
Response: 2334563
Conc: 44.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.692 min
Delta R.T.: 0.000 min
Response: 2645337
Conc: 46.10 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.410 min
Delta R.T.: -0.002 min
Response: 2888027
Conc: 45.74 ng/ml



#41 AR-1268-1


R.T.: 9.264 min
Delta R.T.: 0.000 min
Response: 2195651
Conc: 450.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268CCC500




#41 AR-1268-1

R.T.: 8.306 min
Delta R.T.: -0.001 min
Response: 2427769
Conc: 451.30 ng/ml




#42 AR-1268-2

R.T.: 9.356 min
Delta R.T.: 0.000 min
Response: 2165920
Conc: 447.18 ng/ml



#42 AR-1268-2


R.T.: 8.370 min
Delta R.T.: -0.002 min
Response: 2380752
Conc: 450.99 ng/ml



#43 AR-1268-3


R.T.: 9.573 min
Delta R.T.: 0.000 min
Response: 1816820
Conc: 449.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268CCC500




#43 AR-1268-3

R.T.: 8.579 min
Delta R.T.: -0.002 min
Response: 2009552
Conc: 455.71 ng/ml




#44 AR-1268-4

R.T.: 9.984 min
Delta R.T.: 0.002 min
Response: 748679
Conc: 466.14 ng/ml



#44 AR-1268-4


R.T.: 8.866 min
Delta R.T.: -0.002 min
Response: 820614
Conc: 456.25 ng/ml



#45 AR-1268-5

R.T.: 10.375 min
Delta R.T.: 0.001 min
Response: 5993855
Conc: 466.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268CCC500



#45 AR-1268-5

R.T.: 9.157 min
Delta R.T.: -0.002 min
Response: 6634501
Conc: 464.23 ng/ml