

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110322\
 Data File : PP053017.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Nov 2022 10:10
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 11:05:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 27 04:46:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.404	3.649	103.4E6	96506432	46.810	51.354
2)	SA Decachlor...	10.206	8.729	74340773	85833968	58.568	51.179
Target Compounds							
26)	L6 AR-1254-1	6.437	5.562	35690805	50675502	470.195	471.911
27)	L6 AR-1254-2	6.657	5.712	52541405	44263345	470.101	463.083
28)	L6 AR-1254-3	7.025	6.120	53154519	70179472	477.537	459.334
29)	L6 AR-1254-4	7.311	6.351	35416356	42389050	502.927	466.869
30)	L6 AR-1254-5	7.732	6.774	41094707	65991959	521.949	474.321



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110322\
Data File : PP053017.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2022 10:10
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 11:05:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 27 04:46:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.404 min
Delta R.T.: -0.006 min
Response: 103436202
Conc: 46.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.649 min
Delta R.T.: -0.005 min
Response: 96506432
Conc: 51.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.206 min
Delta R.T.: -0.018 min
Response: 74340773
Conc: 58.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.729 min
Delta R.T.: -0.014 min
Response: 85833968
Conc: 51.18 ng/ml



#26 AR-1254-1


R.T.: 6.437 min
Delta R.T.: -0.010 min
Response: 35690805
Conc: 470.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.562 min
Delta R.T.: -0.010 min
Response: 50675502
Conc: 471.91 ng/ml




#27 AR-1254-2

R.T.: 6.657 min
Delta R.T.: -0.011 min
Response: 52541405
Conc: 470.10 ng/ml



#27 AR-1254-2


R.T.: 5.712 min
Delta R.T.: -0.010 min
Response: 44263345
Conc: 463.08 ng/ml



#28 AR-1254-3


R.T.: 7.025 min
Delta R.T.: -0.011 min
Response: 53154519
Conc: 477.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.120 min
Delta R.T.: -0.011 min
Response: 70179472
Conc: 459.33 ng/ml




#29 AR-1254-4

R.T.: 7.311 min
Delta R.T.: -0.011 min
Response: 35416356
Conc: 502.93 ng/ml



#29 AR-1254-4


R.T.: 6.351 min
Delta R.T.: -0.012 min
Response: 42389050
Conc: 466.87 ng/ml



#30 AR-1254-5

R.T.: 7.732 min
Delta R.T.: -0.011 min
Response: 41094707
Conc: 521.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.774 min
Delta R.T.: -0.012 min
Response: 65991959
Conc: 474.32 ng/ml