

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110420\
 Data File : PP031130.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Nov 2020 21:44
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:09:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 04 08:24:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.799	4.074	3038115	3030592	45.487	45.594
2)	SA Decachlor...	10.679	9.395	1907977	2154874	47.719	51.606
Target Compounds							
3)	L1 AR-1016-1	6.104	5.335	884077	844337	454.052	455.101
4)	L1 AR-1016-2	6.127	5.355	1302903	1236384	456.399	454.855
5)	L1 AR-1016-3	6.193	5.547	806790	670707	463.205	529.536
6)	L1 AR-1016-4	6.299	5.599	673005	490522	465.759	531.878
7)	L1 AR-1016-5	6.613	5.828	625640	613844	501.883	467.680
8)	L2 AR-1221-1	5.041	4.332	96366	85026	159.583	150.967
9)	L2 AR-1221-2	5.143	4.432	146929	131552	327.114	317.084
10)	L2 AR-1221-3	5.228	4.521	526456	508564	390.063	387.799
11)	L3 AR-1232-1	5.228	4.521	526456	508564	430.458	424.561
12)	L3 AR-1232-2	5.812	5.355	694283	1236384	1143.354	1163.440
13)	L3 AR-1232-3	6.127	5.547	1302903	670707	1166.303	1308.194
14)	L3 AR-1232-4	6.299	5.644	673005	603194	1185.810	1381.087
15)	L3 AR-1232-5	6.399	5.828	414404	613844	1195.936	1271.178
16)	L4 AR-1242-1	6.104	5.335	884077	844337	658.045	656.086
17)	L4 AR-1242-2	6.127	5.355	1302903	1236384	667.129	662.106
18)	L4 AR-1242-3	6.193	5.547	806790	670707	661.537	714.387
19)	L4 AR-1242-4	6.299	5.644	673005	603194	664.293	695.562
20)	L4 AR-1242-5	7.078	6.209	100052	482780	104.847	484.958 #
21)	L5 AR-1248-1	6.104	5.335	884077	844337	916.806	912.498
22)	L5 AR-1248-2	6.399	5.599	414404	490522	358.299	445.078
23)	L5 AR-1248-3	6.613	5.644	625640	603194	439.728	498.365
24)	L5 AR-1248-4	7.039	5.828	115031	613844	75.475	415.933 #
25)	L5 AR-1248-5	7.078	6.252	100052	95214	65.950	75.391
26)	L6 AR-1254-1	7.010	6.209	403377	482780	287.306	245.478
27)	L6 AR-1254-2	7.238	6.367	431016	421213	206.775	256.333
28)	L6 AR-1254-3	7.617	6.808	310369	788158	134.473	287.825 #
29)	L6 AR-1254-4	7.910	7.028	197900	104022	120.878	64.465 #
30)	L6 AR-1254-5	8.337	7.456	1212891	1384476	718.485	632.329
31)	L7 AR-1260-1	7.785	6.925	960148	995315	522.718	496.621
32)	L7 AR-1260-2	8.050	7.122	1069004	1200662	496.214	512.706
33)	L7 AR-1260-3	8.415	7.278	757447	1177191	460.155	517.789
34)	L7 AR-1260-4	8.644	7.761	771807	991938	415.094	543.689 #
35)	L7 AR-1260-5	8.957	8.008	1957592	2433231	519.194	540.673
36)	L8 AR-1262-1	8.415	7.498	757447	1094274	376.987	471.116
37)	L8 AR-1262-2	8.957	8.008	1957592	2433231	538.599	586.920
38)	L8 AR-1262-3	9.272	8.295	1306084	589871	505.241	347.918 #
39)	L8 AR-1262-4	9.344	8.357	408241	1863916	198.287	586.957 #
40)	L8 AR-1262-5	9.974	8.855	672919	702897	475.051	437.333
41)	L9 AR-1268-1	9.272	8.295	1306084	589871	298.998	120.491 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110420\
Data File : PP031130.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Nov 2020 21:44
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:09:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 04 08:24:38 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	9.344	8.357	408241	1863916	94.936	391.913 #
43)	L9 AR-1268-3	9.562	8.568	70173	32406	19.403	8.099 #
44)	L9 AR-1268-4	9.974	8.855	672919	702897	424.856	414.862
45)	L9 AR-1268-5	10.363	9.144	249501	207220	20.863	15.452 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110420\
Data File : PP031130.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Nov 2020 21:44
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:09:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 04 08:24:38 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.799 min
Delta R.T.: -0.002 min
Response: 3038115
Conc: 45.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.074 min
Delta R.T.: -0.002 min
Response: 3030592
Conc: 45.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.679 min
Delta R.T.: -0.008 min
Response: 1907977
Conc: 47.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.395 min
Delta R.T.: -0.008 min
Response: 2154874
Conc: 51.61 ng/ml



#3 AR-1016-1


R.T.: 6.104 min
Delta R.T.: -0.003 min
Response: 884077
Conc: 454.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.335 min
Delta R.T.: -0.004 min
Response: 844337
Conc: 455.10 ng/ml




#4 AR-1016-2

R.T.: 6.127 min
Delta R.T.: -0.003 min
Response: 1302903
Conc: 456.40 ng/ml



#4 AR-1016-2


R.T.: 5.355 min
Delta R.T.: -0.004 min
Response: 1236384
Conc: 454.85 ng/ml



#5 AR-1016-3


R.T.: 6.193 min
Delta R.T.: -0.003 min
Response: 806790
Conc: 463.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.547 min
Delta R.T.: -0.004 min
Response: 670707
Conc: 529.54 ng/ml




#6 AR-1016-4

R.T.: 6.299 min
Delta R.T.: -0.003 min
Response: 673005
Conc: 465.76 ng/ml



#6 AR-1016-4


R.T.: 5.599 min
Delta R.T.: -0.004 min
Response: 490522
Conc: 531.88 ng/ml



#7 AR-1016-5


R.T.: 6.613 min
Delta R.T.: -0.003 min
Response: 625640
Conc: 501.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.828 min
Delta R.T.: -0.005 min
Response: 613844
Conc: 467.68 ng/ml




#8 AR-1221-1

R.T.: 5.041 min
Delta R.T.: -0.006 min
Response: 96366
Conc: 159.58 ng/ml



#8 AR-1221-1


R.T.: 4.332 min
Delta R.T.: -0.004 min
Response: 85026
Conc: 150.97 ng/ml



#9 AR-1221-2


R.T.: 5.143 min
Delta R.T.: -0.001 min
Response: 146929
Conc: 327.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#9 AR-1221-2

R.T.: 4.432 min
Delta R.T.: -0.003 min
Response: 131552
Conc: 317.08 ng/ml




#10 AR-1221-3

R.T.: 5.228 min
Delta R.T.: -0.002 min
Response: 526456
Conc: 390.06 ng/ml



#10 AR-1221-3


R.T.: 4.521 min
Delta R.T.: -0.004 min
Response: 508564
Conc: 387.80 ng/ml



#11 AR-1232-1


R.T.: 5.228 min
Delta R.T.: -0.002 min
Response: 526456
Conc: 430.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#11 AR-1232-1

R.T.: 4.521 min
Delta R.T.: -0.004 min
Response: 508564
Conc: 424.56 ng/ml




#12 AR-1232-2

R.T.: 5.812 min
Delta R.T.: -0.002 min
Response: 694283
Conc: 1143.35 ng/ml



#12 AR-1232-2


R.T.: 5.355 min
Delta R.T.: -0.004 min
Response: 1236384
Conc: 1163.44 ng/ml



#13 AR-1232-3


R.T.: 6.127 min
Delta R.T.: -0.003 min
Response: 1302903
Conc: 1166.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#13 AR-1232-3

R.T.: 5.547 min
Delta R.T.: -0.004 min
Response: 670707
Conc: 1308.19 ng/ml




#14 AR-1232-4

R.T.: 6.299 min
Delta R.T.: -0.003 min
Response: 673005
Conc: 1185.81 ng/ml



#14 AR-1232-4


R.T.: 5.644 min
Delta R.T.: -0.004 min
Response: 603194
Conc: 1381.09 ng/ml



#15 AR-1232-5


R.T.: 6.399 min
Delta R.T.: -0.003 min
Response: 414404
Conc: 1195.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#15 AR-1232-5

R.T.: 5.828 min
Delta R.T.: -0.005 min
Response: 613844
Conc: 1271.18 ng/ml




#16 AR-1242-1

R.T.: 6.104 min
Delta R.T.: -0.003 min
Response: 884077
Conc: 658.04 ng/ml



#16 AR-1242-1


R.T.: 5.335 min
Delta R.T.: -0.004 min
Response: 844337
Conc: 656.09 ng/ml



#17 AR-1242-2


R.T.: 6.127 min
Delta R.T.: -0.003 min
Response: 1302903
Conc: 667.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#17 AR-1242-2

R.T.: 5.355 min
Delta R.T.: -0.004 min
Response: 1236384
Conc: 662.11 ng/ml




#18 AR-1242-3

R.T.: 6.193 min
Delta R.T.: -0.003 min
Response: 806790
Conc: 661.54 ng/ml



#18 AR-1242-3


R.T.: 5.547 min
Delta R.T.: -0.004 min
Response: 670707
Conc: 714.39 ng/ml



#19 AR-1242-4


R.T.: 6.299 min
Delta R.T.: -0.003 min
Response: 673005
Conc: 664.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#19 AR-1242-4

R.T.: 5.644 min
Delta R.T.: -0.004 min
Response: 603194
Conc: 695.56 ng/ml




#20 AR-1242-5

R.T.: 7.078 min
Delta R.T.: -0.002 min
Response: 100052
Conc: 104.85 ng/ml



#20 AR-1242-5


R.T.: 6.209 min
Delta R.T.: -0.004 min
Response: 482780
Conc: 484.96 ng/ml



#21 AR-1248-1


R.T.: 6.104 min
Delta R.T.: -0.003 min
Response: 884077
Conc: 916.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#21 AR-1248-1

R.T.: 5.335 min
Delta R.T.: -0.004 min
Response: 844337
Conc: 912.50 ng/ml




#22 AR-1248-2

R.T.: 6.399 min
Delta R.T.: -0.003 min
Response: 414404
Conc: 358.30 ng/ml



#22 AR-1248-2


R.T.: 5.599 min
Delta R.T.: -0.004 min
Response: 490522
Conc: 445.08 ng/ml



#23 AR-1248-3


R.T.: 6.613 min
Delta R.T.: -0.003 min
Response: 625640
Conc: 439.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#23 AR-1248-3

R.T.: 5.644 min
Delta R.T.: -0.004 min
Response: 603194
Conc: 498.37 ng/ml




#24 AR-1248-4

R.T.: 7.039 min
Delta R.T.: -0.003 min
Response: 115031
Conc: 75.47 ng/ml



#24 AR-1248-4


R.T.: 5.828 min
Delta R.T.: -0.005 min
Response: 613844
Conc: 415.93 ng/ml



#25 AR-1248-5


R.T.: 7.078 min
Delta R.T.: -0.002 min
Response: 100052
Conc: 65.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#25 AR-1248-5

R.T.: 6.252 min
Delta R.T.: -0.005 min
Response: 95214
Conc: 75.39 ng/ml




#26 AR-1254-1

R.T.: 7.010 min
Delta R.T.: -0.003 min
Response: 403377
Conc: 287.31 ng/ml



#26 AR-1254-1


R.T.: 6.209 min
Delta R.T.: -0.004 min
Response: 482780
Conc: 245.48 ng/ml



#27 AR-1254-2


R.T.: 7.238 min
Delta R.T.: -0.003 min
Response: 431016
Conc: 206.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#27 AR-1254-2

R.T.: 6.367 min
Delta R.T.: -0.004 min
Response: 421213
Conc: 256.33 ng/ml




#28 AR-1254-3

R.T.: 7.617 min
Delta R.T.: -0.003 min
Response: 310369
Conc: 134.47 ng/ml



#28 AR-1254-3


R.T.: 6.808 min
Delta R.T.: 0.014 min
Response: 788158
Conc: 287.83 ng/ml



#29 AR-1254-4


R.T.: 7.910 min
Delta R.T.: -0.004 min
Response: 197900
Conc: 120.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#29 AR-1254-4

R.T.: 7.028 min
Delta R.T.: -0.004 min
Response: 104022
Conc: 64.47 ng/ml




#30 AR-1254-5

R.T.: 8.337 min
Delta R.T.: -0.004 min
Response: 1212891
Conc: 718.49 ng/ml



#30 AR-1254-5


R.T.: 7.456 min
Delta R.T.: -0.005 min
Response: 1384476
Conc: 632.33 ng/ml



#31 AR-1260-1


R.T.: 7.785 min
Delta R.T.: -0.004 min
Response: 960148
Conc: 522.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#31 AR-1260-1

R.T.: 6.925 min
Delta R.T.: -0.006 min
Response: 995315
Conc: 496.62 ng/ml




#32 AR-1260-2

R.T.: 8.050 min
Delta R.T.: -0.004 min
Response: 1069004
Conc: 496.21 ng/ml



#32 AR-1260-2


R.T.: 7.122 min
Delta R.T.: -0.005 min
Response: 1200662
Conc: 512.71 ng/ml



#33 AR-1260-3


R.T.: 8.415 min
Delta R.T.: -0.004 min
Response: 757447
Conc: 460.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#33 AR-1260-3

R.T.: 7.278 min
Delta R.T.: -0.005 min
Response: 1177191
Conc: 517.79 ng/ml




#34 AR-1260-4

R.T.: 8.644 min
Delta R.T.: -0.004 min
Response: 771807
Conc: 415.09 ng/ml



#34 AR-1260-4


R.T.: 7.761 min
Delta R.T.: -0.006 min
Response: 991938
Conc: 543.69 ng/ml



#35 AR-1260-5


R.T.: 8.957 min
Delta R.T.: -0.005 min
Response: 1957592
Conc: 519.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#35 AR-1260-5

R.T.: 8.008 min
Delta R.T.: -0.006 min
Response: 2433231
Conc: 540.67 ng/ml




#36 AR-1262-1

R.T.: 8.415 min
Delta R.T.: -0.004 min
Response: 757447
Conc: 376.99 ng/ml



#36 AR-1262-1


R.T.: 7.498 min
Delta R.T.: -0.005 min
Response: 1094274
Conc: 471.12 ng/ml



#37 AR-1262-2


R.T.: 8.957 min
Delta R.T.: -0.005 min
Response: 1957592
Conc: 538.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#37 AR-1262-2

R.T.: 8.008 min
Delta R.T.: -0.005 min
Response: 2433231
Conc: 586.92 ng/ml




#38 AR-1262-3

R.T.: 9.272 min
Delta R.T.: 0.009 min
Response: 1306084
Conc: 505.24 ng/ml



#38 AR-1262-3


R.T.: 8.295 min
Delta R.T.: -0.006 min
Response: 589871
Conc: 347.92 ng/ml



#39 AR-1262-4


R.T.: 9.344 min
Delta R.T.: -0.005 min
Response: 408241
Conc: 198.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#39 AR-1262-4

R.T.: 8.357 min
Delta R.T.: -0.007 min
Response: 1863916
Conc: 586.96 ng/ml




#40 AR-1262-5

R.T.: 9.974 min
Delta R.T.: -0.004 min
Response: 672919
Conc: 475.05 ng/ml



#40 AR-1262-5


R.T.: 8.855 min
Delta R.T.: -0.005 min
Response: 702897
Conc: 437.33 ng/ml



#41 AR-1268-1


R.T.: 9.272 min
Delta R.T.: 0.011 min
Response: 1306084
Conc: 299.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#41 AR-1268-1

R.T.: 8.295 min
Delta R.T.: -0.005 min
Response: 589871
Conc: 120.49 ng/ml




#42 AR-1268-2

R.T.: 9.344 min
Delta R.T.: -0.008 min
Response: 408241
Conc: 94.94 ng/ml



#42 AR-1268-2


R.T.: 8.357 min
Delta R.T.: -0.008 min
Response: 1863916
Conc: 391.91 ng/ml



#43 AR-1268-3


R.T.: 9.562 min
Delta R.T.: -0.006 min
Response: 70173
Conc: 19.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#43 AR-1268-3

R.T.: 8.568 min
Delta R.T.: -0.006 min
Response: 32406
Conc: 8.10 ng/ml




#44 AR-1268-4

R.T.: 9.974 min
Delta R.T.: -0.004 min
Response: 672919
Conc: 424.86 ng/ml



#44 AR-1268-4


R.T.: 8.855 min
Delta R.T.: -0.006 min
Response: 702897
Conc: 414.86 ng/ml



#45 AR-1268-5

R.T.: 10.363 min
Delta R.T.: -0.006 min
Response: 249501
Conc: 20.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#45 AR-1268-5

R.T.: 9.144 min
Delta R.T.: -0.007 min
Response: 207220
Conc: 15.45 ng/ml