

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110420\
 Data File : PP031170.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Nov 2020 9:57
 Operator : DD\AJ
 Sample : L4214-09
 Misc : AR1268 MDL 25 PPB
 ALS Vial : 73 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 11:15:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 04 08:24:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.799	4.073	858126	826362	12.848	12.432
2)	SA Decachlor...	10.680	9.393	632806	701582	15.827	16.802
Target Compounds							
9)	L2 AR-1221-2	0.000	4.428	0	9763	N.D.	23.533 #
20)	L4 AR-1242-5	0.000	6.235f	0	5117	N.D.	5.140 #
25)	L5 AR-1248-5	0.000	6.235f	0	5117	N.D.	4.052 #
26)	L6 AR-1254-1	0.000	6.235f	0	5117	N.D.	2.602 #
34)	L7 AR-1260-4	0.000	7.758	0	29987	N.D.	16.436 #
35)	L7 AR-1260-5	0.000	8.007	0	14981	N.D.	3.329 #
36)	L8 AR-1262-1	0.000	7.497	0	27410	N.D.	11.801 #
37)	L8 AR-1262-2	0.000	8.007	0	14981	N.D.	3.614 #
38)	L8 AR-1262-3	9.256	8.293	109182	123069	42.235	72.589 #
39)	L8 AR-1262-4	9.347	8.358	99788	112845	48.468	35.536 #
40)	L8 AR-1262-5	0.000	8.853	0	48332	N.D.	30.071 #
41)	L9 AR-1268-1	9.256	8.293	109182	123069	24.995	25.139
42)	L9 AR-1268-2	9.347	8.358	99788	112845	23.206	23.727
43)	L9 AR-1268-3	9.563	8.566	92852	98618	25.674	24.646
44)	L9 AR-1268-4	0.000	8.853	0	48332	N.D.	28.526 #
45)	L9 AR-1268-5	10.363	9.142	290672	310446	24.306	23.149



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110420\
Data File : PP031170.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Nov 2020 9:57
Operator : DD\AJ
Sample : L4214-09
Misc : AR1268 MDL 25 PPB
ALS Vial : 73 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 11:15:54 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 04 08:24:38 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.799 min
Delta R.T.: -0.002 min
Response: 858126
Conc: 12.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.073 min
Delta R.T.: -0.002 min
Response: 826362
Conc: 12.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.680 min
Delta R.T.: -0.008 min
Response: 632806
Conc: 15.83 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.393 min
Delta R.T.: -0.011 min
Response: 701582
Conc: 16.80 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 5.144 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.428 min
Delta R.T.: -0.008 min
Response: 9763
Conc: 23.53 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T.: 7.081 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 6.235 min
Delta R.T.: 0.022 min
Response: 5117
Conc: 5.14 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 7.080 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.235 min
Delta R.T.: -0.022 min
Response: 5117
Conc: 4.05 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 7.013 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 6.235 min
Delta R.T.: 0.022 min
Response: 5117
Conc: 2.60 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T.: 8.648 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#34 AR-1260-4

R.T.: 7.758 min
Delta R.T.: -0.009 min
Response: 29987
Conc: 16.44 ng/ml




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T.: 8.962 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.007 min
Delta R.T.: -0.007 min
Response: 14981
Conc: 3.33 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 8.419 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#36 AR-1262-1

R.T.: 7.497 min
Delta R.T.: -0.007 min
Response: 27410
Conc: 11.80 ng/ml




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T.: 8.963 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.007 min
Delta R.T.: -0.006 min
Response: 14981
Conc: 3.61 ng/ml



#38 AR-1262-3


R.T.: 9.256 min
Delta R.T.: -0.007 min
Response: 109182
Conc: 42.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#38 AR-1262-3

R.T.: 8.293 min
Delta R.T.: -0.008 min
Response: 123069
Conc: 72.59 ng/ml




#39 AR-1262-4

R.T.: 9.347 min
Delta R.T.: -0.002 min
Response: 99788
Conc: 48.47 ng/ml



#39 AR-1262-4


R.T.: 8.358 min
Delta R.T.: -0.006 min
Response: 112845
Conc: 35.54 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 9.978 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#40 AR-1262-5

R.T.: 8.853 min
Delta R.T.: -0.007 min
Response: 48332
Conc: 30.07 ng/ml




#41 AR-1268-1

R.T.: 9.256 min
Delta R.T.: -0.005 min
Response: 109182
Conc: 24.99 ng/ml



#41 AR-1268-1


R.T.: 8.293 min
Delta R.T.: -0.007 min
Response: 123069
Conc: 25.14 ng/ml



#42 AR-1268-2


R.T.: 9.347 min
Delta R.T.: -0.005 min
Response: 99788
Conc: 23.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#42 AR-1268-2

R.T.: 8.358 min
Delta R.T.: -0.008 min
Response: 112845
Conc: 23.73 ng/ml




#43 AR-1268-3

R.T.: 9.563 min
Delta R.T.: -0.004 min
Response: 92852
Conc: 25.67 ng/ml



#43 AR-1268-3


R.T.: 8.566 min
Delta R.T.: -0.008 min
Response: 98618
Conc: 24.65 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 9.978 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#44 AR-1268-4

R.T.: 8.853 min
Delta R.T.: -0.008 min
Response: 48332
Conc: 28.53 ng/ml



#45 AR-1268-5

R.T.: 10.363 min
Delta R.T.: -0.006 min
Response: 290672
Conc: 24.31 ng/ml



#45 AR-1268-5

R.T.: 9.142 min
Delta R.T.: -0.009 min
Response: 310446
Conc: 23.15 ng/ml