Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110420\
Data File : PP031081.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Nov 2020 18:36
Operator : DD\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument : ECD_P
ClientSampleId : AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 08:04:39 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 04 08:04:22 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.801	4.077	2438989	2458389	50.000	50.000
2)	SA Decachlor...	10.687	9.403	1507079	1646512	50.000	50.000
	Target Compounds						
26)	L6 AR-1254-1	7.013	6.213	701998	983348	500.000	500.000
27)	L6 AR-1254-2	7.242	6.372	1042233	821613	500.000	500.000
28)	L6 AR-1254-3	7.620	6.793	1154020	1369161	500.000	500.000
29)	L6 AR-1254-4	7.914	7.032	818595	806806	500.000	500.000
30)	L6 AR-1254-5	8.341	7.462	844061	1094743	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110420\
Data File : PP031081.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Nov 2020 18:36
Operator : DD\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 08:04:39 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 04 08:04:22 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.801 min
Delta R.T.: 0.000 min
Response: 2438989
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500



#1 Tetrachloro-m-xylene

R.T.: 4.077 min
Delta R.T.: 0.000 min
Response: 2458389
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.687 min
Delta R.T.: 0.000 min
Response: 1507079
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.403 min
Delta R.T.: 0.000 min
Response: 1646512
Conc: 50.00 ng/ml



#26 AR-1254-1

R.T.: 7.013 min
Delta R.T.: 0.000 min
Response: 701998
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500



#26 AR-1254-1

R.T.: 6.213 min
Delta R.T.: 0.000 min
Response: 983348
Conc: 500.00 ng/ml



#27 AR-1254-2

R.T.: 7.242 min
Delta R.T.: 0.000 min
Response: 1042233
Conc: 500.00 ng/ml



#27 AR-1254-2

R.T.: 6.372 min
Delta R.T.: 0.000 min
Response: 821613
Conc: 500.00 ng/ml



#28 AR-1254-3

R.T.: 7.620 min
Delta R.T.: 0.000 min
Response: 1154020
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500



#28 AR-1254-3

R.T.: 6.793 min
Delta R.T.: 0.000 min
Response: 1369161
Conc: 500.00 ng/ml



#29 AR-1254-4

R.T.: 7.914 min
Delta R.T.: 0.000 min
Response: 818595
Conc: 500.00 ng/ml



#29 AR-1254-4

R.T.: 7.032 min
Delta R.T.: 0.000 min
Response: 806806
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 8.341 min
Delta R.T.: 0.000 min
Response: 844061
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 7.462 min
Delta R.T.: 0.000 min
Response: 1094743
Conc: 500.00 ng/ml