

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110421\
 Data File : PP040697.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Nov 2021 21:59
 Operator : AJ\MA
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 11/04/2021
 Supervised By :mohammad ahmed 11/05/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 11:19:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 04 11:19:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.765	3.768	1075059	1281116	50.000	50.000
2) SA Decachlor...	10.670	9.037	1116968	863659	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.392f	7.134	626514	288135	500.000	500.000
37) L8 AR-1262-2	8.937f	7.688	1136210	890293	500.000	500.000
38) L8 AR-1262-3	9.234f	7.973	531603	387675	532.166m	500.000
39) L8 AR-1262-4	9.308	8.037	293192	700765	500.000	500.000
40) L8 AR-1262-5	9.960f	8.537	455297	347211	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110421\
Data File : PP040697.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Nov 2021 21:59
Operator : AJ\MA
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/04/2021
Supervised By :mohammad ahmed 11/05/2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 11:19:59 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110421.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 04 11:19:41 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 1075059
Conc: 50.00 ng/ml

Instrument :

ECD_P


ClientSampleId :

AR1262ICC500

Manual Integrations


APPROVED

Reviewed By :Ankita Jodhani 11/04/2021
Supervised By :mohammad ahmed 11/05/2021




#1 Tetrachloro-m-xylene

R.T.: 3.768 min
Delta R.T.: 0.000 min
Response: 1281116
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.670 min
Delta R.T.: 0.002 min
Response: 1116968
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.037 min
Delta R.T.: 0.002 min
Response: 863659
Conc: 50.00 ng/ml




#36 AR-1262-1

R.T.: 8.392 min
Delta R.T.: 0.018 min
Response: 626514
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/04/2021
Supervised By :mohammad ahmed 11/05/2021




#36 AR-1262-1

R.T.: 7.134 min
Delta R.T.: -0.001 min
Response: 288135
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 8.937 min
Delta R.T.: 0.019 min
Response: 1136210
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 890293
Conc: 500.00 ng/ml




#38 AR-1262-3

R.T.: 9.234 min
Delta R.T.: 0.020 min
Response: 531603
Conc: 532.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/04/2021
Supervised By :mohammad ahmed 11/05/2021




#38 AR-1262-3

R.T.: 7.973 min
Delta R.T.: 0.000 min
Response: 387675
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 9.308 min
Delta R.T.: 0.004 min
Response: 293192
Conc: 500.00 ng/ml



#39 AR-1262-4

R.T.: 8.037 min
Delta R.T.: 0.000 min
Response: 700765
Conc: 500.00 ng/ml




#40 AR-1262-5

R.T.: 9.960 min
Delta R.T.: 0.023 min
Response: 455297
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/04/2021
Supervised By :mohammad ahmed 11/05/2021



#40 AR-1262-5

R.T.: 8.537 min
Delta R.T.: 0.000 min
Response: 347211
Conc: 500.00 ng/ml