

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110422\
 Data File : PP053065.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2022 09:56
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 11:53:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 27 04:46:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.404	3.649	98328177	88756323	44.498	47.230
2) SA Decachlor...	10.208	8.730	69649997	81182711	54.872	48.406
Target Compounds						
16) L4 AR-1242-1	5.577	4.748	23417520	23734441	416.084	425.292
17) L4 AR-1242-2	5.600	4.767	33558900	32814201	406.244	425.455
18) L4 AR-1242-3	5.663	4.946	21098988	17811935	414.865	434.140
19) L4 AR-1242-4	5.762	5.030	17001463	18513455	419.690	436.504
20) L4 AR-1242-5	6.502	5.561	17152652	22743813	430.237	449.939



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110422\
Data File : PP053065.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2022 09:56
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 11:53:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 27 04:46:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.404 min
Delta R.T.: -0.006 min
Response: 98328177
Conc: 44.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.649 min
Delta R.T.: -0.005 min
Response: 88756323
Conc: 47.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.208 min
Delta R.T.: -0.016 min
Response: 69649997
Conc: 54.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.730 min
Delta R.T.: -0.013 min
Response: 81182711
Conc: 48.41 ng/ml



#16 AR-1242-1


R.T.: 5.577 min
Delta R.T.: -0.009 min
Response: 23417520
Conc: 416.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.748 min
Delta R.T.: -0.009 min
Response: 23734441
Conc: 425.29 ng/ml




#17 AR-1242-2

R.T.: 5.600 min
Delta R.T.: -0.008 min
Response: 33558900
Conc: 406.24 ng/ml



#17 AR-1242-2


R.T.: 4.767 min
Delta R.T.: -0.009 min
Response: 32814201
Conc: 425.45 ng/ml



#18 AR-1242-3


R.T.: 5.663 min
Delta R.T.: -0.008 min
Response: 21098988
Conc: 414.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.946 min
Delta R.T.: -0.010 min
Response: 17811935
Conc: 434.14 ng/ml




#19 AR-1242-4

R.T.: 5.762 min
Delta R.T.: -0.009 min
Response: 17001463
Conc: 419.69 ng/ml



#19 AR-1242-4


R.T.: 5.030 min
Delta R.T.: -0.009 min
Response: 18513455
Conc: 436.50 ng/ml



#20 AR-1242-5

R.T.: 6.502 min
Delta R.T.: -0.009 min
Response: 17152652
Conc: 430.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.561 min
Delta R.T.: -0.010 min
Response: 22743813
Conc: 449.94 ng/ml