

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110422\
 Data File : PP053090.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2022 19:55
 Operator : YP\AJ
 Sample : N5417-06
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 40183

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/07/2022
 Supervised By :Ankita Jodhani 11/07/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 07:36:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 27 04:46:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.406	3.651	42508629	32731657	19.237	17.418m
2) SA Decachlor...	10.209	8.732	33375704	33455467	26.294	19.948

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110422\
Data File : PP053090.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2022 19:55
Operator : YP\AJ
Sample : N5417-06
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
40183


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/07/2022
Supervised By :Ankita Jodhani 11/07/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 07:36:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102622.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 27 04:46:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.406 min
Delta R.T.: -0.004 min
Response: 42508629
Conc: 19.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
40183


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/07/2022
Supervised By :Ankita Jodhani 11/07/2022




#1 Tetrachloro-m-xylene

R.T.: 3.651 min
Delta R.T.: -0.002 min
Response: 32731657
Conc: 17.42 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.209 min
Delta R.T.: -0.015 min
Response: 33375704
Conc: 26.29 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: -0.011 min
Response: 33455467
Conc: 19.95 ng/ml