

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110525\
 Data File : PP076326.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Nov 2025 10:53
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/06/2025
 Supervised By :Yogesh Patel 11/06/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 01:16:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.640	3.779	78321870	379.6E6	54.805	57.358
2) SA Decachlor...	10.410	8.773	58106699	382.0E6	53.976	46.452
Target Compounds						
21) L5 AR-1248-1	5.792	4.872	19719505	181.5E6	549.740	483.190
22) L5 AR-1248-2	6.063	5.091	26507921	125.7E6	543.815	511.268
23) L5 AR-1248-3	6.266	5.131	28822086	142.0E6	531.138	494.949m
24) L5 AR-1248-4	6.664	5.304	34306808	145.7E6	516.948	515.333
25) L5 AR-1248-5	6.703	5.693	33831641	293.7E6	521.605	565.268

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110525\
Data File : PP076326.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Nov 2025 10:53
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/06/2025
Supervised By :Yogesh Patel 11/06/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 01:16:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.640 min
Delta R.T.: -0.002 min
Response: 78321870
Conc: 54.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/06/2025
Supervised By :Yogesh Patel 11/06/2025




#1 Tetrachloro-m-xylene

R.T.: 3.779 min
Delta R.T.: -0.003 min
Response: 379608522
Conc: 57.36 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.410 min
Delta R.T.: -0.009 min
Response: 58106699
Conc: 53.98 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.773 min
Delta R.T.: -0.009 min
Response: 381976889
Conc: 46.45 ng/ml




#21 AR-1248-1

R.T.: 5.792 min
Delta R.T.: -0.008 min
Response: 19719505
Conc: 549.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/06/2025
Supervised By :Yogesh Patel 11/06/2025




#21 AR-1248-1

R.T.: 4.872 min
Delta R.T.: -0.004 min
Response: 181494277
Conc: 483.19 ng/ml




#22 AR-1248-2

R.T.: 6.063 min
Delta R.T.: -0.008 min
Response: 26507921
Conc: 543.81 ng/ml



#22 AR-1248-2

R.T.: 5.091 min
Delta R.T.: -0.004 min
Response: 125724937
Conc: 511.27 ng/ml




#23 AR-1248-3

R.T.: 6.266 min
Delta R.T.: -0.008 min
Response: 28822086
Conc: 531.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/06/2025
Supervised By :Yogesh Patel 11/06/2025




#23 AR-1248-3

R.T.: 5.131 min
Delta R.T.: -0.005 min
Response: 141997411
Conc: 494.95 ng/ml m




#24 AR-1248-4

R.T.: 6.664 min
Delta R.T.: -0.008 min
Response: 34306808
Conc: 516.95 ng/ml



#24 AR-1248-4

R.T.: 5.304 min
Delta R.T.: -0.005 min
Response: 145690406
Conc: 515.33 ng/ml




#25 AR-1248-5

R.T.: 6.703 min
Delta R.T.: -0.008 min
Response: 33831641
Conc: 521.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/06/2025
Supervised By :Yogesh Patel 11/06/2025



#25 AR-1248-5

R.T.: 5.693 min
Delta R.T.: -0.005 min
Response: 293746634
Conc: 565.27 ng/ml