

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110620\
 Data File : PP031183.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Nov 2020 9:16
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 01:32:02 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 05 13:03:18 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.799	4.073	3142625	3140038	47.051	47.241
2)	SA Decachlor...	10.682	9.394	2017673	2206078	50.462	52.832
Target Compounds							
3)	L1 AR-1016-1	6.105	5.334	915768	883608	470.328	476.268
4)	L1 AR-1016-2	6.128	5.354	1363715	1292484	477.701	475.494
5)	L1 AR-1016-3	6.194	5.546	844988	715387	485.135	564.812
6)	L1 AR-1016-4	6.299	5.598	708188	519383	490.107	563.173
7)	L1 AR-1016-5	6.614	5.827	647714	663004	519.591	505.134
8)	L2 AR-1221-1	5.042	4.332	99118	92219	169.292	160.949
9)	L2 AR-1221-2	5.143	4.431	151789	141483	354.043	330.440
10)	L2 AR-1221-3	5.228	4.520	544547	532192	402.595	403.928
11)	L3 AR-1232-1	5.228	4.520	544547	532192	445.250	444.286
12)	L3 AR-1232-2	5.812	5.354	709677	1292484	1168.706	1216.230
13)	L3 AR-1232-3	6.128	5.546	1363715	715387	1220.739	1395.341
14)	L3 AR-1232-4	6.299	5.642	708188	634178	1247.801	1452.028
15)	L3 AR-1232-5	6.399	5.827	506058	663004	1460.444	1372.980
16)	L4 AR-1242-1	6.105	5.334	915768	883608	681.633	686.601
17)	L4 AR-1242-2	6.128	5.354	1363715	1292484	698.267	692.149
18)	L4 AR-1242-3	6.194	5.546	844988	715387	692.858	761.977
19)	L4 AR-1242-4	6.299	5.642	708188	634178	699.021	731.291
20)	L4 AR-1242-5	7.079	6.207	98574	469069	103.298	471.185 #
21)	L5 AR-1248-1	6.105	5.334	915768	883608	949.670	954.939
22)	L5 AR-1248-2	6.399	5.598	506058	519383	437.545	471.265
23)	L5 AR-1248-3	6.614	5.642	647714	634178	455.243	523.964
24)	L5 AR-1248-4	7.040	5.827	118619	663004	77.829	449.242 #
25)	L5 AR-1248-5	7.079	6.250	98574	87762	64.976	69.490
26)	L6 AR-1254-1	7.010	6.207	412357	469069	293.702	238.506
27)	L6 AR-1254-2	7.239	6.366	440021	418048	211.095	254.407
28)	L6 AR-1254-3	7.617	6.806	275598	803935	119.408	293.587 #
29)	L6 AR-1254-4	7.911	7.027	197748	109806	120.785	68.050 #
30)	L6 AR-1254-5	8.338	7.455	1337809	1423225	792.483	650.026
31)	L7 AR-1260-1	7.786	6.924	950676	1042667	517.561	520.248
32)	L7 AR-1260-2	8.051	7.120	1121859	1232654	520.749	526.367
33)	L7 AR-1260-3	8.417	7.276	876065	1204671	532.216	529.876
34)	L7 AR-1260-4	8.644	7.759	1016105	1020936	546.483	559.583
35)	L7 AR-1260-5	8.959	8.006	2012971	2488273	533.882	552.904
36)	L8 AR-1262-1	8.417	7.496	876065	1108249	436.024	477.133
37)	L8 AR-1262-2	8.959	8.006	2012971	2488273	553.836	600.197
38)	L8 AR-1262-3	9.273	8.293	1347151	567674	521.128	334.826 #
39)	L8 AR-1262-4	9.346	8.356	426082	1896976	206.953	597.368 #
40)	L8 AR-1262-5	9.975	8.853	651983	720630	460.271	448.367
41)	L9 AR-1268-1	9.273	8.293	1347151	567674	308.399	115.957 #
42)	L9 AR-1268-2	9.346	8.356	426082	1896976	99.085	398.864 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110620\
 Data File : PP031183.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Nov 2020 9:16
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 01:32:02 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 05 13:03:18 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	0.000	8.566	0	35063	N.D.	8.763 #
44) L9	AR-1268-4	9.975	8.853	651983	720630	411.638	425.329
45) L9	AR-1268-5	10.364	9.142	189107	205493	15.813	15.323



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110620\
Data File : PP031183.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Nov 2020 9:16
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 01:32:02 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 05 13:03:18 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.799 min
Delta R.T.: -0.002 min
Response: 3142625
Conc: 47.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.073 min
Delta R.T.: -0.002 min
Response: 3140038
Conc: 47.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.682 min
Delta R.T.: -0.006 min
Response: 2017673
Conc: 50.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.394 min
Delta R.T.: -0.010 min
Response: 2206078
Conc: 52.83 ng/ml



#3 AR-1016-1


R.T.: 6.105 min
Delta R.T.: -0.003 min
Response: 915768
Conc: 470.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.334 min
Delta R.T.: -0.005 min
Response: 883608
Conc: 476.27 ng/ml




#4 AR-1016-2

R.T.: 6.128 min
Delta R.T.: -0.002 min
Response: 1363715
Conc: 477.70 ng/ml



#4 AR-1016-2


R.T.: 5.354 min
Delta R.T.: -0.005 min
Response: 1292484
Conc: 475.49 ng/ml



#5 AR-1016-3


R.T.: 6.194 min
Delta R.T.: -0.002 min
Response: 844988
Conc: 485.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.546 min
Delta R.T.: -0.006 min
Response: 715387
Conc: 564.81 ng/ml




#6 AR-1016-4

R.T.: 6.299 min
Delta R.T.: -0.003 min
Response: 708188
Conc: 490.11 ng/ml



#6 AR-1016-4


R.T.: 5.598 min
Delta R.T.: -0.006 min
Response: 519383
Conc: 563.17 ng/ml



#7 AR-1016-5


R.T.: 6.614 min
Delta R.T.: -0.003 min
Response: 647714
Conc: 519.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.827 min
Delta R.T.: -0.006 min
Response: 663004
Conc: 505.13 ng/ml




#8 AR-1221-1

R.T.: 5.042 min
Delta R.T.: -0.002 min
Response: 99118
Conc: 169.29 ng/ml



#8 AR-1221-1


R.T.: 4.332 min
Delta R.T.: -0.001 min
Response: 92219
Conc: 160.95 ng/ml



#9 AR-1221-2


R.T.: 5.143 min
Delta R.T.: 0.000 min
Response: 151789
Conc: 354.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#9 AR-1221-2

R.T.: 4.431 min
Delta R.T.: 0.000 min
Response: 141483
Conc: 330.44 ng/ml




#10 AR-1221-3

R.T.: 5.228 min
Delta R.T.: 0.001 min
Response: 544547
Conc: 402.60 ng/ml



#10 AR-1221-3


R.T.: 4.520 min
Delta R.T.: 0.000 min
Response: 532192
Conc: 403.93 ng/ml



#11 AR-1232-1


R.T.: 5.228 min
Delta R.T.: -0.002 min
Response: 544547
Conc: 445.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#11 AR-1232-1

R.T.: 4.520 min
Delta R.T.: -0.004 min
Response: 532192
Conc: 444.29 ng/ml




#12 AR-1232-2

R.T.: 5.812 min
Delta R.T.: -0.002 min
Response: 709677
Conc: 1168.71 ng/ml



#12 AR-1232-2


R.T.: 5.354 min
Delta R.T.: -0.005 min
Response: 1292484
Conc: 1216.23 ng/ml



#13 AR-1232-3


R.T.: 6.128 min
Delta R.T.: -0.002 min
Response: 1363715
Conc: 1220.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#13 AR-1232-3

R.T.: 5.546 min
Delta R.T.: -0.005 min
Response: 715387
Conc: 1395.34 ng/ml




#14 AR-1232-4

R.T.: 6.299 min
Delta R.T.: -0.003 min
Response: 708188
Conc: 1247.80 ng/ml



#14 AR-1232-4


R.T.: 5.642 min
Delta R.T.: -0.006 min
Response: 634178
Conc: 1452.03 ng/ml



#15 AR-1232-5


R.T.: 6.399 min
Delta R.T.: -0.002 min
Response: 506058
Conc: 1460.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#15 AR-1232-5

R.T.: 5.827 min
Delta R.T.: -0.006 min
Response: 663004
Conc: 1372.98 ng/ml




#16 AR-1242-1

R.T.: 6.105 min
Delta R.T.: -0.002 min
Response: 915768
Conc: 681.63 ng/ml



#16 AR-1242-1


R.T.: 5.334 min
Delta R.T.: -0.006 min
Response: 883608
Conc: 686.60 ng/ml



#17 AR-1242-2


R.T.: 6.128 min
Delta R.T.: -0.002 min
Response: 1363715
Conc: 698.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#17 AR-1242-2

R.T.: 5.354 min
Delta R.T.: -0.005 min
Response: 1292484
Conc: 692.15 ng/ml




#18 AR-1242-3

R.T.: 6.194 min
Delta R.T.: -0.002 min
Response: 844988
Conc: 692.86 ng/ml



#18 AR-1242-3


R.T.: 5.546 min
Delta R.T.: -0.006 min
Response: 715387
Conc: 761.98 ng/ml



#19 AR-1242-4


R.T.: 6.299 min
Delta R.T.: -0.003 min
Response: 708188
Conc: 699.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#19 AR-1242-4

R.T.: 5.642 min
Delta R.T.: -0.006 min
Response: 634178
Conc: 731.29 ng/ml




#20 AR-1242-5

R.T.: 7.079 min
Delta R.T.: -0.002 min
Response: 98574
Conc: 103.30 ng/ml



#20 AR-1242-5


R.T.: 6.207 min
Delta R.T.: -0.006 min
Response: 469069
Conc: 471.19 ng/ml



#21 AR-1248-1


R.T.: 6.105 min
Delta R.T.: -0.003 min
Response: 915768
Conc: 949.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#21 AR-1248-1

R.T.: 5.334 min
Delta R.T.: -0.006 min
Response: 883608
Conc: 954.94 ng/ml




#22 AR-1248-2

R.T.: 6.399 min
Delta R.T.: -0.003 min
Response: 506058
Conc: 437.55 ng/ml



#22 AR-1248-2


R.T.: 5.598 min
Delta R.T.: -0.006 min
Response: 519383
Conc: 471.27 ng/ml



#23 AR-1248-3


R.T.: 6.614 min
Delta R.T.: -0.002 min
Response: 647714
Conc: 455.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#23 AR-1248-3

R.T.: 5.642 min
Delta R.T.: -0.006 min
Response: 634178
Conc: 523.96 ng/ml




#24 AR-1248-4

R.T.: 7.040 min
Delta R.T.: -0.002 min
Response: 118619
Conc: 77.83 ng/ml



#24 AR-1248-4


R.T.: 5.827 min
Delta R.T.: -0.006 min
Response: 663004
Conc: 449.24 ng/ml



#25 AR-1248-5


R.T.: 7.079 min
Delta R.T.: -0.002 min
Response: 98574
Conc: 64.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#25 AR-1248-5

R.T.: 6.250 min
Delta R.T.: -0.006 min
Response: 87762
Conc: 69.49 ng/ml




#26 AR-1254-1

R.T.: 7.010 min
Delta R.T.: -0.003 min
Response: 412357
Conc: 293.70 ng/ml



#26 AR-1254-1


R.T.: 6.207 min
Delta R.T.: -0.006 min
Response: 469069
Conc: 238.51 ng/ml



#27 AR-1254-2


R.T.: 7.239 min
Delta R.T.: -0.003 min
Response: 440021
Conc: 211.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#27 AR-1254-2

R.T.: 6.366 min
Delta R.T.: -0.006 min
Response: 418048
Conc: 254.41 ng/ml




#28 AR-1254-3

R.T.: 7.617 min
Delta R.T.: -0.003 min
Response: 275598
Conc: 119.41 ng/ml



#28 AR-1254-3


R.T.: 6.806 min
Delta R.T.: 0.013 min
Response: 803935
Conc: 293.59 ng/ml



#29 AR-1254-4


R.T.: 7.911 min
Delta R.T.: -0.003 min
Response: 197748
Conc: 120.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#29 AR-1254-4

R.T.: 7.027 min
Delta R.T.: -0.005 min
Response: 109806
Conc: 68.05 ng/ml




#30 AR-1254-5

R.T.: 8.338 min
Delta R.T.: -0.003 min
Response: 1337809
Conc: 792.48 ng/ml



#30 AR-1254-5


R.T.: 7.455 min
Delta R.T.: -0.007 min
Response: 1423225
Conc: 650.03 ng/ml



#31 AR-1260-1


R.T.: 7.786 min
Delta R.T.: -0.003 min
Response: 950676
Conc: 517.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#31 AR-1260-1

R.T.: 6.924 min
Delta R.T.: -0.007 min
Response: 1042667
Conc: 520.25 ng/ml




#32 AR-1260-2

R.T.: 8.051 min
Delta R.T.: -0.004 min
Response: 1121859
Conc: 520.75 ng/ml



#32 AR-1260-2


R.T.: 7.120 min
Delta R.T.: -0.007 min
Response: 1232654
Conc: 526.37 ng/ml



#33 AR-1260-3


R.T.: 8.417 min
Delta R.T.: -0.003 min
Response: 876065
Conc: 532.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#33 AR-1260-3

R.T.: 7.276 min
Delta R.T.: -0.007 min
Response: 1204671
Conc: 529.88 ng/ml




#34 AR-1260-4

R.T.: 8.644 min
Delta R.T.: -0.003 min
Response: 1016105
Conc: 546.48 ng/ml



#34 AR-1260-4


R.T.: 7.759 min
Delta R.T.: -0.008 min
Response: 1020936
Conc: 559.58 ng/ml



#35 AR-1260-5


R.T.: 8.959 min
Delta R.T.: -0.003 min
Response: 2012971
Conc: 533.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#35 AR-1260-5

R.T.: 8.006 min
Delta R.T.: -0.008 min
Response: 2488273
Conc: 552.90 ng/ml




#36 AR-1262-1

R.T.: 8.417 min
Delta R.T.: -0.002 min
Response: 876065
Conc: 436.02 ng/ml



#36 AR-1262-1


R.T.: 7.496 min
Delta R.T.: -0.007 min
Response: 1108249
Conc: 477.13 ng/ml



#37 AR-1262-2


R.T.: 8.959 min
Delta R.T.: -0.003 min
Response: 2012971
Conc: 553.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#37 AR-1262-2

R.T.: 8.006 min
Delta R.T.: -0.007 min
Response: 2488273
Conc: 600.20 ng/ml




#38 AR-1262-3

R.T.: 9.273 min
Delta R.T.: 0.011 min
Response: 1347151
Conc: 521.13 ng/ml



#38 AR-1262-3


R.T.: 8.293 min
Delta R.T.: -0.008 min
Response: 567674
Conc: 334.83 ng/ml



#39 AR-1262-4


R.T.: 9.346 min
Delta R.T.: -0.003 min
Response: 426082
Conc: 206.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#39 AR-1262-4

R.T.: 8.356 min
Delta R.T.: -0.008 min
Response: 1896976
Conc: 597.37 ng/ml




#40 AR-1262-5

R.T.: 9.975 min
Delta R.T.: -0.003 min
Response: 651983
Conc: 460.27 ng/ml



#40 AR-1262-5


R.T.: 8.853 min
Delta R.T.: -0.007 min
Response: 720630
Conc: 448.37 ng/ml



#41 AR-1268-1


R.T.: 9.273 min
Delta R.T.: 0.013 min
Response: 1347151
Conc: 308.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#41 AR-1268-1

R.T.: 8.293 min
Delta R.T.: -0.007 min
Response: 567674
Conc: 115.96 ng/ml




#42 AR-1268-2

R.T.: 9.346 min
Delta R.T.: -0.006 min
Response: 426082
Conc: 99.09 ng/ml



#42 AR-1268-2


R.T.: 8.356 min
Delta R.T.: -0.010 min
Response: 1896976
Conc: 398.86 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.568 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#43 AR-1268-3

R.T.: 8.566 min
Delta R.T.: -0.008 min
Response: 35063
Conc: 8.76 ng/ml




#44 AR-1268-4

R.T.: 9.975 min
Delta R.T.: -0.003 min
Response: 651983
Conc: 411.64 ng/ml



#44 AR-1268-4


R.T.: 8.853 min
Delta R.T.: -0.008 min
Response: 720630
Conc: 425.33 ng/ml



#45 AR-1268-5

R.T.: 10.364 min
Delta R.T.: -0.005 min
Response: 189107
Conc: 15.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500



#45 AR-1268-5

R.T.: 9.142 min
Delta R.T.: -0.009 min
Response: 205493
Conc: 15.32 ng/ml