

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110623\
 Data File : PP061461.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2023 17:50
 Operator : YP\AJ
 Sample : 05255-02MS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 P001-WC01-01MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/07/2023
 Supervised By :Ankita Jodhani 11/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 23:16:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 27 18:44:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.572	3.712	19226470	19173179	7.783m	12.829m#
2)	SA Decachlor...	10.558	8.778	36655067	32520442	19.438	21.176
Target Compounds							
3)	L1 AR-1016-1	5.763	4.806	92440945	30817802	1135.936	595.462 #
4)	L1 AR-1016-2	5.788	4.825	92552963	50769309	795.905m	717.838m
5)	L1 AR-1016-3	5.852	5.006f	50231326	35141299	710.319m	890.742 #
6)	L1 AR-1016-4	5.954	5.049f	43058835	63262413	737.669	2011.108 #
7)	L1 AR-1016-5	6.251	5.261	90275201	58623982	1572.476	1394.537m
26)	L6 AR-1254-1	6.636f	5.618f	218.7E6	155.8E6	2468.863	1976.966m
27)	L6 AR-1254-2	6.863f	5.776f	409.0E6	129.8E6	3066.443	1858.494m#
28)	L6 AR-1254-3	7.231f	6.176f	534.5E6	185.5E6	3864.495	1634.841m#
29)	L6 AR-1254-4	7.517f	6.404f	202.6E6	112.4E6	2035.669m	1617.638
30)	L6 AR-1254-5	7.940f	6.823	553.6E6	532.8E6	5030.757	5035.761
31)	L7 AR-1260-1	7.395	6.304	319.4E6	219.1E6	2839.615	2692.850
32)	L7 AR-1260-2	7.646	6.494	1072.3E6	229.5E6	8288.613	2352.906 #
33)	L7 AR-1260-3	8.017	6.648	186.4E6	243.2E6	2195.119	2627.678
34)	L7 AR-1260-4	8.252	7.123	341.4E6	122.4E6	3377.519	1669.792 #
35)	L7 AR-1260-5	8.592	7.367	488.2E6	371.9E6	2520.442	2178.141

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110623\
Data File : PP061461.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2023 17:50
Operator : YP\AJ
Sample : 05255-02MS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
P001-WC01-01MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/07/2023
Supervised By :Ankita Jodhani 11/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 23:16:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 27 18:44:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

