

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110623\
 Data File : PP061442.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2023 10:39
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 23:08:59 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 27 18:44:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.562	3.697	135.4E6	83475471	54.817	55.856
2) SA Decachlor...	10.540	8.766	87837161	66040764	46.581	43.004
Target Compounds						
21) L5 AR-1248-1	5.754	4.794	26129915	16296176	535.444	524.381
22) L5 AR-1248-2	6.033	5.034	38148027	22582109	506.221	517.694
23) L5 AR-1248-3	6.241	5.076	42120968	23918020	505.063	514.632
24) L5 AR-1248-4	6.650	5.249	44537081	28484147	505.229	511.176
25) L5 AR-1248-5	6.690	5.646	43364192	26535881	511.409	509.991



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110623\
Data File : PP061442.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2023 10:39
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 23:08:59 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 27 18:44:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.562 min
Delta R.T.: 0.003 min
Response: 135408580
Conc: 54.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.697 min
Delta R.T.: 0.003 min
Response: 83475471
Conc: 55.86 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.540 min
Delta R.T.: -0.005 min
Response: 87837161
Conc: 46.58 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.766 min
Delta R.T.: -0.004 min
Response: 66040764
Conc: 43.00 ng/ml



#21 AR-1248-1


R.T.: 5.754 min
Delta R.T.: 0.002 min
Response: 26129915
Conc: 535.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.794 min
Delta R.T.: 0.001 min
Response: 16296176
Conc: 524.38 ng/ml




#22 AR-1248-2

R.T.: 6.033 min
Delta R.T.: 0.002 min
Response: 38148027
Conc: 506.22 ng/ml



#22 AR-1248-2


R.T.: 5.034 min
Delta R.T.: 0.001 min
Response: 22582109
Conc: 517.69 ng/ml



#23 AR-1248-3


R.T.: 6.241 min
Delta R.T.: 0.001 min
Response: 42120968
Conc: 505.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.076 min
Delta R.T.: 0.002 min
Response: 23918020
Conc: 514.63 ng/ml




#24 AR-1248-4

R.T.: 6.650 min
Delta R.T.: 0.002 min
Response: 44537081
Conc: 505.23 ng/ml



#24 AR-1248-4


R.T.: 5.249 min
Delta R.T.: 0.001 min
Response: 28484147
Conc: 511.18 ng/ml



#25 AR-1248-5

R.T.: 6.690 min
Delta R.T.: 0.002 min
Response: 43364192
Conc: 511.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.646 min
Delta R.T.: 0.002 min
Response: 26535881
Conc: 509.99 ng/ml