

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110623\
 Data File : PP061483.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Nov 2023 00:10
 Operator : YP\AJ
 Sample : 05258-10
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 10-A-10-B-10-C

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/07/2023
 Supervised By :Ankita Jodhani 11/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 06:40:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 27 18:44:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.572	3.704	51860653	34126502	20.995	22.835
2) SA Decachlor...	10.567f	8.779	38448778	24113515	20.390m	15.702

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110623\
Data File : PP061483.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Nov 2023 00:10
Operator : YP\AJ
Sample : 05258-10
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
10-A-10-B-10-C


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/07/2023
Supervised By :Ankita Jodhani 11/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 06:40:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 27 18:44:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.572 min
Delta R.T.: 0.013 min
Response: 51860653
Conc: 20.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
10-A-10-B-10-C


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/07/2023
Supervised By :Ankita Jodhani 11/07/2023




#1 Tetrachloro-m-xylene

R.T.: 3.704 min
Delta R.T.: 0.011 min
Response: 34126502
Conc: 22.83 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.567 min
Delta R.T.: 0.031 min
Response: 38448778
Conc: 20.39 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.779 min
Delta R.T.: 0.016 min
Response: 24113515
Conc: 15.70 ng/ml