

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110723\
 Data File : PP061525.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Nov 2023 21:59
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 04:27:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 08 04:27:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.582	3.712	254.3E6	155.1E6	99.285	95.339
2) SA Decachlor...	10.590	8.795	331.5E6	252.1E6	95.566	95.803
Target Compounds						
41) L9 AR-1268-1	8.944	7.666	273.5E6	219.3E6	972.639	955.287
42) L9 AR-1268-2	9.047	7.731	250.3E6	199.5E6	971.906	972.457
43) L9 AR-1268-3	9.295	7.938	215.0E6	174.2E6	976.128	959.284
44) L9 AR-1268-4	9.758	8.232	91060901	72350741	965.282	953.519
45) L9 AR-1268-5	10.214	8.532	722.7E6	537.0E6	993.373	977.678



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110723\
Data File : PP061525.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Nov 2023 21:59
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 04:27:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110723.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 08 04:27:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.582 min
Delta R.T.: 0.003 min
Response: 254326330
Conc: 99.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.712 min
Delta R.T.: 0.003 min
Response: 155069924
Conc: 95.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.590 min
Delta R.T.: 0.012 min
Response: 331456315
Conc: 95.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.795 min
Delta R.T.: 0.006 min
Response: 252093483
Conc: 95.80 ng/ml



#41 AR-1268-1


R.T.: 8.944 min
Delta R.T.: 0.006 min
Response: 273501062
Conc: 972.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.666 min
Delta R.T.: 0.005 min
Response: 219345351
Conc: 955.29 ng/ml




#42 AR-1268-2

R.T.: 9.047 min
Delta R.T.: 0.007 min
Response: 250293551
Conc: 971.91 ng/ml



#42 AR-1268-2


R.T.: 7.731 min
Delta R.T.: 0.004 min
Response: 199508278
Conc: 972.46 ng/ml



#43 AR-1268-3


R.T.: 9.295 min
Delta R.T.: 0.009 min
Response: 214975307
Conc: 976.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.938 min
Delta R.T.: 0.005 min
Response: 174184039
Conc: 959.28 ng/ml




#44 AR-1268-4

R.T.: 9.758 min
Delta R.T.: 0.010 min
Response: 91060901
Conc: 965.28 ng/ml



#44 AR-1268-4


R.T.: 8.232 min
Delta R.T.: 0.006 min
Response: 72350741
Conc: 953.52 ng/ml



#45 AR-1268-5

R.T.: 10.214 min
Delta R.T.: 0.011 min
Response: 722652400
Conc: 993.37 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.532 min
Delta R.T.: 0.005 min
Response: 537044835
Conc: 977.68 ng/ml