

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110723\
 Data File : PP061501.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Nov 2023 15:27
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/09/2023
 Supervised By :Ankita Jodhani 11/09/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 15:59:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 07 15:47:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.586	3.715	64786175	40924200	25.250	26.578m
2) SA Decachlor...	10.599	8.804	49519310	37738025	25.973	26.474
Target Compounds						
3) L1 AR-1016-1	5.785	4.822	21090625	14165862	262.302	278.462
4) L1 AR-1016-2	5.808	4.841	30581345	18992752	262.881	272.582
5) L1 AR-1016-3	5.872	5.019	18738930	10677405	264.167	273.213
6) L1 AR-1016-4	5.973	5.063	15256340	8557816	262.223	275.866
7) L1 AR-1016-5	6.274	5.279	15296760	11475346	266.837	281.190
31) L7 AR-1260-1	7.418	6.325	26718387	21016423	255.577	272.065
32) L7 AR-1260-2	7.678	6.515	31807445	25435853	262.745	275.138
33) L7 AR-1260-3	8.041	6.669	21407645	24253421	260.269	270.466
34) L7 AR-1260-4	8.277	7.145	26232616	18760479	262.961	269.244
35) L7 AR-1260-5	8.618	7.388	48615967	42923193	259.376	263.353

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110723\
Data File : PP061501.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Nov 2023 15:27
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/09/2023
Supervised By :Ankita Jodhani 11/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 15:59:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110723.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 07 15:47:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

