

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110725\
 Data File : PP076389.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Nov 2025 15:31
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/10/2025
 Supervised By :Yogesh Patel 11/10/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 10 01:19:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.638	3.776	78597225	381.7E6	54.998	57.667
2) SA Decachlor...	10.409	8.771	57448293	425.5E6	53.364	51.749
Target Compounds						
21) L5 AR-1248-1	5.788	4.868	18989577	179.4E6	529.391m	477.509m
22) L5 AR-1248-2	6.059	5.090	24513555	115.9E6	502.900m	471.330
23) L5 AR-1248-3	6.262	5.130	26950703	145.5E6	496.652m	507.022m
24) L5 AR-1248-4	6.660	5.302	32416705	140.5E6	488.467m	497.126
25) L5 AR-1248-5	6.699	5.691	30868867	294.2E6	475.926m	566.176

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110725\
Data File : PP076389.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Nov 2025 15:31
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/10/2025
Supervised By :Yogesh Patel 11/10/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 01:19:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.638 min
Delta R.T.: -0.004 min
Response: 78597225
Conc: 55.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/10/2025
Supervised By :Yogesh Patel 11/10/2025




#1 Tetrachloro-m-xylene

R.T.: 3.776 min
Delta R.T.: -0.005 min
Response: 381657041
Conc: 57.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.409 min
Delta R.T.: -0.009 min
Response: 57448293
Conc: 53.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.771 min
Delta R.T.: -0.011 min
Response: 425531301
Conc: 51.75 ng/ml




#21 AR-1248-1

R.T.: 5.788 min
Delta R.T.: -0.012 min
Response: 18989577
Conc: 529.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/10/2025
Supervised By :Yogesh Patel 11/10/2025




#21 AR-1248-1

R.T.: 4.868 min
Delta R.T.: -0.008 min
Response: 179360634
Conc: 477.51 ng/ml m




#22 AR-1248-2

R.T.: 6.059 min
Delta R.T.: -0.012 min
Response: 24513555
Conc: 502.90 ng/ml m



#22 AR-1248-2

R.T.: 5.090 min
Delta R.T.: -0.006 min
Response: 115903723
Conc: 471.33 ng/ml




#23 AR-1248-3

R.T.: 6.262 min
Delta R.T.: -0.012 min
Response: 26950703
Conc: 496.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/10/2025
Supervised By :Yogesh Patel 11/10/2025




#23 AR-1248-3

R.T.: 5.130 min
Delta R.T.: -0.006 min
Response: 145461041
Conc: 507.02 ng/ml m




#24 AR-1248-4

R.T.: 6.660 min
Delta R.T.: -0.012 min
Response: 32416705
Conc: 488.47 ng/ml m



#24 AR-1248-4

R.T.: 5.302 min
Delta R.T.: -0.007 min
Response: 140543165
Conc: 497.13 ng/ml




#25 AR-1248-5

R.T.: 6.699 min
Delta R.T.: -0.012 min
Response: 30868867
Conc: 475.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/10/2025
Supervised By :Yogesh Patel 11/10/2025



#25 AR-1248-5

R.T.: 5.691 min
Delta R.T.: -0.007 min
Response: 294218466
Conc: 566.18 ng/ml