

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110821\
 Data File : PP040787.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Nov 2021 18:37
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 00:37:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 04 11:28:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.763	3.767	446480	557784	22.039	23.243
2)	SA Decachlor...	10.668	9.035	458166	376146	21.513	21.539
Target Compounds							
9)	L2 AR-1221-2	5.149f	0.000	17677	0	117.120	N.D. #
10)	L2 AR-1221-3	0.000	4.210	0	27138	N.D.	41.343 #
11)	L3 AR-1232-1	0.000	4.210	0	27138	N.D.	50.284 #
28)	L6 AR-1254-3	7.615f	0.000	6077	0	4.514	N.D. #
32)	L7 AR-1260-2	8.055f	0.000	6542	0	5.651	N.D. #
33)	L7 AR-1260-3	8.410	0.000	20933	0	23.707	N.D. #
37)	L8 AR-1262-2	8.908	0.000	35057	0	15.427	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110821\
Data File : PP040787.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Nov 2021 18:37
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 00:37:53 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110421.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 04 11:28:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.763 min
Delta R.T.: -0.007 min
Response: 446480
Conc: 22.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.767 min
Delta R.T.: 0.000 min
Response: 557784
Conc: 23.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.668 min
Delta R.T.: -0.008 min
Response: 458166
Conc: 21.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.035 min
Delta R.T.: -0.002 min
Response: 376146
Conc: 21.54 ng/ml



#9 AR-1221-2


R.T.: 5.149 min
Delta R.T.: 0.015 min
Response: 17677
Conc: 117.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 4.125 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 5.220 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 4.210 min
Delta R.T.: -0.003 min
Response: 27138
Conc: 41.34 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T.: 5.218 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.210 min
Delta R.T.: -0.001 min
Response: 27138
Conc: 50.28 ng/ml




#28 AR-1254-3

R.T.: 7.615 min
Delta R.T.: 0.019 min
Response: 6077
Conc: 4.51 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 6.466 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 8.055 min
Delta R.T.: 0.023 min
Response: 6542
Conc: 5.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 6.802 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 8.410 min
Delta R.T.: 0.014 min
Response: 20933
Conc: 23.71 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 6.956 min
Response: 0
Conc: N.D.



#37 AR-1262-2

R.T.: 8.908 min
Delta R.T.: -0.010 min
Response: 35057
Conc: 15.43 ng/ml

Instrument :
ECD_P
ClientSampleId :



#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.689 min
Response: 0
Conc: N.D.