

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110821\
 Data File : PP040766.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Nov 2021 12:33
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 00:33:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 04 11:28:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.763	3.767	1145122	1356305	56.525	56.517
2)	SA Decachlor...	10.668	9.035	1083036	876858	50.853	50.212
Target Compounds							
26)	L6 AR-1254-1	6.985	5.886	412543	502894	534.367	516.994
27)	L6 AR-1254-2	7.214	6.046	642818	431788	517.848	514.491
28)	L6 AR-1254-3	7.594	6.465	687766	657537	510.861	520.648
29)	L6 AR-1254-4	7.890	6.705	502620	392583	515.142	527.630
30)	L6 AR-1254-5	8.317	7.134	558354	564199	514.642	525.389



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110821\
Data File : PP040766.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Nov 2021 12:33
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 00:33:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110421.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 04 11:28:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.763 min
Delta R.T.: -0.007 min
Response: 1145122
Conc: 56.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.767 min
Delta R.T.: 0.000 min
Response: 1356305
Conc: 56.52 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.668 min
Delta R.T.: -0.008 min
Response: 1083036
Conc: 50.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.035 min
Delta R.T.: -0.002 min
Response: 876858
Conc: 50.21 ng/ml



#26 AR-1254-1


R.T.: 6.985 min
Delta R.T.: 0.000 min
Response: 412543
Conc: 534.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.886 min
Delta R.T.: 0.000 min
Response: 502894
Conc: 516.99 ng/ml




#27 AR-1254-2

R.T.: 7.214 min
Delta R.T.: 0.000 min
Response: 642818
Conc: 517.85 ng/ml



#27 AR-1254-2


R.T.: 6.046 min
Delta R.T.: 0.000 min
Response: 431788
Conc: 514.49 ng/ml



#28 AR-1254-3


R.T.: 7.594 min
Delta R.T.: -0.001 min
Response: 687766
Conc: 510.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.465 min
Delta R.T.: 0.000 min
Response: 657537
Conc: 520.65 ng/ml




#29 AR-1254-4

R.T.: 7.890 min
Delta R.T.: -0.002 min
Response: 502620
Conc: 515.14 ng/ml



#29 AR-1254-4


R.T.: 6.705 min
Delta R.T.: 0.000 min
Response: 392583
Conc: 527.63 ng/ml



#30 AR-1254-5

R.T.: 8.317 min
Delta R.T.: 0.000 min
Response: 558354
Conc: 514.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.134 min
Delta R.T.: 0.000 min
Response: 564199
Conc: 525.39 ng/ml