

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110822\
 Data File : PP053167.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Nov 2022 22:58
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 05:35:47 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:34:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.400	3.646	154.8E6	126.1E6	73.893	74.149
2) SA Decachlor...	10.199	8.724	175.9E6	251.8E6	72.772	72.984
Target Compounds						
41) L9 AR-1268-1	8.668	7.600	143.4E6	202.0E6	734.386	737.233
42) L9 AR-1268-2	8.763	7.666	128.3E6	181.0E6	730.805	736.259
43) L9 AR-1268-3	8.995	7.873	113.0E6	156.0E6	732.284	734.229
44) L9 AR-1268-4	9.427	8.165	42570456	63392871	724.424	733.829
45) L9 AR-1268-5	9.852	8.463	355.4E6	512.9E6	741.488	736.656



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110822\
Data File : PP053167.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Nov 2022 22:58
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 05:35:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:34:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.400 min
Delta R.T.: 0.001 min
Response: 154807705
Conc: 73.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: 0.000 min
Response: 126072945
Conc: 74.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.199 min
Delta R.T.: 0.000 min
Response: 175903720
Conc: 72.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.724 min
Delta R.T.: 0.000 min
Response: 251768976
Conc: 72.98 ng/ml



#41 AR-1268-1


R.T.: 8.668 min
Delta R.T.: 0.000 min
Response: 143391060
Conc: 734.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.600 min
Delta R.T.: 0.000 min
Response: 201956269
Conc: 737.23 ng/ml




#42 AR-1268-2

R.T.: 8.763 min
Delta R.T.: 0.000 min
Response: 128333380
Conc: 730.80 ng/ml



#42 AR-1268-2


R.T.: 7.666 min
Delta R.T.: 0.001 min
Response: 181015175
Conc: 736.26 ng/ml



#43 AR-1268-3


R.T.: 8.995 min
Delta R.T.: 0.001 min
Response: 113013813
Conc: 732.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.873 min
Delta R.T.: 0.000 min
Response: 156017311
Conc: 734.23 ng/ml




#44 AR-1268-4

R.T.: 9.427 min
Delta R.T.: 0.000 min
Response: 42570456
Conc: 724.42 ng/ml



#44 AR-1268-4


R.T.: 8.165 min
Delta R.T.: 0.001 min
Response: 63392871
Conc: 733.83 ng/ml



#45 AR-1268-5

R.T.: 9.852 min
Delta R.T.: 0.002 min
Response: 355362087
Conc: 741.49 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.463 min
Delta R.T.: 0.000 min
Response: 512896236
Conc: 736.66 ng/ml