

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110822\
 Data File : PP053147.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Nov 2022 17:30
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 04:01:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 03:52:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.402	3.648	8407710	6420963	4.166	4.027
2) SA Decachlor...	10.202	8.726	6316750	8885707	4.487	4.567
Target Compounds						
3) L1 AR-1016-1	5.575	4.747	2972490	2665925	45.701	45.864
4) L1 AR-1016-2	5.598	4.766	4390057	3514291	45.812	42.748
5) L1 AR-1016-3	5.660	4.946	2630447	1899024	44.476	42.951
6) L1 AR-1016-4	5.759	4.988	2073300	1639266	43.558	44.599
7) L1 AR-1016-5	6.056	5.204	1981538	2193060	41.266	46.497
31) L7 AR-1260-1	7.187	6.251	4006348	4395072	47.696	46.119
32) L7 AR-1260-2	7.444	6.441	4641658	5823611	49.222	49.195
33) L7 AR-1260-3	7.805	6.597	3072484	5052184	48.089	45.176
34) L7 AR-1260-4	8.032	7.072	3923639	3806939	51.862	43.978
35) L7 AR-1260-5	8.355	7.316	5831655	8454148	44.235	42.515



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110822\
Data File : PP053147.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Nov 2022 17:30
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 04:01:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 03:52:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.402 min
Delta R.T.: 0.000 min
Response: 8407710
Conc: 4.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: 0.000 min
Response: 6420963
Conc: 4.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.202 min
Delta R.T.: 0.000 min
Response: 6316750
Conc: 4.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.726 min
Delta R.T.: -0.001 min
Response: 8885707
Conc: 4.57 ng/ml



#3 AR-1016-1


R.T.: 5.575 min
Delta R.T.: 0.000 min
Response: 2972490
Conc: 45.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#3 AR-1016-1

R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 2665925
Conc: 45.86 ng/ml




#4 AR-1016-2

R.T.: 5.598 min
Delta R.T.: 0.000 min
Response: 4390057
Conc: 45.81 ng/ml



#4 AR-1016-2


R.T.: 4.766 min
Delta R.T.: 0.000 min
Response: 3514291
Conc: 42.75 ng/ml



#5 AR-1016-3


R.T.: 5.660 min
Delta R.T.: 0.000 min
Response: 2630447
Conc: 44.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#5 AR-1016-3

R.T.: 4.946 min
Delta R.T.: 0.000 min
Response: 1899024
Conc: 42.95 ng/ml




#6 AR-1016-4

R.T.: 5.759 min
Delta R.T.: 0.000 min
Response: 2073300
Conc: 43.56 ng/ml



#6 AR-1016-4


R.T.: 4.988 min
Delta R.T.: 0.000 min
Response: 1639266
Conc: 44.60 ng/ml



#7 AR-1016-5


R.T.: 6.056 min
Delta R.T.: 0.000 min
Response: 1981538
Conc: 41.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#7 AR-1016-5

R.T.: 5.204 min
Delta R.T.: 0.000 min
Response: 2193060
Conc: 46.50 ng/ml




#31 AR-1260-1

R.T.: 7.187 min
Delta R.T.: 0.000 min
Response: 4006348
Conc: 47.70 ng/ml



#31 AR-1260-1


R.T.: 6.251 min
Delta R.T.: 0.000 min
Response: 4395072
Conc: 46.12 ng/ml



#32 AR-1260-2


R.T.: 7.444 min
Delta R.T.: 0.000 min
Response: 4641658
Conc: 49.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#32 AR-1260-2

R.T.: 6.441 min
Delta R.T.: 0.000 min
Response: 5823611
Conc: 49.19 ng/ml




#33 AR-1260-3

R.T.: 7.805 min
Delta R.T.: 0.000 min
Response: 3072484
Conc: 48.09 ng/ml



#33 AR-1260-3


R.T.: 6.597 min
Delta R.T.: 0.000 min
Response: 5052184
Conc: 45.18 ng/ml



#34 AR-1260-4


R.T.: 8.032 min
Delta R.T.: 0.000 min
Response: 3923639
Conc: 51.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#34 AR-1260-4

R.T.: 7.072 min
Delta R.T.: 0.000 min
Response: 3806939
Conc: 43.98 ng/ml



#35 AR-1260-5

R.T.: 8.355 min
Delta R.T.: 0.000 min
Response: 5831655
Conc: 44.23 ng/ml



#35 AR-1260-5

R.T.: 7.316 min
Delta R.T.: 0.000 min
Response: 8454148
Conc: 42.51 ng/ml