

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP110918.M
 Title : 8080.M
 Last Update : Fri Nov 09 17:06:51 2018
 Response Via : Initial Calibration

Calibration Files

750 =PP028016.D 1000 =PP028017.D 1500 =PP028018.D
 500 =PP028015.D 200 =PP028014.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	6.363	6.242	6.080	6.246	6.454	6.277	E5	2.25
2) T	3,5-DICHLOROBENZOI...	5.428	5.230	5.088	5.374	5.688	5.361	E5	4.20
3) T	4-Nitrophenol	5.470	5.339	5.360	5.213	4.936	5.264	E5	3.89
4) S	2,4-DCAA	2.644	2.581	2.528	2.600	2.663	2.603	E5	2.06
5) T	DICAMBA	1.673	1.623	1.576	1.654	1.747	1.655	E6	3.83
6) T	MCPP	4.817	4.847	4.679	4.487	3.207	4.407	E5	15.56
7) T	MCPA	7.604	7.575	7.283	6.962	5.303	6.946	E5	13.74
8) T	DICHLORPROP	4.775	4.590	4.395	4.085	3.551	4.279	E5	11.23
9) T	2,4-D	3.890	3.785	3.676	3.856	3.991	3.840	E5	3.07
10) T	Pentachlorophenol	8.482	8.024	7.451	8.650	9.269	8.375	E6	8.15
11) T	2,4,5-TP (SILVEX)	3.046	2.924	2.788	3.054	3.242	3.011	E6	5.60
12) T	2,4,5-T	2.800	2.697	2.577	2.809	2.984	2.773	E6	5.43
13) T	2,4-DB	2.323	2.286	2.258	2.200	2.244	2.262	E5	2.03
14) T	DINOSEB	2.284	2.192	2.108	2.292	2.472	2.270	E6	5.99
15) T	Picloram	4.881	4.761	4.616	4.727	4.653	4.728	E6	2.18
16) T	DCPA	4.608	4.413	4.177	4.629	4.780	4.521	E6	5.14

Signal #2 Calibration Files

750 =PP028016.D 1000 =PP028017.D 1500 =PP028018.D
 500 =PP028015.D 200 =PP028014.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.868	5.720	5.572	5.826	6.248	5.847	E5	4.30
2) T	3,5-DICHLOROBENZOI...	5.257	5.116	5.007	5.130	5.359	5.174	E5	2.64
3) T	4-Nitrophenol	2.375	2.337	2.318	2.250	2.093	2.274	E5	4.90
4) S	2,4-DCAA	4.295	4.183	4.099	4.199	4.446	4.244	E5	3.12
5) T	DICAMBA	1.708	1.651	1.596	1.675	1.767	1.679	E6	3.80
6) T	MCPP	8.888	8.910	8.690	8.253	6.767	8.302	E5	10.81
7) T	MCPA	1.389	1.395	1.389	1.267	1.024	1.293	E6	12.35
8) T	DICHLORPROP	4.174	4.032	3.931	4.094	4.365	4.119	E5	3.98
9) T	2,4-D	4.935	4.796	4.693	4.821	4.933	4.835	E5	2.10
10) T	Pentachlorophenol	6.521	6.097	5.593	6.666	7.356	6.447	E6	10.20
11) T	2,4,5-TP (SILVEX)	2.502	2.394	2.274	2.483	2.621	2.455	E6	5.27
12) T	2,4,5-T	2.225	2.137	2.045	2.187	2.273	2.173	E6	4.03
13) T	2,4-DB	2.727	2.657	2.666	2.540	2.375	2.593	E5	5.38
14) T	DINOSEB	1.826	1.748	1.687	1.800	1.928	1.798	E6	5.01
15) T	Picloram	3.703	3.564	3.395	3.630	3.702	3.599	E6	3.54
16) T	DCPA	3.253	3.075	2.877	3.278	3.588	3.214	E6	8.22

(#) = Out of Range