

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110922\
 Data File : PP053192.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Nov 2022 14:02
 Operator : YP\AJ
 Sample : N5495-04
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB-31A

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/10/2022
 Supervised By :Ankita Jodhani 11/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 14:33:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.399	3.645	54247736	45379048	28.220	30.167
2) SA Decachlor...	10.197	8.723	37746903	45080640	27.838	24.051m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110922\
Data File : PP053192.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Nov 2022 14:02
Operator : YP\AJ
Sample : N5495-04
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SB-31A


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/10/2022
Supervised By :Ankita Jodhani 11/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 14:33:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:50:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.399 min
Delta R.T.: -0.004 min
Response: 54247736
Conc: 28.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-31A


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/10/2022
Supervised By :Ankita Jodhani 11/10/2022




#1 Tetrachloro-m-xylene

R.T.: 3.645 min
Delta R.T.: -0.004 min
Response: 45379048
Conc: 30.17 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.197 min
Delta R.T.: -0.006 min
Response: 37746903
Conc: 27.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.723 min
Delta R.T.: -0.006 min
Response: 45080640
Conc: 24.05 ng/ml m