

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110922\
 Data File : PP053203.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Nov 2022 17:13
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 21:03:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.400	3.646	114.2E6	93741812	59.385	62.318
2) SA Decachlor...	10.197	8.724	77359980	102.6E6	57.052	54.728
Target Compounds						
26) L6 AR-1254-1	6.433	5.559	44048963	51288230	623.071	565.780
27) L6 AR-1254-2	6.652	5.708	59905386	46503892	570.258	572.745
28) L6 AR-1254-3	7.020	6.117	57152152	71334139	557.183	561.180
29) L6 AR-1254-4	7.307	6.348	40076420	44939431	576.072	569.909
30) L6 AR-1254-5	7.727	6.771	42345557	68364803	534.620	541.248



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110922\
Data File : PP053203.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Nov 2022 17:13
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 21:03:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:50:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.400 min
Delta R.T.: -0.003 min
Response: 114157204
Conc: 59.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: -0.003 min
Response: 93741812
Conc: 62.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.197 min
Delta R.T.: -0.005 min
Response: 77359980
Conc: 57.05 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.724 min
Delta R.T.: -0.005 min
Response: 102579299
Conc: 54.73 ng/ml



#26 AR-1254-1


R.T.: 6.433 min
Delta R.T.: 0.000 min
Response: 44048963
Conc: 623.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.559 min
Delta R.T.: 0.000 min
Response: 51288230
Conc: 565.78 ng/ml




#27 AR-1254-2

R.T.: 6.652 min
Delta R.T.: 0.000 min
Response: 59905386
Conc: 570.26 ng/ml



#27 AR-1254-2


R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 46503892
Conc: 572.74 ng/ml



#28 AR-1254-3


R.T.: 7.020 min
Delta R.T.: 0.000 min
Response: 57152152
Conc: 557.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.117 min
Delta R.T.: 0.000 min
Response: 71334139
Conc: 561.18 ng/ml




#29 AR-1254-4

R.T.: 7.307 min
Delta R.T.: 0.000 min
Response: 40076420
Conc: 576.07 ng/ml



#29 AR-1254-4


R.T.: 6.348 min
Delta R.T.: 0.000 min
Response: 44939431
Conc: 569.91 ng/ml



#30 AR-1254-5

R.T.: 7.727 min
Delta R.T.: 0.000 min
Response: 42345557
Conc: 534.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.771 min
Delta R.T.: 0.000 min
Response: 68364803
Conc: 541.25 ng/ml