

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111020\
 Data File : PP031210.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Nov 2020 19:46
 Operator : DD\AJ
 Sample : PB132915BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB132915BS

Manual Integrations
 APPROVED

Ankita
 11/11/2020 9:01:04 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 00:41:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 05 13:03:18 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.798	4.072	1068179	959291	15.993	14.432
2)	SA Decachlor...	10.681	9.391	728327	794535	18.216	19.028m
Target Compounds							
3)	L1 AR-1016-1	6.104	5.332	667207	622751	342.670	335.666
4)	L1 AR-1016-2	6.127	5.352	985399	891198	345.179	327.864
5)	L1 AR-1016-3	6.192	5.544	608784	495693	349.523	391.359
6)	L1 AR-1016-4	6.297	5.596	513494	351423	355.367	381.051
7)	L1 AR-1016-5	6.613	5.825	468013	456001	375.436	347.421
31)	L7 AR-1260-1	7.785	6.922	759343	809191	413.397	403.753
32)	L7 AR-1260-2	8.050	7.119	879475	962634	408.238	411.063
33)	L7 AR-1260-3	8.415	7.274	578678	943024	351.551	414.790
34)	L7 AR-1260-4	8.643	7.758	726546	699761	390.752	383.545
35)	L7 AR-1260-5	8.959	8.004	1377190	1714904	365.259	381.058

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111020\
Data File : PP031210.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Nov 2020 19:46
Operator : DD\AJ
Sample : PB132915BS
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampled :
PB132915BS


Manual Integrations
APPROVED

Ankita
11/11/2020 9:01:04 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 00:41:54 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 05 13:03:18 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.798 min
Delta R.T.: -0.003 min
Response: 1068179
Conc: 15.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB132915BS


Manual Integrations
APPROVED

Ankita
11/11/2020 9:01:04 AM




#1 Tetrachloro-m-xylene

R.T.: 4.072 min
Delta R.T.: -0.004 min
Response: 959291
Conc: 14.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.681 min
Delta R.T.: -0.007 min
Response: 728327
Conc: 18.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.391 min
Delta R.T.: -0.013 min
Response: 794535
Conc: 19.03 ng/ml m




#3 AR-1016-1

R.T.: 6.104 min
Delta R.T.: -0.004 min
Response: 667207
Conc: 342.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB132915BS


Manual Integrations
APPROVED

Ankita
11/11/2020 9:01:04 AM




#3 AR-1016-1

R.T.: 5.332 min
Delta R.T.: -0.007 min
Response: 622751
Conc: 335.67 ng/ml




#4 AR-1016-2

R.T.: 6.127 min
Delta R.T.: -0.003 min
Response: 985399
Conc: 345.18 ng/ml



#4 AR-1016-2

R.T.: 5.352 min
Delta R.T.: -0.007 min
Response: 891198
Conc: 327.86 ng/ml




#5 AR-1016-3

R.T.: 6.192 min
Delta R.T.: -0.004 min
Response: 608784
Conc: 349.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB132915BS


Manual Integrations
APPROVED

Ankita
11/11/2020 9:01:04 AM




#5 AR-1016-3

R.T.: 5.544 min
Delta R.T.: -0.007 min
Response: 495693
Conc: 391.36 ng/ml




#6 AR-1016-4

R.T.: 6.297 min
Delta R.T.: -0.004 min
Response: 513494
Conc: 355.37 ng/ml



#6 AR-1016-4

R.T.: 5.596 min
Delta R.T.: -0.008 min
Response: 351423
Conc: 381.05 ng/ml




#7 AR-1016-5

R.T.: 6.613 min
Delta R.T.: -0.004 min
Response: 468013
Conc: 375.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB132915BS


Manual Integrations
APPROVED

Ankita
11/11/2020 9:01:04 AM




#7 AR-1016-5

R.T.: 5.825 min
Delta R.T.: -0.008 min
Response: 456001
Conc: 347.42 ng/ml




#31 AR-1260-1

R.T.: 7.785 min
Delta R.T.: -0.004 min
Response: 759343
Conc: 413.40 ng/ml



#31 AR-1260-1

R.T.: 6.922 min
Delta R.T.: -0.009 min
Response: 809191
Conc: 403.75 ng/ml




#32 AR-1260-2

R.T.: 8.050 min
Delta R.T.: -0.005 min
Response: 879475
Conc: 408.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB132915BS


Manual Integrations
APPROVED

Ankita
11/11/2020 9:01:04 AM




#32 AR-1260-2

R.T.: 7.119 min
Delta R.T.: -0.008 min
Response: 962634
Conc: 411.06 ng/ml




#33 AR-1260-3

R.T.: 8.415 min
Delta R.T.: -0.004 min
Response: 578678
Conc: 351.55 ng/ml



#33 AR-1260-3

R.T.: 7.274 min
Delta R.T.: -0.009 min
Response: 943024
Conc: 414.79 ng/ml




#34 AR-1260-4

R.T.: 8.643 min
Delta R.T.: -0.005 min
Response: 726546
Conc: 390.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB132915BS


Manual Integrations
APPROVED

Ankita
11/11/2020 9:01:04 AM




#34 AR-1260-4

R.T.: 7.758 min
Delta R.T.: -0.009 min
Response: 699761
Conc: 383.54 ng/ml



#35 AR-1260-5

R.T.: 8.959 min
Delta R.T.: -0.004 min
Response: 1377190
Conc: 365.26 ng/ml



#35 AR-1260-5

R.T.: 8.004 min
Delta R.T.: -0.009 min
Response: 1714904
Conc: 381.06 ng/ml