

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111022\
 Data File : PP053220.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Nov 2022 09:37
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 10 20:47:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.399	3.645	99781216	84820752	51.906	56.387
2) SA Decachlor...	10.198	8.723	71762098	94149427	52.924	50.230
Target Compounds						
21) L5 AR-1248-1	5.572	4.745	18844132	18121328	505.241	528.533
22) L5 AR-1248-2	5.848	4.984	26578151	25012199	485.777	520.846
23) L5 AR-1248-3	6.053	5.027	28740752	25807304	482.319	520.807
24) L5 AR-1248-4	6.458	5.201	30162789	30755316	488.894	520.757
25) L5 AR-1248-5	6.497	5.599	29489792	28987163	488.874	526.904



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111022\
Data File : PP053220.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Nov 2022 09:37
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 20:47:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:50:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.399 min
Delta R.T.: -0.004 min
Response: 99781216
Conc: 51.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.645 min
Delta R.T.: -0.004 min
Response: 84820752
Conc: 56.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.198 min
Delta R.T.: -0.005 min
Response: 71762098
Conc: 52.92 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.723 min
Delta R.T.: -0.005 min
Response: 94149427
Conc: 50.23 ng/ml



#21 AR-1248-1


R.T.: 5.572 min
Delta R.T.: -0.002 min
Response: 18844132
Conc: 505.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.745 min
Delta R.T.: -0.001 min
Response: 18121328
Conc: 528.53 ng/ml




#22 AR-1248-2

R.T.: 5.848 min
Delta R.T.: -0.001 min
Response: 26578151
Conc: 485.78 ng/ml



#22 AR-1248-2


R.T.: 4.984 min
Delta R.T.: -0.002 min
Response: 25012199
Conc: 520.85 ng/ml



#23 AR-1248-3


R.T.: 6.053 min
Delta R.T.: -0.002 min
Response: 28740752
Conc: 482.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.027 min
Delta R.T.: -0.002 min
Response: 25807304
Conc: 520.81 ng/ml




#24 AR-1248-4

R.T.: 6.458 min
Delta R.T.: -0.001 min
Response: 30162789
Conc: 488.89 ng/ml



#24 AR-1248-4


R.T.: 5.201 min
Delta R.T.: -0.002 min
Response: 30755316
Conc: 520.76 ng/ml



#25 AR-1248-5

R.T.: 6.497 min
Delta R.T.: -0.001 min
Response: 29489792
Conc: 488.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.599 min
Delta R.T.: -0.002 min
Response: 28987163
Conc: 526.90 ng/ml