

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111120\
 Data File : PP031222.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Nov 2020 8:58
 Operator : DD\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 16:02:11 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 05 13:03:18 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	0.000	4.076	0	75100	N.D.	1.130 #
Target Compounds							
3)	L1 AR-1016-1	0.000	5.365f	0	13346	N.D.	7.193 #
4)	L1 AR-1016-2	0.000	5.365	0	13346	N.D.	4.910 #
5)	L1 AR-1016-3	0.000	5.522f	0	33254	N.D.	26.255 #
8)	L2 AR-1221-1	0.000	4.333	0	151843	N.D.	265.011 #
9)	L2 AR-1221-2	0.000	4.435	0	70168	N.D.	163.880 #
10)	L2 AR-1221-3	0.000	4.519	0	48663	N.D.	36.935 #
11)	L3 AR-1232-1	0.000	4.524	0	43711	N.D.	36.491 #
12)	L3 AR-1232-2	0.000	5.365	0	13346	N.D.	12.558 #
13)	L3 AR-1232-3	0.000	5.522f	0	33254	N.D.	64.861 #
16)	L4 AR-1242-1	0.000	5.365f	0	13346	N.D.	10.370 #
17)	L4 AR-1242-2	0.000	5.365	0	13346	N.D.	7.147 #
18)	L4 AR-1242-3	0.000	5.522f	0	33254	N.D.	35.420 #
20)	L4 AR-1242-5	0.000	6.232f	0	73127	N.D.	73.457 #
21)	L5 AR-1248-1	0.000	5.365f	0	13346	N.D.	14.423 #
24)	L5 AR-1248-4	7.033	0.000	30060	0	19.723	N.D. #
25)	L5 AR-1248-5	0.000	6.232f	0	73127	N.D.	57.902 #
26)	L6 AR-1254-1	7.033f	6.232f	30060	73127	21.410	37.183 #
36)	L8 AR-1262-1	0.000	7.507	0	3260	N.D.	1.403 #
38)	L8 AR-1262-3	0.000	8.290	0	23613	N.D.	13.927 #
41)	L9 AR-1268-1	0.000	8.290	0	23613	N.D.	4.823 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111120\
Data File : PP031222.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Nov 2020 8:58
Operator : DD\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 16:02:11 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110420.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 05 13:03:18 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 4.801 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.076 min
Delta R.T.: 0.000 min
Response: 75100
Conc: 1.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.688 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 9.404 min
Response: 0
Conc: N.D.



#3 AR-1016-1


R.T.: 0.000 min
Exp R.T. : 6.107 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.365 min
Delta R.T.: 0.026 min
Response: 13346
Conc: 7.19 ng/ml




#4 AR-1016-2

R.T.: 0.000 min
Exp R.T. : 6.130 min
Response: 0
Conc: N.D.



#4 AR-1016-2


R.T.: 5.365 min
Delta R.T.: 0.005 min
Response: 13346
Conc: 4.91 ng/ml



#5 AR-1016-3


R.T.: 0.000 min
Exp R.T.: 6.196 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.522 min
Delta R.T.: -0.029 min
Response: 33254
Conc: 26.25 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T.: 5.044 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 4.333 min
Delta R.T.: 0.000 min
Response: 151843
Conc: 265.01 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 5.143 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.435 min
Delta R.T.: 0.003 min
Response: 70168
Conc: 163.88 ng/ml




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T.: 5.227 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 4.519 min
Delta R.T.: -0.002 min
Response: 48663
Conc: 36.93 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T. : 5.230 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.524 min
Delta R.T.: 0.000 min
Response: 43711
Conc: 36.49 ng/ml




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 5.814 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 5.365 min
Delta R.T.: 0.005 min
Response: 13346
Conc: 12.56 ng/ml



#13 AR-1232-3


R.T.: 0.000 min
Exp R.T.: 6.130 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.522 min
Delta R.T.: -0.029 min
Response: 33254
Conc: 64.86 ng/ml




#16 AR-1242-1

R.T.: 0.000 min
Exp R.T.: 6.107 min
Response: 0
Conc: N.D.



#16 AR-1242-1


R.T.: 5.365 min
Delta R.T.: 0.025 min
Response: 13346
Conc: 10.37 ng/ml



#17 AR-1242-2


R.T.: 0.000 min
Exp R.T.: 6.130 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.365 min
Delta R.T.: 0.005 min
Response: 13346
Conc: 7.15 ng/ml




#18 AR-1242-3

R.T.: 0.000 min
Exp R.T.: 6.196 min
Response: 0
Conc: N.D.



#18 AR-1242-3


R.T.: 5.522 min
Delta R.T.: -0.030 min
Response: 33254
Conc: 35.42 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 7.081 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#20 AR-1242-5

R.T.: 6.232 min
Delta R.T.: 0.019 min
Response: 73127
Conc: 73.46 ng/ml




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T.: 6.108 min
Response: 0
Conc: N.D.



#21 AR-1248-1


R.T.: 5.365 min
Delta R.T.: 0.026 min
Response: 13346
Conc: 14.42 ng/ml



#24 AR-1248-4


R.T.: 7.033 min
Delta R.T.: -0.009 min
Response: 30060
Conc: 19.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.833 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 7.080 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.232 min
Delta R.T.: -0.024 min
Response: 73127
Conc: 57.90 ng/ml



#26 AR-1254-1


R.T.: 7.033 min
Delta R.T.: 0.020 min
Response: 30060
Conc: 21.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 6.232 min
Delta R.T.: 0.020 min
Response: 73127
Conc: 37.18 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 8.419 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 7.507 min
Delta R.T.: 0.004 min
Response: 3260
Conc: 1.40 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 9.262 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#38 AR-1262-3

R.T.: 8.290 min
Delta R.T.: -0.011 min
Response: 23613
Conc: 13.93 ng/ml



#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 9.261 min
Response: 0
Conc: N.D.



#41 AR-1268-1

R.T.: 8.290 min
Delta R.T.: -0.011 min
Response: 23613
Conc: 4.82 ng/ml