

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111122\
 Data File : PP053291.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2022 14:52
 Operator : YP\AJ
 Sample : N5512-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 20:19:47 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.402	3.648	45266480	38747177	23.548	25.758
2)	SA Decachlor...	10.206	8.729	37293427	48784275	27.503	26.027
Target Compounds							
3)	L1 AR-1016-1	5.566	0.000	925942	0	14.884	N.D. #
4)	L1 AR-1016-2	5.592	0.000	1220641	0	13.207	N.D. #
5)	L1 AR-1016-3	5.652	0.000	763039	0	13.482	N.D. #
6)	L1 AR-1016-4	5.754	4.989	156984	447625	3.451	12.807 #
7)	L1 AR-1016-5	6.059	0.000	123063	0	2.719	N.D. #
8)	L2 AR-1221-1	4.610	0.000	837421	0	37.259	N.D. #
9)	L2 AR-1221-2	4.710	3.952	692130	1934843	41.654	132.089 #
10)	L2 AR-1221-3	4.763	0.000	151076	0	3.000	N.D. #
11)	L3 AR-1232-1	4.784	0.000	264298	0	6.892	N.D. #
12)	L3 AR-1232-2	5.335f	0.000	1950700	0	86.668	N.D. #
13)	L3 AR-1232-3	5.592	0.000	1220641	0	30.074	N.D. #
14)	L3 AR-1232-4	5.754	5.018	156984	122482	7.778	6.958
15)	L3 AR-1232-5	5.868f	0.000	298201	0	17.985	N.D. #
16)	L4 AR-1242-1	5.566	0.000	925942	0	19.000	N.D. #
17)	L4 AR-1242-2	5.592	0.000	1220641	0	16.931	N.D. #
18)	L4 AR-1242-3	5.652	0.000	763039	0	17.039	N.D. #
19)	L4 AR-1242-4	5.754	5.018	156984	122482	4.365	3.503
20)	L4 AR-1242-5	6.495	5.562	765064	426876	20.955	9.748 #
21)	L5 AR-1248-1	5.566	0.000	925942	0	24.826	N.D. #
22)	L5 AR-1248-2	5.868f	4.989	298201	447625	5.450	9.321 #
23)	L5 AR-1248-3	6.059	5.018	123063	122482	2.065	2.472
24)	L5 AR-1248-4	6.437f	0.000	1802120	0	29.210	N.D. #
25)	L5 AR-1248-5	6.495	0.000	765064	0	12.683	N.D. #
26)	L6 AR-1254-1	6.437	5.562	1802120	426876	25.491	4.709 #
27)	L6 AR-1254-2	6.655	5.710	751176	361996	7.151	4.458 #
28)	L6 AR-1254-3	7.026	6.121	1124988	985068	10.968	7.749 #
29)	L6 AR-1254-4	7.309	6.350	481433	369947	6.920	4.692 #
30)	L6 AR-1254-5	7.734	6.774	529310	1088518	6.683	8.618 #
31)	L7 AR-1260-1	7.193	6.247	207989	714980	2.544	7.800 #
32)	L7 AR-1260-2	7.447	6.439	403540	1869024	4.358	16.242 #
33)	L7 AR-1260-3	7.806	6.596	157912	717974	2.532	6.691 #
34)	L7 AR-1260-4	8.036	7.076	310356	647426	4.143	7.825 #
35)	L7 AR-1260-5	8.361	7.319	245888	213532	1.943	1.128 #
36)	L8 AR-1262-1	7.806	0.000	157912	0	1.643	N.D. #
37)	L8 AR-1262-2	8.346	7.076	90254	647426	0.596	5.583 #
38)	L8 AR-1262-3	8.661	7.608	18095	484241	0.175	5.287 #
39)	L8 AR-1262-4	8.771	7.649f	362795	459546	4.487	2.802 #
40)	L8 AR-1262-5	9.424	8.173	212028	621450	4.170	8.180 #
41)	L9 AR-1268-1	8.661	7.608	18095	484241	0.096	1.812 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111122\
 Data File : PP053291.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2022 14:52
 Operator : YP\AJ
 Sample : N5512-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 20:19:47 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.771	7.649f	362795	459546	2.139	1.925
43) L9	AR-1268-3	8.992	7.872	-22573	296491	N.D.	1.442 #
44) L9	AR-1268-4	9.424	8.173	212028	621450	3.788	7.550 #
45) L9	AR-1268-5	9.847	8.472	442673	181945	0.961	0.274 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111122\
Data File : PP053291.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2022 14:52
Operator : YP\AJ
Sample : N5512-02
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 20:19:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:50:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.402 min
Delta R.T.: 0.000 min
Response: 45266480
Conc: 23.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: -0.001 min
Response: 38747177
Conc: 25.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.206 min
Delta R.T.: 0.003 min
Response: 37293427
Conc: 27.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.729 min
Delta R.T.: 0.000 min
Response: 48784275
Conc: 26.03 ng/ml



#3 AR-1016-1


R.T.: 5.566 min
Delta R.T.: -0.011 min
Response: 925942
Conc: 14.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.749 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.592 min
Delta R.T.: -0.006 min
Response: 1220641
Conc: 13.21 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.767 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.652 min
Delta R.T.: -0.010 min
Response: 763039
Conc: 13.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.946 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.754 min
Delta R.T.: -0.006 min
Response: 156984
Conc: 3.45 ng/ml



#6 AR-1016-4


R.T.: 4.989 min
Delta R.T.: 0.000 min
Response: 447625
Conc: 12.81 ng/ml



#7 AR-1016-5


R.T.: 6.059 min
Delta R.T.: 0.002 min
Response: 123063
Conc: 2.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.206 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.610 min
Delta R.T.: 0.000 min
Response: 837421
Conc: 37.26 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.863 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.710 min
Delta R.T.: 0.014 min
Response: 692130
Conc: 41.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.952 min
Delta R.T.: 0.002 min
Response: 1934843
Conc: 132.09 ng/ml




#10 AR-1221-3

R.T.: 4.763 min
Delta R.T.: -0.010 min
Response: 151076
Conc: 3.00 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 4.028 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.784 min
Delta R.T.: 0.010 min
Response: 264298
Conc: 6.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.027 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.335 min
Delta R.T.: 0.029 min
Response: 1950700
Conc: 86.67 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T. : 4.766 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.592 min
Delta R.T.: -0.005 min
Response: 1220641
Conc: 30.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.945 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.754 min
Delta R.T.: -0.005 min
Response: 156984
Conc: 7.78 ng/ml



#14 AR-1232-4


R.T.: 5.018 min
Delta R.T.: -0.011 min
Response: 122482
Conc: 6.96 ng/ml



#15 AR-1232-5


R.T.: 5.868 min
Delta R.T.: 0.017 min
Response: 298201
Conc: 17.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.204 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.566 min
Delta R.T.: -0.009 min
Response: 925942
Conc: 19.00 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.747 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.592 min
Delta R.T.: -0.005 min
Response: 1220641
Conc: 16.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.766 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.652 min
Delta R.T.: -0.007 min
Response: 763039
Conc: 17.04 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 4.945 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.754 min
Delta R.T.: -0.005 min
Response: 156984
Conc: 4.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.018 min
Delta R.T.: -0.011 min
Response: 122482
Conc: 3.50 ng/ml




#20 AR-1242-5

R.T.: 6.495 min
Delta R.T.: -0.004 min
Response: 765064
Conc: 20.96 ng/ml



#20 AR-1242-5


R.T.: 5.562 min
Delta R.T.: 0.002 min
Response: 426876
Conc: 9.75 ng/ml



#21 AR-1248-1


R.T.: 5.566 min
Delta R.T.: -0.008 min
Response: 925942
Conc: 24.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.746 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.868 min
Delta R.T.: 0.019 min
Response: 298201
Conc: 5.45 ng/ml



#22 AR-1248-2


R.T.: 4.989 min
Delta R.T.: 0.002 min
Response: 447625
Conc: 9.32 ng/ml



#23 AR-1248-3


R.T.: 6.059 min
Delta R.T.: 0.004 min
Response: 123063
Conc: 2.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.018 min
Delta R.T.: -0.010 min
Response: 122482
Conc: 2.47 ng/ml




#24 AR-1248-4

R.T.: 6.437 min
Delta R.T.: -0.022 min
Response: 1802120
Conc: 29.21 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 5.203 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.495 min
Delta R.T.: -0.003 min
Response: 765064
Conc: 12.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.601 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.437 min
Delta R.T.: 0.005 min
Response: 1802120
Conc: 25.49 ng/ml



#26 AR-1254-1


R.T.: 5.562 min
Delta R.T.: 0.002 min
Response: 426876
Conc: 4.71 ng/ml



#27 AR-1254-2


R.T.: 6.655 min
Delta R.T.: 0.002 min
Response: 751176
Conc: 7.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.710 min
Delta R.T.: 0.002 min
Response: 361996
Conc: 4.46 ng/ml




#28 AR-1254-3

R.T.: 7.026 min
Delta R.T.: 0.005 min
Response: 1124988
Conc: 10.97 ng/ml



#28 AR-1254-3


R.T.: 6.121 min
Delta R.T.: 0.004 min
Response: 985068
Conc: 7.75 ng/ml



#29 AR-1254-4


R.T.: 7.309 min
Delta R.T.: 0.002 min
Response: 481433
Conc: 6.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.350 min
Delta R.T.: 0.002 min
Response: 369947
Conc: 4.69 ng/ml




#30 AR-1254-5

R.T.: 7.734 min
Delta R.T.: 0.006 min
Response: 529310
Conc: 6.68 ng/ml



#30 AR-1254-5


R.T.: 6.774 min
Delta R.T.: 0.003 min
Response: 1088518
Conc: 8.62 ng/ml



#31 AR-1260-1


R.T.: 7.193 min
Delta R.T.: 0.004 min
Response: 207989
Conc: 2.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.247 min
Delta R.T.: -0.005 min
Response: 714980
Conc: 7.80 ng/ml




#32 AR-1260-2

R.T.: 7.447 min
Delta R.T.: 0.001 min
Response: 403540
Conc: 4.36 ng/ml



#32 AR-1260-2


R.T.: 6.439 min
Delta R.T.: -0.003 min
Response: 1869024
Conc: 16.24 ng/ml



#33 AR-1260-3


R.T.: 7.806 min
Delta R.T.: -0.002 min
Response: 157912
Conc: 2.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.596 min
Delta R.T.: 0.000 min
Response: 717974
Conc: 6.69 ng/ml




#34 AR-1260-4

R.T.: 8.036 min
Delta R.T.: 0.001 min
Response: 310356
Conc: 4.14 ng/ml



#34 AR-1260-4


R.T.: 7.076 min
Delta R.T.: 0.003 min
Response: 647426
Conc: 7.83 ng/ml



#35 AR-1260-5


R.T.: 8.361 min
Delta R.T.: 0.005 min
Response: 245888
Conc: 1.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.319 min
Delta R.T.: 0.001 min
Response: 213532
Conc: 1.13 ng/ml




#36 AR-1262-1

R.T.: 7.806 min
Delta R.T.: 0.002 min
Response: 157912
Conc: 1.64 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 6.863 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.346 min
Delta R.T.: -0.006 min
Response: 90254
Conc: 0.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.076 min
Delta R.T.: 0.006 min
Response: 647426
Conc: 5.58 ng/ml




#38 AR-1262-3

R.T.: 8.661 min
Delta R.T.: -0.008 min
Response: 18095
Conc: 0.17 ng/ml



#38 AR-1262-3


R.T.: 7.608 min
Delta R.T.: 0.008 min
Response: 484241
Conc: 5.29 ng/ml



#39 AR-1262-4


R.T.: 8.771 min
Delta R.T.: 0.012 min
Response: 362795
Conc: 4.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.649 min
Delta R.T.: -0.016 min
Response: 459546
Conc: 2.80 ng/ml




#40 AR-1262-5

R.T.: 9.424 min
Delta R.T.: -0.003 min
Response: 212028
Conc: 4.17 ng/ml



#40 AR-1262-5


R.T.: 8.173 min
Delta R.T.: 0.009 min
Response: 621450
Conc: 8.18 ng/ml



#41 AR-1268-1


R.T.: 8.661 min
Delta R.T.: -0.005 min
Response: 18095
Conc: 0.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.608 min
Delta R.T.: 0.008 min
Response: 484241
Conc: 1.81 ng/ml




#42 AR-1268-2

R.T.: 8.771 min
Delta R.T.: 0.009 min
Response: 362795
Conc: 2.14 ng/ml



#42 AR-1268-2


R.T.: 7.649 min
Delta R.T.: -0.017 min
Response: 459546
Conc: 1.92 ng/ml



#43 AR-1268-3


R.T.: 8.992 min
Delta R.T.: -0.003 min
Response: -22573
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.872 min
Delta R.T.: 0.000 min
Response: 296491
Conc: 1.44 ng/ml




#44 AR-1268-4

R.T.: 9.424 min
Delta R.T.: -0.002 min
Response: 212028
Conc: 3.79 ng/ml



#44 AR-1268-4


R.T.: 8.173 min
Delta R.T.: 0.009 min
Response: 621450
Conc: 7.55 ng/ml



#45 AR-1268-5

R.T.: 9.847 min
Delta R.T.: -0.004 min
Response: 442673
Conc: 0.96 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.472 min
Delta R.T.: 0.010 min
Response: 181945
Conc: 0.27 ng/ml