

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111122\
 Data File : PP053281.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2022 11:07
 Operator : YP\AJ
 Sample : N5553-04RE
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TOTE-COMPRE

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/14/2022
 Supervised By :Ankita Jodhani 11/14/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 20:08:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 09 05:50:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.414	3.653	588.4E6	740.8E6	306.071	492.458 #
2) SA Decachlor...	10.209	8.745	22840403	23993903	16.845	12.801m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111122\
Data File : PP053281.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2022 11:07
Operator : YP\AJ
Sample : N5553-04RE
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
TOTE-COMPRE


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2022
Supervised By :Ankita Jodhani 11/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 20:08:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP110822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 09 05:50:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.414 min
Delta R.T.: 0.011 min
Response: 588372139
Conc: 306.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
TOTE-COMPRE


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2022
Supervised By :Ankita Jodhani 11/14/2022




#1 Tetrachloro-m-xylene

R.T.: 3.653 min
Delta R.T.: 0.004 min
Response: 740779723
Conc: 492.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.209 min
Delta R.T.: 0.006 min
Response: 22840403
Conc: 16.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.745 min
Delta R.T.: 0.017 min
Response: 23993903
Conc: 12.80 ng/ml m